A-A-51478
20 June 1986

COMMERCIAL ITEM DESCRIPTION

SCREEN, X-RAY PROTECTIVE

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a mobile, x-ray protective screen for protecting medical personnel from exposure to secondary radiation during radiologic procedures.

Salient characteristics:

The screen shall be mobile, x-ray protective, and constructed of a suitable corrosion resisting metal and having a viewing panel made of lead-acrylic.

The viewing panel shall extend from the top of the screen down 24 inches (60.9 cm), and extend across the full 30 inch (76.2 cm) width. The viewing panel shall be approximately 12 mm thick with a lead equivalency of 0.5 mm minimum.

The opaque lower panel shall be of corrosion resisting steel construction containing a 1/32 inch (0.8 mm) thick (minimum) lead screen.

The screen shall be mounted on four smooth rolling casters.

The screen shall have a nominal height of 72 inches and a nominal width of 30 inches.

The screen shall not weigh more than 130 lbs (58.9 kg).

The welds shall be durable, sound in every detail and in accordance with the latest practices recommended by The American Welding Society.

The screen shall be free of burrs, sharpness, and sharp projections.

Workmanship. The x-ray protective screen shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One screen, as specified constitutes one unit.

AMSC N/A FSC 6525

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6525-01-166-9033

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:56Z
Check the source to verify that this is the current version before use.

*MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

Project No. 6525-1196

PGC 53037

Location: CID/SCREEN-XRAY-PROTECTIVE/H13

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.