A-A-51484
23 June 1986

COMMERCIAL ITEM DESCRIPTION

DENTAL WAX SOLVENT

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers dental wax solvent.

Salient characteristics:

The dental wax solvent shall be suitable for use in removing residual dental waxes from plastic and porcelain teeth and gypsum molds, and for smoothing the surface of wax models in denture prosthesis.

The dental wax solvent shall be a clear water white or a very pale amber, nonviscous fluid containing deodorized kerosene, a suitable detergent, and methyl salicylate. The deodorized kerosene shall be of the type known as "K-220". The detergent shall be odorless and shall be a surface active agent that acts as an emulsifier in addition to enhancing the solvent action of the kerosene. The methyl salicylate shall be United States Pharmacopeia (U.S.P.) grade and shall serve as a masking agent for the objectionable odor which may result from the interaction of the kerosene and detergent. The flashpoint of the solvent shall be no less than 140 degrees Fahrenheit (^{o}F), when determined by American Society for Testing and Materials (ASTM) method ASTM D56.

Prosthetic performance tests shall be performed to determine performance by setting up both acrylic and porcelain teeth in wax for processing into a partial or full acrylic resin denture. The solvent shall have no harmful effect on resinous denture base and other materials commonly used in the construction of dentures. The solvent shall be noninjurious to skin or mucosa during such accidental contact as may occur. The solvent shall be effective in removing dental waxes from the surface of plastic and porcelain teeth after wax up. The solvent shall dissolve wax from the gypsum molds and leave the molds clean of any wax residue when flushed with boiling water. The solvent shall produce a polished surface on dental baseplate wax when applied with soft cotton wads.

Workmanship. The dental wax solvent shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Can (CN). One can containing one quart, minimum, of solvent, as specified, constitutes one unit.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more characteristics. The unit of product for inspection purposes shall be one can of dental wax solvent. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:56Z
Check the source to verify that this is the current version before use.

In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6520-00-299-9643	DENTAL WAX SOLVENT

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Civil Agency Coordinating Activity:
VA-OSS

Project No. 6520-1935

47565

Location: NCR

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:56Z
Check the source to verify that this is the current version before use.