A-A-54575
24 January 1992

COMMERCIAL ITEM DESCRIPTION

DRAINAGE KIT, CLOSED SUCTION JACKSON-PRATT

The General Services Administrations has authorized the use of this commercial item description.

This commercial item description covers two sizes of drainage kits for suction of closed wounds. Each kit consists of a suction reservoir and a Jackson-Pratt style surgical drain.

Salient Characteristics.

Material. The drain and reservoir shall be fabricated from a silicone rubber compound. The reservoir material shall be transparent to view graduations from the opposite side of reservoir.

Drain. The drain shall consist of a length of round tubing bonded to a flattened section. The drain shall be of the following sizes:

Size	Flat Width
1	7 mm ± 1 mm
2	10 mm ± 1 mm

The tubing shall be 79 ± 8 cm in length and shall have an inside diameter of 2.4 to 3.2 mm (3/32 to 1/8 inch). The flattened section shall have 3 to 4 rows of perforations into the lumen of the drain with a minimum of 60 perforations. The perforations shall be 1 to 2 mm in diameter. The rows of perforations shall be placed within the distal 75 percent of the flattened section. The flattened section shall be radiopaque throughout its length and shall have ribs on the inside surface. The juncture of the tubing to the flattened section shall be smoothly tapered to facilitate atraumatic withdrawal of the drain from a wound.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

Reservoir physical properties. The silicone compound used in the fabrication of the bulb proper shall have the properties shown below:

Physical Properties and Test Methods

Property	Value	ASTM Standard
Hardness (Shore durometer)	50-70A	D-2240
Tear Resistance	200 pounds force per inch (minimum)	D-624
Tensile Strength	950 pounds force per square inch (minimum)	D-412
Elongation	400 percent (minimum)	D-412

Style and design and dimensions. The reservoir shall be an oval-shaped aspirator bulb with a permanently attached semi-rigid silicone rubber head. The reservoir head shall have a drainage port, a suction port, and an attached bulb strap. The suction port shall have a plug attached to it which makes a mechanically secure and leak-proof connection with the drainage port.

The suction port shall protrude from the reservoir head 0.562 ± 0.005 inches. The suction port shall have three steps ranging in outside diameter from 0.156 ± 0.007 inches at the top of a step to 0.187 ± 0.007 inches at the base of a step. The top step shall be 0.16 ± 0.01 inches in length, with the top of the first step forming the distal tip of the suction port. The second and third steps shall be 0.15 ± 0.01 inches long. The drainage port plug shall be attached to the suction port between the third step base and the reservoir top. The suction port shall have an inside diameter of 0.110 ± 0.007 inches. The suction port shall connect to an integral Bunsen valve on the interior of the reservoir which allows flow of fluids into the reservoir but prevents flow from the reservoir into the suction port. The Bunsen valve shall also prevent the loss of negative pressure in the suction tubing when the reservoir drainage port is opened.

The reservoir drainage port shall protrude from the reservoir head 0.250 ± 0.020 inches, shall have an inside diameter of 0.156 ± 0.010 inches, and shall have an outside diameter of 0.219 ± 0.01 inches.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

The reservoir shall have an overall length of 4.75 ± 0.25 inches when measured from the suction reservoir tip to the bottom of the reservoir. The reservoir shall have an outside diameter of 2.20 ± 0.2 inches at the point of maximum diameter. The reservoir shall have a minimum internal volume of 125 milliliters (ml).

Graduations. The reservoir shall be graduated in 25 ml increments from 25 to 100 ml. Each graduation line shall extend around the circumference of the bulb and shall be numbered. The reservoir shall bear the letters "cc" or "ml" beside or near the 100 ml graduations. Graduations shall be accurate to within ± 10 percent of the labeled volume. Accuracy shall be tested with water and a graduated cylinder.

Performance. Finished reservoirs shall be tested in accordance with the vacuum and flow rate test.

Vacuum test. Attach a manometer or pressure gauge to the suction port. Manually or mechanically evacuate the air from the reservoir and install the reservoir drain plug. Release the reservoir. The reservoir shall create a minimum vacuum of 85 mm of mercury, and shall hold the vacuum for a minimum of 2 minutes.

Flow rate test. Mount the reservoir by the strap above a supply basin of water at room temperature such that a hydrostatic head of 7 ± 1/8 inches exists from the water level in the basin to the entrance of the reservoir. The reservoir may be held by a mechanism other than the strap as long as the reservoir shape recovery and the vacuum induced are not influenced by the mechanism and the 7 inch head is maintained. Manually or mechanically evacuate the air from the reservoir, install the reservoir drain plug, and release the reservoir. The reservoir shall collect 33-45 ml of water in 2 minutes.

Radiopacity. The flat portion shall be radiopaque throughout its length in accordance with ASTM Designation F640.

Sterility and Pyrogenicity. The drainage kit shall be sterile and free from pyrogens and shall be tested for sterility and pyrogenicity in accordance with the United States Pharmacopia (USP). Sterility assurance level shall be 0.000001.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

Biocompatibility. Material used in the manufacture of the drainage kit shall meet the requirements of the USP for class IV plastics and USP Interdermal Sensitivity Test.

Workmanship. The item shall be processed in such a manner as to be uniform and free from defects that will affect life, serviceability, or appearance.

Unit. Package (PG). One package containing 5 individually packaged drainage kits of one size only, as specified, constitutes one unit. Each kit shall be individually packaged in a sealed peel-open container capable of maintaining sterility of contents unless package is damaged or opened.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in specification where such inspection are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records related.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one drainage kit. Sampling shall be inspection level II with an acceptable quality level (AQL) of 1.0.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for test shall be conducting in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one drainage kit. Sampling shall be inspection level S-2 with an AQL of 1.0.

Tests. Tests shall be conducted to determined compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers and their respective Item Identifications are covered by this document:

NSNS	ITEM IDENTIFICATION
6515-01-341-4523	DRAINAGE KIT, CLOSED SUCTION, Jackson-Pratt, Radiopaque, 97.5 cm Long 10 mm Wide Drain, 100 ml Reservoir, Disposable, 5's
6515-01-341-4524	DRAINAGE KIT, CLOSED SUCTION, Jackson-Pratt, Radiopaque, 97.5 cm Long 7 mm Wide Drain, 100 ml Reservoir, Disposable, 5's

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-5317
Location: drainage.doc/S-11

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.

2 The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER A-A-54575	2 DOCUMENT DATE (YYMMDD) 24 January 1992

3 DOCUMENT TITLE
DRAINAGE KIT, CLOSED SUCTION JACKSON PRATT

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME (Last, First, Middle Initial)	b ORGANIZATION	
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7 DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME	b TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON
c ADDRESS (Include Zip Code)	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 Previous editions are obsolete

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that the document is complete before use.