A-A-54579
28 February 1992

COMMERCIAL ITEM DESCRIPTION

INSTRUMENT SET, SURGICAL, DISSECTION, VITREORETINAL

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers an autoclavable vitreoretinal dissection instrument set.

Salient characteristics:

The instrument set shall consist of three (3) surgical dissection instruments contained in a protective sterilization case. The surgical instruments shall be designed and constructed to facilitate ophthalmic surgical procedure involving the dissection of delicate tissue planes in the vitreous cavity, including separation of cortical vitreous and epiretinal membranes from the inner surface of the retina.

The sterilization case shall have the capacity to hold six (6) surgical instruments and shall be suitable for use as a container for holding the surgical instruments while they are being sterilized in an autoclave. The sterilization case shall also be suitable for storage and transportation of the surgical instruments, when outside of an autoclave.

Surgical Instruments. Shall be supplied with three surgical dissection instruments having different tip configurations. The surgical instruments shall be supplied with the following three (3) tip configurations:

One tip with a 90° blunt hook to engage the proximal edge of membranes.

One tip with a 130° blunt hook to engage the distal edge with dissection in the direction of the incision quadrant.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

One right angled blunt spatula to elevate vitreous and other tissue bridging, or break delicate vitreoretinal adhesions.

The handle and tip of each of the required surgical instruments shall conform to the requirements in figure 1 for dimensions and configuration.

Each required surgical instrument shall be light weight, corrosion-resistant, autoclavable, and designed to be handheld and non-powered. Each required surgical instrument shall be six (6) inches long and shall consist of:

A four (4) inch diamond knurled handle (see figure 1) fabricated from Beta III titanium alloy.

A two (2) inch tapered tip (see figure 1) fabricated from Beta III titanium alloy or 17-4PH corrosion-resisting steel.

Sterilization Case. Shall be supplied with a protective sterilization case that is designed to hold surgical instruments securely in place without inflicting any damage to the tip ends. The sterilization case shall have nominal overall dimensions of 8 inches long X 3 inches wide X 1 inch deep, and shall include the following:

A housing with a low reflective glare, that is fabricated from 22 gauge type 304 corrosion-resisting steel and includes drain holes for condensate collected during a sterilization process (auotclaving).

A hinged corrosion-resisting metal cover with a compression spring catch.

A pivotable metal block that is secured toward an inner edge of the housing (short dimension)

A bracket that is secured within the housing at a location above the pivotable metal block.

All metal parts shall be suitably protected against corrosion.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

The block and bracket shall have aligned slots or cutouts that provide specific positions for holding and separating six (6) surgical instruments. The block shall support the notched sectors of surgical instrument handles (see figure 1). The bracket shall support the surgical instruments at their shafts which are located above the knurled handles (see figure 1). The closure cover and drain holes shall permit the tips of the surgical instruments to receive proper steam flow during autoclaving.

Workmanship. The instrument set and sterilization case shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One instrument set, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination and testing. Examination and testing shall be conducted on the surgical instruments and sterilization case in accordance with commercial practice. The sampling for examination shall be in accordance with commercial practice.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. (Intermediate/exterior package quantities labeling, and marking must be specified in the contract and/or order).

NOTE: The following National Stock Number is covered by this document:

NSN	Item Identification
6515-01-168-0554	INSTRUMENT SET, SURGICAL, DISSECTION VITREORETINAL

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-5203

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54579	27 January 1992

3. DOCUMENT TITLE

INSTRUMENT SET, SURGICAL DISSECTION, VITREORETINAL

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code)	
Defense Personnel Support Center	(1) Commercial (215) 737-2870	(2) AUTOVON 444-2870
c. ADDRESS (Include Zip Code) ATTN: MSE 2800 S. 20th Street Philadelphia, PA 19145.	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.