INCH-POUND

A-A-54599
6 March 1992

COMMERCIAL ITEM DESCRIPTION

BANDS, PERSONAL IDENTIFICATION

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for A-A-30044B, which is cancelled.

This Commercial Item Description covers three sizes of a personal identification band to be worn on the wrist for identification of hospital patients.

Salient Characteristics:

The personal identification bands shall be disposable, made from plastic that is soft, pliable, and clear or natural color with a rigidity of not more than 50 grams per centimeter when tested as specified in ASTM D2923. The bands shall fasten by means of a suitably finished, nonirritating, corrosion-resisting, metal or plastic closing device or an adhesive that will form a bond that cannot be opened by hand without easily visible destruction of the bond. The metal or plastic closing device shall be so designed that it prevents removal and reattachment to the patient without clearly visible signs of damage. No element of the band or closing device shall be dermally toxic nor act as a skin sensitizer or irritant. The bands shall be provided with white identification cards capable of being inserted into the band and capable of accepting at least two lines of typed, written or imprinted information. A pocket in the sleeve of each band to accept the identification card shall be provided and shall be such that cards ranging in size from 3/8 to 13/16 inch wide by 2-3/4 to 4-1/16 inches long shall be accommodated. Each unit of issue shall include directions for use and ballpoint pen containing waterproof ink. The bands shall be available in three sizes as follows:

AMSC/NA FSC 6530

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Size 1 - Mother-infant patients. Each band shall have three easily separable sections. One shall be for the mother, of sufficient length to form an 8-1/2 inch circumference when clasped. The other two, for the infant, shall be of sufficient length to form a 5 inch circumference when clasped. The band for the mother shall be at least 7/16 inch, but not more than 1 inch, in width. The infant bands shall be at least 7/16 inch, but not more than 1/2 inch, in width. Each of the three sections shall be permanently and legibly imprinted with the same unique serial number suitable for correlated mother-infant identification. All bands within a lot shall have different unique serial numbers.

Size 2 - Pediatric patient. The band shall be of sufficient length to form a 7 inch circumference when clasped, and shall be at least 7/16 inch, but not more than 7/8 inch, in width.

Size 3 - Adult patient. The band shall be of sufficient length to form a 9 inch circumference when clasped, and shall be at least 7/16 inch, but not more than 15/16 inch, in width.

Workmanship. The personal identification bands shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG).

For Size 1. One package containing 200 bands and a minimum of 720 perforated cards, as specified, constitutes one unit.

For Size 2 and Size 3. One package containing 400 bands and a minimum of 440 perforated cards, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be five identification bands and five cards.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document.

6530-00-104-7639 - Size 1
6530-00-982-5118 - Size 2
6530-00-104-7631 - Size 3

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA - OSS
USPHS
FDA - MPQAS

Project Number: 6530-2294

Location: IDMC11.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54599	2. DOCUMENT DATE (YYMMDD) 6 March 1992

3. DOCUMENT TITLE
BANDS, PERSONAL IDENTIFICATION

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	e. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.