INCH-POUND

A-A-54601
20 March 1992

COMMERCIAL ITEM DESCRIPTION

ULTRASONIC PROPHYLAXIS UNIT, DENTAL, PORTABLE

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers an ultrasonic prophylaxis unit suitable for use in oral prophylaxis.

Salient characteristics:

Material. The material used shall be of the highest quality throughout and shall be suitable in all respects for the intended purpose.

Ultrasonic prophylaxis unit. The ultrasonic prophylaxis unit shall be portable and shall consist of four major components: an electronic generator, a handpiece assembly, a foot switch, and three interchangeable inserts (30K-3, 30K-10 and 30K-EWPP), see Note. The unit shall be suitable for use in oral prophylaxis.

Electronic generator. The generator and water system shall be housed in a portable table model cabinet having the following nominal dimensions: 4 inches high, 7-3/8 inches wide and 7-5/8 inches deep, and shall have a nominal weight of 5.3 lbs. The front panel of the cabinet shall have a power selector dial, an on-indicator, a water flow adjustment knob, and the handpiece receptacle. The power selector dial shall be capable of variable power selection. The water flow adjustments dial shall control the volume and temperature of water flowing into the handpiece. The unit shall be properly fused, and the rear of the cabinet shall provide for power line connection and foot switch connection. The water connection shall be made on the underside of the unit. The unit shall have an internal water regulator to provide a constant flow of water, and shall have an electrical interlock to prevent water flow when handpiece is disconnected. The unit shall operate on 120 volt, 50/60 Hz, AC current input, and shall draw approximately 92 watts. The unit shall provide maximum power output to the handpiece at 30,000 $\pm$ 1500 hertz per second, at a nominal 0.8 amps and 30 watts.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Handpiece assembly. The handpiece assembly shall consist of the handpiece proper (housing and handle for the interchangeable prophylaxis inserts) attached to a cable fitted with a cable plug suitable for use with the receptacle on the front panel of the cabinet.

Foot switch. The foot switch shall control the unit (when power switch is in any position and the main switch in "on"), which shall activate the handpiece and turn on the water flow to the prophylactic tip. Removal of foot from switch shall shut off both handpiece and water flow. The switch shall be attached to a cable fitted with a cable plug suitable for use with the receptacle on the rear panel of the cabinet.

Inserts. The inserts shall be fabricated of a stainless steel tip, polysulfane molded nozzle, eyelet, O-ring, connecting body and transducer (magnetostrictive stock). All inserts shall contain a means for delivery of water to the tip. The O-ring shall act as a water seal when the insert is placed into the open end of the handpiece. The transducer shall convert electrical energy to mechanical energy in the form of oscillations which shall be transmitted through the connecting body to drive the insert tip. The inserts shall be quickly and easily interchanged (snap-in and snap-out).

Identification marking. Each prophylaxis unit shall be permanently and legibly marked with the manufacturer, model designation, serial number, and electrical characteristics, such as voltage, cycle and amperage.

Service data. The manufacturer shall furnish, with each unit, two copies each of a technical manual which contains complete instructions for installation, operation, and maintenance and a list of component parts.

Workmanship. The ultrasonic prophylaxis unit and inserts shall be free from defects which detract from their appearance or may impair their serviceability.

Unit. Each (EA). One ultrasonic prophylaxis unit and three inserts, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection, examination, and test characteristics. The unit of product for inspection, examination, test purposes shall be one ultrasonic unit and each insert unless otherwise specified, sampling for examination shall be conducted at inspection level II, except for dimensions which shall be conducted at inspection level S-1, AQL 2.5 (percent defective) and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0 (percent defective).

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Tests. Tests shall be conducted to determine compliance with specification requirements..

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Chemical analysis. In the absence of contractor certification for the raw material, a chemical analysis shall be conducted. The chemical analysis shall be performed by reagent solutions, combustion methods, nonemission methods, or by spectrochemical analysis. Any ASTM method shall be acceptable.

Operational test. Each ultrasonic prophylaxis unit shall be given an operational test in accordance with the operational instructions accompanying the ultrasonic prophylaxis unit. Non-conforming units shall be rejected.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

NOTES:

1. The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-01-313-6250	ULTRASONIC PROPHYLAXIS UNIT, DENTAL, PORTABLE, 120 volt, 50/60 Hz, AC

2. The following National Stock Numbers are referenced by this document:

National Stock Number	Item Identification
6520-01-281-8793	INSERT, HANDPIECE, DENTAL, ULTRASONIC PROPHYLAXIS UNIT, No. 10, Curette Type
6520-01-313-6269	INSERT, HANDPIECE, DENTAL, ULTRASONIC PROPHYLAXIS UNIT, No. 3, Flat-Hoe Chisel Type
6520-01-313-6270	INSERT, HANDPIECE, DENTAL, ULTRASONIC PROPHYLAXIS UNIT, No. EWPP, Probe Type

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2514
LOCATION. KAD/#2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.