DATA ITEM DESCRIPTION

Form Approved
OMB No. 0704-0188

TITLE	1. IDENTIFICATION NUMBER
REQUEST FOR APPROVAL OF NONSTANDARD GENERAL PURPOSE ELECTRONIC TEST EQUIPMENT	DI-MISC-80056B

3. DESCRIPTION / PURPOSE

3.1 This request is used to obtain approval from the Navy for use of nonstandard general purpose electronic test equipment (GPETE) which is not listed as a standard in MIL-STD-1364.

4. APPROVAL DATE (YYMMDD)	5. OFFICE OF PRIMARY RESPONSIBILITY (OPR)	6a. DTIC REQUIRED	6b. GIDEP REQUIRED
890717	NAVY - SH/CEL DS		

7. APPLICATION / INTERRELATIONSHIP

7.1 This Data Item Description contains format and content preparation instructions for that data generated under the work task described by paragraph 5.4 of MIL-STD-1364H.

7.2 These data are used to describe nonstandard GPETE which has been identified by a contractor or other activity and is required to support military systems, subsystems, and other equipment within the Navy.

7.3 This DID supersedes DI-MISC-80056A.

8. APPROVAL LIMITATION	9a. APPLICABLE FORMS	9b. AMSC NUMBER
	NAVSEA 10550/2A	N4790

EPARATION INSTRUCTIONS

10.1 Reference document. The applicable issue of the documents cited herein, including their approval date, and dates of any applicable notices and revisions shall be as reflected in the contract.

10.2 Format. Application for approval of nonstandard GPETE shall be in the format of NAVSEA 10550/2A. Instructions for completing NAVSEA 10550/2A are as follows:

10.2.1 Block 1 of NAVSEA 10550/2A. Only one equipment type (with applicable options, plug-ins, and ancillaries) shall be included in an individual submission. The equipment request shall be completely identified in accordance with a through d:

a. Equipment name. The equipment name shall be the complete noun name assigned by the manufacturer. Acronyms and abbreviations should be avoided wherever possible.
b. Manufacturer and model. The manufacturer and model shall be exactly as identified by the manufacturer. Required options, plug-ins, and ancillaries shall be included as described by the manufacturer.
c. Quantity required and unit cost. The quantity shall be the total required for the specified application. The unit cost shall be the manufacturer's standard single unit catalog price. (NOTE: In order to avoid resubmittal, out-year requirements, if any, should be identified by fiscal year of intended procurement.)

TRIBUTION STATEMENT

DISTRIBUTION STATEMENT A Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2020-10-20T10:34Z
Check the source to verify that this is the current version before use.

10. PREPARATION INSTRUCTIONS (Continued)

d. Functional description. An attachment that provides a complete description of the equipment shall be included with the application. Reference to a manufacturer's current catal g and page number is acceptable provided the catalog entry completely describes the equipment.

10.2.2 Block 2 of NAVSEA 10550/2A. Indicate whether approval to procure the equipment identified in paragraph 1 has been previously submitted by the applying activity. If "has" is indicated, the original NESEA Control Number and any extension or reapplication control numbers shall be identified.

10.2.3 Block 3 of NAVSEA 10550/2A. The military or commercial model numbers(s) and noun name(s) of the prime system requiring the capabilities provided by the requested equipment shall be identified. Model number(s) alone or noun names(s) alone constitute incomplete information and are not acceptable.

10.2.4 Block 4 of NAVSEA 10550/2A. Test data, test reports, or test requirements that support the need for the nonstandard GPETE shall accurately reflect the minimum measurement parameters of the prime system and shall not be merely a list of the characteristics of the proposed nonstandard GPETE. (Any classified information must be forwarded in accordance with regulations governing transmission of such data.)

10.2.5 Block 5 of NAVSEA 10550/2A. If a similar GPETE item exists in MIL-STD-1364, it shall be identified and compared to the nonstandard item. When there is no similar standard GPETE item, the word "none" shall be entered. The measurement parameter(s) required but not provided in the nearest similar standard GPETE item shall be identified.

10.2.6 Block 6 of NAVSEA 10550/2A. In cases where criteria other than measurement parameters are the basis for the application (such as configuration, cost, dimensions), a comparison to the MIL-STD-1364 GPETE item shall be provided.

10.2.7 Block 7 of NAVSEA 10550/2A. The MIL-T-28800 type, class, and style of the nonstandard GPETE (if known) shall be indicated.

10.2.7.1 Block 7.1 of NAVSEA 10550/2A. Any specific environmental conditions applicable to the intended use shall be identified.

10.2.8 Block 8 of NAVSEA 10550/2A (UIC). If applicable, the Unit Identification Code(s) (UIC) of the end user(s) and the MARK FOR location of intended use shall be supplied.

10.2.9 Block 9 of NAVSEA 10550/2A (Originator). The originator of the application shall be completely identified. Applications prepared by the contractor will not be processed without the appropriate Navy program manager's certification. The Test and Monitoring Systems (TAMS) Manager shall be identified as NAEC (Code 92413); Navy Weapons Station, Earle (Code 8052); NAVSEA (Code 07031); or NESEA (Code 2261), as applicable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:34Z
Check the source to verify that this is the current version before use.

APPLICATION FOR APPROVAL OF NON-STANDARD GENERAL PURPOSE ELECTRONIC TEST EQUIPMENT (GPETE)

Naval Sea Systems Command
NAVSEA CEL-DS1
Washington, DC 20362-5101

OMB NO. 0704-0188

DATE

1. In accordance with the provisions of MIL-STD-1364(NAVY), the undersigned hereby applies for approval of the following non-standard GPETE:

a. Equipment Name

b. Manufacturer and Model

c. Quantity Required and Unit Cost

d. Functional Description (Attach catalog, specification, or other information)

2. Application for approval of this equipment (has) (has not) been previously submitted by this activity.

Approved? ☐ YES ☐ NO

NONSTANDARD GPETE CONTROL NUMBER(S)

3. The non-standard GPETE is to be used in support of the following prime system(s) or equipment

4. Applicable minimum measurement parameters required by the prime system

4.1 Describe test data, test reports, test requirements, etc., which support the requirement for the nonstandard GPETE. (Reference by document number only is not acceptable.)

5. Nearest similar MIL-STD-1364 item

5.1 Measurement parameters required, but not provided, in nearest similar MIL-STD-1364 item

6. If optional configuration, dimensions, or cost are the basis for the application, provide comparison data for the nonstandard GPETE and the MIL-STD-1364 GPETE.

7. Required type, class, and style of the nonstandard GPETE from MIL-T-28800 (if known)

7.1 Identify specific environmental conditions that the nonstandard GPETE will be subjected to. Include the maximum temperature humidity range(s) under which the instrument will be operated/stored.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:34Z
Check the source to verify that this is the current version before use.

8. UIC of end user(s); include "mark for" instructions.	
9. Originator (Type or Print Name)	☐ Government ☐ Contractor (requires Navy certification, see 8.1)
Signature (with code)	Telephone (Autovon & Commercial)
Activity/Location	Mailing Address
TAMS Manager (NAEC, NAVWPNSTA EARLE, NAVSEA, NESEA) Validation	
9.1 Navy Program Manager Certification (Type or Print Name)	Signature (with code)
Title	Telephone (Autovon & Commercial)
Activity/Location	
10. Additional Comments/Remarks	

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:34Z
Check the source to verify that this is the current version before use.