12 June 1975
SUPERSEDING
MIL-S-43710
7 August 1970

MILITARY SPECIFICATION

SCALE, DIAL INDICATING, AUTOPSY

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers one type of dial indicating autopsy scale (see 6.1).

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

FEDERAL

PPP-B-1122 - Balances, Scales, and Accessories, Packaging and Packing Of

STANDARD

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Standard product. The scale covered by this specification shall be the current, standard product of the manufacturer, except for any changes necessary to comply with this specification.

Source: https://assist.dla.mil -- Downloaded: 202610-21T22:05Z
Check the source to verify that this is the current version before use.

3.2 Material. Material not definitely specified shall be of the quality normally used by the manufacturer for scales, providing the completed item complies with all the provisions of this specification.

* 3.3 Design and construction. The scale shall be a double dial indicating, multi-revolution, hanging style, spring scale with a round pan load receiver. Revolutions shall be counted by a revolution indicator slide marked at 10 kilograms (kg). The rated capacity of the scale shall be 15 (kg). The scale shall show no evidence of bends, deformation, breakage, or misalignment when loaded with 1.25 times the rated capacity (see 4.3.2). The accuracy of the scal shall be ± 1 graduation (when tested as specified in 4.3.1). The dials shall be circular and shall have one dial facing front and one dial facing the back. The dial faces shall be white, with black numbers and graduations, and shall be provided with a pointer. The dial shall be graduated to 10 kg by 25 grams. The dial shall be protected by clear glass windows and have a corrosion-resisting steel carbon steel sash. The diameter of the dial shall be not less than 8 inches. The scale shall be provided with one or more air dash pot. The load receiver shall be fabricated of corrosion-resisting steel. The load receiver for the scale shall be 15 inches, ± 1/4 inch, in diameter and 4-1/4 inches, ± 3/4 inch, deep. The bottom of load receivers shall be perforated. The load receiver shall be removable. The hangers for the load receiver shall be fabricated of corrosion resistant metal.

* 3.4 Finish. The load receivers shall be finished with a No. 4 or better polish. Exterior ferrous metal surfaces other than corrosion resisting steel shall be enameled or pleated in accordance with the manufacturer's commercial practice. The dial shall be enameled.

3.5 Marking. The scale shall be marked with the total capacity of the scale and the manufacturer's name or trademark of such known character to be identifiable to the manufacturer.

3.6 Workmanship. The finished scale shall be free of sharp edges, burrs, slivers, splinters, or tacky and uneven finish. The windows on the dial shall not be loose, cracked, chipped, or contain opaque spots. The numbers and graduation markings on the dial shall be legible, clean and evenly spaced.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T22:05Z
Check the source to verify that this is the current version before use.

the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.2 Inspection. Sampling for inspection shall be performed in accordance with MIL-STD-105, except where otherwise indicated hereinafter.

4.2.1 Component and material inspection. In accordance with 4.1 above, components and materials should be inspected in accordance with all the requirements of referenced specifications, and standards unless otherwise excluded, amended, modified, or qualified in this specification or applicable purchase document.

4.2.2 End item inspection. A lot shall consist of all the scales offered for inspection at one time. The sample unit shall be one completely fabricated scale.

* 4.2.2.1 Visual examination. The end item shall be examined for the defects classified in table I. The inspection level shall be II with an acceptable quality level (AQL) of 1.0 for major defects and 2.5 for total defects expressed in terms of defects per hundred units.

TABLE I. Classification of defects

Examine	Defect	Classification Major	Minor
Finish	Exterior ferrous metal surfaces not enameled, when required	X	
	Load receivers not finished as specified	X	
Construction and workmanship (applicable to all components)	Component missing	X	
	Not multirevolution type	X	
	Component not properly assembled or secured	X	
	Component fractured or malformed	X	
	Burr, sliver, or splinter	X	

Source: https://assist.dla.mil -- Downloaded: 202610-21T22:05Z
Check the source to verify that this is the currentversion before use.

TABLE I. Classification of defects (cont'd)

Examine	Defect	Classification Major	Minor
Dial and graduations	Pointer missing	X	
	Not white with black graduation	X	
	Not graduated as specified	X	
	Graduations not even	X	
	Not double dial		X
	Not circular	X	
	Window loose, cracked, chipped, opaque, or not clean		X
	Revolution indicator slide missing	X	
Load receivers	Not perforated	X	
	Not removable	X	
Instruction manual	Missing, or contents not as specified		X
Identification marking	Missing, incomplete, not legible		X

* 4.2.2.2 Dimensional examination. Inspection shall be made of the end item for compliance with dimensions specified. Any deviation from the specified requirements shall constitute a defect. The inspection level shall be S-2 and the AQL shall be 2.5 expressed in terms of defects per hundred units.

4.2.2.3 Testing of the end item. The scale shall be tested as specified in 4.3.1 and 4.3.2. The inspection level shall be S-2. Any defect shall be cause for the rejection of the lot.

4.2.3 Examination of preparation for delivery. Examination of preparation for delivery shall be made in accordance with PPP-B-1122.

4.3 Tests.

* 4.3.1 Accuracy test. The tolerance of the scales shall be determined at 1, 10, and 15 kg. Class F weights, traceable to the National Bureau of Standards, shall be applied at the center of the load receiver. The test shall be repeated 10 times and the dial reading observed each time. A reading in excess of 1 graduation from the weight applied shall constitute failure of the test.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T22:05Z
Check the source to verify that this is the current version before use.

* 4.3.2 Overload test. A load of 18.75 kg shall be applied to the center of the load receiver. A bend, deformation, breakage, or misalignment shall constitute failure of this test.

5. PREPARATION FOR DELIVERY

5.1 Preservation, packaging, packing and marking. Preservation, packaging, packing and marking for level A, B, or C, shall be in accordance with PPP-B-1122.

6. NOTES

6.1 Intended use. The scale covered by this specification is intended to weigh specimens during autopsies.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number and date of this specification.
(b) Selection of applicable levels of preservation, packaging, and packing (see 5.1).

6.2.1 Contract data requirements. Any data items to be delivered under any contract for items covered by this specification should be specifically called for in the contract in accordance with the applicable regulations of the procuring activity.

* 6.3 The margins of this specification have been marked with an asterisk (*) to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and suppliers are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Custodians:

Army - GL
Navy - MS
Air Force - 03

Review activity:

Army - MD
DSA - GS, DM

Preparing activity:

Army - GL

Project No. 6670-0115

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T22:05Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T22:05Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-S-43710A - SCALE, DIAL INDICATING, AUTOPSY

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T22:05Z
Check the source to verify that this is the current version before use.

FOLD

DEPARTMENT OF THE ARMY
US ARMY NATICK DEVELOPMENT CENTER
NATICK, MA 01760

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

AMXNM-TD T

POSTAGE AND FEES PAID
DEPARTMENT OF THE ARMY
DOD – 314

U.S. MAIL

Commander
US Army Natick Development Center
ATTN: AMXNM-TDT
Natick, MA 01760

NATICK OP 1 MAY 75

FOLD

Source: https://assist.dla.mil -- Downloaded: 202610-21T22:05Z
Check the source to verify that this is the current version before use.