FEDERAL SPECIFICATION

CATHETERS, URETHRAL, RUBBER

This amendment, which forms a part of Federal Specification ZZ-C-101a, dated January 4, 1951, was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

Page 1, paragraph 2.1: Delete parenthetical paragraph and substitute the following:

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

(Copies of Military Specifications and Standards required by the contractor in connection with specific procurement functions, should be obtained from the procuring activity or as directed by the contracting officer.)

Page 5, after paragraph 6.4, add new paragraph as follows:

6.5 Transportation description.—Transportation description applicable to this item is:

Rubber goods, not otherwise indexed by name. Subject to Rule 34, Uniform Freight Classification.

Carload minimum weight 20,000 pounds.

Truckload minimum weight 20,000 pounds

Section 7 (pages 5, 6 and 7), delete in its entirety and substitute:

7. DEPARTMENTAL REQUIREMENTS

7.1 The following specifications and standard, of the issues in effect on date of invitation for bids, and special requirements, form a part of this specification for purchases made under this specification by the respective departments.

7.2 Army, Navy and Air Force.

7.2.1 *Applicable specifications and standard:*

Federal Specifications:

PPP–B–566 — Boxes, Folding, Paperboard.

PPP–B–585—Boxes; Wood, Wirebound.

PPP–B–601—Boxes, Wood, Cleated-Plywood.

PPP–B–621—Boxes, Wood, Nailed and Lock-Corner.

PPP–B–676 — Boxes, Set-Up, Paperboard.

Military Specifications:

MIL–B–10377—Box, Wood, Cleated, Veneer, Paper Overlaid.

MIL–L–10547 — Liners, Case, Waterproof.

Military Standard:

MIL–STD–129 — Marking for Shipment and Storage.

402174/57/No. 328

FSC Class 6515

Source: https://assist.dla.mil - Downloaded: 2026-10-26T12:25Z
Check the source to verify that this is the current version before use.

7.2.2 *Packaging.*

7.2.2.1 *Unit of issue.* — One catheter, as specified, constitutes one unit.

7.2.2.2 *Quantities.*—Packaging and packing quantities shall be as specified in table V.

TABLE V.—*Packaging and packing quantities*

Unit pack	Intermediate pack	Shipping container
1 unit	6 units	360 units

7.2.2.2.1 *Packing variation permitted.*—If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

7.2.2.3 *Level A.*

7.2.2.3.1 *Unit package.* — Each catheter shall be packaged (not coiled) in a waxed paper or glassine envelope and secured. As an alternate to the above, each unit may be completely wrapped (not coiled) in waxed or glassine paper and secured.

7.2.2.3.2 *Intermediate package.*—The number of unit packages specified in table V shall be packaged in a snug-fitting carton conforming to Federal Specification PPP–B–566 or PPP–B–676. Carton shall be secured to prevent accidental opening.

7.2.2.3.3 *Level C.*—Unless otherwise specified, each catheter shall be packaged in accordance with accepted commercial practice.

7.2.3 *Packing.*

7.2.3.1 *Level A.*

7.2.3.1.1 *Exterior container.*—The number of units specified in table V shall be overpacked in an exterior container designed for type 2 load and constructed in accordance with Federal Specification PPP–B–585, style 3, for class 3 use; PPP–B–601, table II, using type III, class 1 plywood; PPP–B–621, table III; or Military Specification MIL–B–10377, table II.

7.2.3.1.2 *Case liner.*—Each box shall be lined with a waterproof case liner conforming to Military Specification MIL–L–10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto.

7.2.3.1.3 *Strapping.*—Strapping (when required) shall be in accordance with applicable box specification.

Note.—Strapping will not be required for shipment forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

7.2.3.2 *Level B.*

7.2.3.2.1 *Exterior container.*—The number of units specified in table V shall be overpacked in an exterior container conforming to Federal Specification PPP–B–601, table I, using class 1 plywood; or an exterior container designed for type 2 load and constructed in accordance with Federal Specification PPP–B–585, style 3, for class 2 use; PPP–B–621, table I; or Military Specification MIL–B–10377, table I.

7.2.3.3 *Level C.*—The subject commodity shall be packaged and packed in substantial commercial containers of the type, size, and kind, commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

7.2.4 *Marking.*—Each unit package, each intermediate carton, and each shipping container shall be marked in accordance with Military Standard MIL–STD–129. Service color marking and symbol are not required on containers for level A packs when addressed to domestic establishments, for the Medical Services of the Armed Forces.

Page 7, bottom of page: Delete "For Sale" statement and substitute the following: "Copies of this specification may be purchased for 5 cents."

★ U. S. GOVERNMENT PRINTING OFFICE: 1957

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:25Z
Check the source to verify that this is the current version before use.