RR-B-96b
JULY 18, 1957
SUPERSEDING
Int. Fed. Spec. RR-B-0096a (GSA-FSS)
June 18, 1956 and
Fed. Spec. RR-B-96
July 9, 1946

FEDERAL SPECIFICATION

BAND SET, COPPER, DENTAL AND COPPER STRIP, DENTAL

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope.—This specification covers copper bands and strips suitable for dental use.

1.2 Classification.

1.2.1 *Types.*—Copper bands and strips shall be of the following types, as specified (see 6.2):

Type I.—Copper bands.

Type II.—Copper strips.

1.2.2 *Sizes.*—Copper bands shall be of the sizes shown in table I (see 6.2).

2. APPLICABLE SPECIFICATIONS AND STANDARDS

2.1 The following specifications and standards, of the issues in effect on date of invitation for bids, form a part of this specification:

Federal Specifications:

GG-D-441 — Dispensers; for Articulating Paper, Celluloid, or Steel Rolls (Dental).

PPP-B-566 — Boxes, Folding, Paperboard.

PPP-B-585—Boxes; Wood, Wirebound.

PPP-B-601—Boxes, Wood, Cleated-Plywood.

PPP-B-621—Boxes, Wood, Nailed and Lock-Corner.

PPP-B-676 — Boxes, Set-Up, Paperboard.

Federal Standards:

Fed. Std. No. 102—Preservation, Packaging and Packing Levels.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Military Specifications:

MIL-P-116—Preservation, Methods Of.

MIL-B-10377—Box, Wood, Cleated, Veneer, Paper Overlaid.

MIL-L-10547 — Liners, Case, Waterproof.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:08Z
Check the source to verify that this is the current version before use.

Military Standards:

MIL–STD–105 — Sampling Procedures and Tables for Inspection by Attributes.

MIL–STD–129—Marking for Shipment and Storage.

(Copies of Military Specifications and Standards required by the contractor in connection with specific procurement functions should be obtained from the procuring activity or as direced by the contracting officer.)

3. REQUIREMENTS

3.1 Material.—Bands and strips shall be made of copper, having a purity of at least 99.88 percent, silver being counted as copper.

3.2 Type I.

3.2.1 *Description.*—Copper bands shall be seamless, with smooth edges, and unannealed.

3.2.2 *Dimensions.*—The bands shall measure 0.50 ± 0.015 inch in length, with a wall thickness of 0.005 ± 0.0015 inch. Inside diameter shall be as indicated in table I for the respective size, with a tolerance of ± 0.002 inch.

TABLE I.—*Inside diameters of type I bands*

Size	Inside diameter
	Inch
1	0.188
2	0.203
3	0.219
4	0.234
5	0.250
6	0.266
7	0.281
8	0.297
9	0.313
10	0.328
11	0.344
12	0.359
13	0.375
14	0.391
15	0.406
16	0.422
17	0.438
18	0.453
19	0.469
20	0.484

3.2.3 *Band set.*—Unless otherwise specified, a band set shall consist of 100 assorted bands in the sizes and number specified in table II.

TABLE II.—*Band set*

Size	Number of bands
5	4
6	5
7	8
8	12
9	10
10	6
11	6
12	6
13	6
14	4
15	4
16	6
17	7
18	7
19	5
20	4

3.3 Type II.

3.3.1 *Description.*—Strips shall be soft annealed, have a smooth surface and shall fit dispensers in accordance with requirements of Federal Specification GG–D–441.

3.3.2 *Dimensions.*—Copper strips shall be furnished in rolls of not less than 20 feet long. Strips shall have a thickness of 0.005 ± 0.0015 inch, and a width of not less than 0.50 inch nor more than 0.5625 inch. The maximum outside diameter of the roll shall be 1.6 inches.

3.3.3 *Bend test.*—The strips shall withstand the bend test specified in 4.3.2.

3.4 Workmanship.—Workmanship shall be first class throughout. Bands and strips shall be clean, smooth, and free from defects which detract from their appearance or may impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:08Z
Check the source to verify that this is the current version before use.

4. SAMPLING, INSPECTION AND TEST PROCEDURES

4.1 Sampling.—Unless otherwise specified, samples shall be submitted in accordance with instructions furnished by the procuring activity (see 6.2). In addition, sampling procedures shall be conducted in accordance with the procedures set forth in Military Standard MIL–STD–105.

4.2 Inspection.—Inspection may be made througout the entire process of manufacture. The passing as satisfactory of any detail of construction or materials shall not relieve the contractor of responsibility for faulty workmanship of materials which may be discovered at any time prior to final acceptance. Final inspection of the finished article shall be made either at point of production or at point of delivery, as specified (see 6.2). In case of factory inspection, every facility shall be afforded inspectors, by the manufacturer, for the prosecution of their work.

4.3 Tests.—Tests shall be conducted in accordance with the applicable specifications, as necessary, to determine compliance with this specification.

4.3.1 *Chemical analysis.*—Copper content may be determined by chemical analysis or spectroanalysis.

4.3.2 *Cold bend test.*—From each lot, a specimen of copper strip 6 inches long shall withstand bending flat on itself without fracture on the outer surface.

5. PREPARATION FOR DELIVERY

5.1 Unit of issue.

5.1.1 *Stock No. 6520–501–5250.*—One (1) band set, as specified in 3.2.3, constitutes one unit of issue.

5.1.1.1 *Stock No. 6520–557–9350.*—One (1) roll, as specified in 3.3.2, constitutes one unit of issue.

5.1.2 *Packaging quantities.*—The number of units of issue indicated in table III shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in procurement document.

TABLE III.—*Packaging quantities*

Federal stock number	Unit package	Intermediate package	Shipping container
6520–501–5250	1	Not required	96
6520–557–9350	1	12	288

5.1.2.1 *Packing variation permitted.*—If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.1.3 *Level A.*

5.1.3.1 *Unit package.*

5.1.3.1.1 *Stock No. 6520–501–5250.*—Each unit shall be packaged in a carton of appropriate size constructed in accordance with Federal Specification PPP–B–566 or PPP–B–676. Carton shall be provided with partitions so that each size of bands is contained in an individual compartment. Each carton shall be adequately secured to prevent accidental opening. Carton shall be further packaged in accordance with method IA-8 of Military Specification MIL–P–116.

5.1.3.1.2 *Stock No. 6520–557–9350.*—Each strip shall be neatly rolled and packaged in a carton of appropriate size constructed in accordance with Federal Specification PPP–B–566 or PPP–B–676. Carton shall be adequately secured to prevent accidental opening.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:08Z
Check the source to verify that this is the current version before use.

5.1.3.2 *Intermediate package.*

5.1.3.2.1 *Stock No. 6520–501–5250.*—Intermediate package is not required.

5.1.3.2.2 *Stock No. 6520–557–9350.*—Intermediate package shall be a carton of appropriate size and design, or a wrapping of kraft paper of suitable weight. Carton or wrapping shall be adequately secured to prevent accidental opening.

5.1.4 *Level B.* (See section 6).

5.1.4.1 *Unit package.*

5.1.4.1.1 *Stock No. 6520–501–5250.*—Each unit shall be packaged in a carton of appropriate size constructed in accordance with Federal Specification PPP–B–566 or PPP–B–676. Carton shall be provided with partitions so that each size of bands is contained in an individual compartment. Each carton shall be adequately secured to prevent accidental opening.

5.1.4.1.2 *Stock No. 6520–557–9350.*—Unit package shall be as specified in 5.1.3.1.2.

5.1.4.2 *Intermediate package.*—Intermediate package shall be as specified in 5.1.3.2.1 or 5.1.3.2.2.

5.1.5 *Level C.*—Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose within the continental limits of the United States of America.

5.2 **Packing.**

5.2.1 *Level A.*

5.2.1.1 *Exterior container.*—Exterior container shall be designed for type 2 load and constructed in accordance with Federal Specification PPP–B–585, style 3, for class 3 use; PPP–B–601, table II, using type III, class 1 plywood; PPP–B–621, table III, or MIL–B–10377, table II.

5.2.1.2 *Case liner.*—Each box shall be lined with a waterproof case liner conforming to Military Specification MIL–L–10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto.

5.2.1.3 *Strapping.*—Strapping, when required, shall be in accordance with applicable box specification.

Note.—Strapping will not be required for shipment forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 *Level B.*

5.2.2.1 *Exterior container.*—Exterior container shall conform to Federal Specification PPP–B–601, table I, using type III, class 1 plywood, or shall be designed for type 2 load and constructed in accordance with Federal Specification PPP–B–585, style 3, for class 2 use; PPP–B–621, table I, or Military Specification MIL–B–10377, table I.

5.2.3 *Level C.*—The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 **Marking.**

5.3.1 *Interior packages.*—Each unit carton, barrier bag, and intermediate package shall be marked in accordance with section 5 of Military Standard MIL–STD–129.

5.3.2 *Shipping container.*—Shipping container shall be marked in accordance with Military Standard MIL–STD–129. Service color marking and symbol are not required on container for level A packs when addressed to domestic establishments.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:08Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Federal Standard 102 should be referred to for definitions and applications of the various levels of packaging protection for supplies and equipment.

6.2 **Ordering data.** — Procurement documents should specify the following:

(*a*) Title, number, and date of this specification.

(*b*) Type required (see 1.2.1).

(*c*) Size required (see 1.2.2).

(*d*) Sizes for sets (see 3.2.3).

(*e*) Sampling required (see 4.1).

(*f*) Selection of applicable levels of packaging and packing.

(1) Level B packaging is intended to provide economical but limited protection, and should be specified only when it is determined to be appropriate. Effective period of protection is estimated to be approximately twelve months from date of initial packaging.

6.3 **Storage surveillance.**—Items packaged in accordance with level B requirements must be inspected to determine condition when not used within the time period indicated in (1). Items not used within such time period must be repackaged either in accordance with level B requirements in this specification or with level A requirements if storage beyond an additional year is anticipated.

6.4 It is believed that this specification adequately describes the characteristics necessary to secure the desired material, and that normally no samples will be necessary prior to award to determine compliance with this specification. If, for any particular purpose, samples with bids are necessary, they should be specifically asked for in the invitation for bids, and the particular purpose to be served by the bid sample should be definitely stated, the specification to apply in all other respects.

6.5 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of the specification, or which are commercially available, but is intended to cover the type, etc., normally procured to meet Federal Government requirements.

6.6 **Transportation description.** — Transportation description applicable to this item is:

Copper Articles, not otherwise indexed by name
Carload minimum weight 30,000 pounds.
Motor volume minimum weight 30,000 pounds.

6.7 This specification covers the following items appearing in the Armed Services Medical Stock List:

6520–501–5250 — Band Set, Copper, Dental, Assorted, 100's.

6520–557–9350 — Copper Strip, Dental 20 feet.

Notice. — When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility, nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise, as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

MILITARY INTERESTS:
Army—M
Navy—Md
Air Force.

★ U. S. GOVERNMENT PRINTING OFFICE: 1957

Copies of this specification may be purchased for 10 cents each.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:08Z
Check the source to verify that this is the current version before use.