A-A-51316
20 December 1985

COMMERCIAL ITEM DESCRIPTION

OPHTHALMOSCOPE, RETINOSCOPE AND TRANSILLUMINATOR SET

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a combination ophthalmoscope, retinoscope and transilluminator set for diagnosing and treating eye injuries or diseases, and measurement of refraction error. A power handle shall be used common to all three heads. A protective case and required spare bulbs shall also be furnished with this set.

Salient characteristics:

Ophthalmoscope head shall be fitted with a gas filled bulb designed to operate at not more than 3.5 volt. The bulb shall be precentered prefocused with indexing pin to provide proper bulb filament to reflecting optics alignment.

Ophthalmoscope head lens system shall consist of at least 23 lenses plus one zero power aperture in a single dial. Minimum ranges shall be from minus 1D to minus 25D, plus 1D to plus 40D. Each lens shall be indexed with a number shown on the window. A five-position selector shall provide one each red-free and yellow filter or grid line, clear slit, pin hole and large aperture.

Retinoscope head shall be streak type with mirror reflecting optics, fitted with gas filled or vacuum enhanced bulb designed to operate at not more than 3.5 volt.

Transilluminator head shall be Finoff type fitted with gas filled or vacuum enhanced bulb designed to operate at not more than 3.5 volt.

Power handle shall be able to accommodate two "C" cells and one conversion filler. As an alternate, provide a 2.5 or 3.5 volt rechargeable Ni-Chad cell with a built-in 115 volt AC recharger. All handles shall have a variable rheostat to vary the brightness and power "OFF" position lock.

Provide one spare bulb for the ophthalmoscope, one for retinoscope and two spare bulbs for the transilluminator. Bulbs shall be identical to the ones installed.

AMSC N/A FSC 6540

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Provide a soft carrying case with zipper closure and form-fitting insert to store instruments and spare bulbs. Material for the case shall be either rubber base or soft vinyl.

Two copies of instruction booklet containing information on bulb replacement, list of component parts and operating procedure shall be provided in the carrying case.

Workmanship. The ophthalmoscope shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Set (SE). One ophthalmoscope set, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

6540-00-299-8688

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS

Project No. 6540-1159

Location: CID/ophthalmoscope/ncr-attc

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.