A-A-51326
23 January 1986

COMMERCIAL ITEM DESCRIPTION

BEAKER, LABORATORY, (Polypropylene)

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-B-36268.

This commercial item description covers one type of polypropylene laboratory beaker for use in laboratory procedures.

Salient characteristics:

The beakers shall be the Griffin low form fabricated of suitable chemically resistant polypropylene material capable of withstanding 250 degrees Fahrenheit autoclave temperatures. The beakers shall be suitable for laboratory procedures. The walls of the beakers have a nominal capacity as listed. As an alternate the beaker may have a double pour out lip. The beakers shall have an accuracy of ± 5% of the total volume in the respective size. Markings shall be well defined, permanent and legible. Each beaker shall have the manufacturers name or registered trademark.

Workmanship. The beakers shall be free from defects which detract from their appearance or impair their serviceability.

Unit. One beaker, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspections. Inspections, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same shall be used for the determination of two or more inspection characteristics. Unless otherwise specified, sampling for examination shall be conducted at inspection level S-2, AQL 1.0, and sampling for tests (when applicable shall be conducted at inspection level S-1, AQL 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:54Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6640-00-982-1293	BEAKER, LABORATORY, Polypropylene, 50 ml
6640-00-982-1292	BEAKER, LABORATORY, Polypropylene, 100 ml
6640-00-982-1291	BEAKER, LABORATORY, Polypropylene, 250 ml
6640-00-982-1290	BEAKER, LABORATORY, Polypropylene, 400 ml
6640-00-982-1834	BEAKER, LABORATORY, Polypropylene, 600 ml
6640-00-982-1289	BEAKER, LABORATORY, Polypropylene, 1000 ml
6640-00-982-1288	BEAKER, LABORATORY, Polypropylene, 2000 ml

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS
FDA-MPQAS

Project No. 6640-1320

Location: CID/BEAKER, LAB
NCR disk, 7 &
S3-5-13

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.