A-A-51333A
30 July 1986

COMMERCIAL ITEM DESCRIPTION

BOTTLE, STERILE FLUID STORAGE

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers one type and two sizes of a sterile fluid storage bottle.

Salient characteristics:

The bottle shall be a cylindrical flask suitable for use in sterilizing solutions and suitable for storing them under vacuum for prolonged periods. The bottle shall be made of transparent, lead-free, heat-resistant borosilicate glass equal to Duran 8330 or Pyrex 7740. The glass shall be highly resistant to thermal shock and shall be certified to have a linear coefficient of expansion less than 33×10^{-7} cm/cm ^{o}C between 0° and 300°C. The glass shall be annealed for the bottle to withstand 45 pounds per square inch bursting pressure.

The bottle shall withstand typical autoclave solution-sterilization cycle as follows:

1. The bottle filled with 0.9 percent saline solution at ambient temperature (approximately 70°F) shall be capped and placed in chamber of autoclave.

2. The temperature of autoclave shall be raised to 260°F in not more than 10 minutes.

3. The bottle shall remain in autoclave at 260°F for not less than 30 minutes.

4. Autoclave shall be exhausted to ambient temperature and pressure in not more than 15 minutes. Door shall be opened, and bottle shall remain in chamber for an additional 15 minutes.

5. Bottle shall be removed and examined.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:54Z
Check the source to verify that this is the current version before use.

The mouth and neck of the bottle shall conform to Figure 1. The bottle shall be suitable for use with a commercial cap.

The bottle shall be available in two sizes, with white silk-screened marking, in raised letters and lines, as follows:

Size I. Five hundred milliliters capacity. The bottle shall be marked to indicate 500 ml capacity with a fill line and with horizontal lines and numbers at the 200, 300, and 400 levels for both upright and inverted positions.

Size II. One thousand milliliters capacity. The bottle shall be marked to indicate 1000 ml capacity with a fill line and with horizontal lines and numbers at the 400, 600, and 800 levels for both the upright and inverted positions and intermediate lines only at the 300, 500, 700 and 900 levels.

Both size bottles shall be marked with the manufacturer's name and part number and "CAUTION: Open upright on a flat surface only".

Workmanship. The bottle shall be free from defects which detract from its appearance or impair its serviceability.


Unit. Package (PG). One package containing ten bottles, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.



NOTE: ALL LINEAR DIMENSIONS ARE INCHES.

TOLERANCES:
FRACTIONS $\pm \frac{1}{64}$
ANGLES $\pm 2°$

FIGURE 1 MOUTH OF STERILE FLUID STORAGE BOTTLE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6530-01-233-1909 (Size I)
6530-01-232-2026 (Size II)

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

Project No. 6530-1904

YQP H1064
H1063

Location: CID/BOTTLE/H11-7

☆U.S. GOVERNMENT PRINTING OFFICE: 1986 — 605-036/44939

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:54Z
Check the source to verify that this is the current version before use.