A-A-51334
28 January 86

COMMERCIAL ITEM DESCRIPTION

LENS MEASURING INSTRUMENT, OPHTHALMIC

The General Services Adminstration has authorized the use of this commercial item description in preference to Military Specification MIL-L-36304D.

This commercial item description covers one type of lens measuring instrument for use in analyzing and marking prescription lenses. Operating voltage is 115 volt, 60 Hz, AC.

Salient characteristics:

The instrument shall be capable of determining the following optical parameters.

Parameter	Range (minimum)
Sphere power	0 to ± 20 Diopters
Cylinder power	0 to ± 20 Diopters
Prism power	0 to 5 Diopters (prism)
Cylinder axis	0 deg. to 180 deg.
Prism axis	0 deg. to 180 deg.
Multifocal segment power	0 to + 20 Diopters

The above optical parameters shall be within tolerances specified in the American National Standard Institute (ANSI) Z80.1, latest edition.

In addition, the instrument shall facilitate marking the optical center location, measurement of prism power imbalance, addition power and marginal power (with modification).

The instrument housing shall be all metal construction and the base made of cast metal. The housing shall have an adjustable inclination from zero to 60 degrees minimum. All metal surfaces shall have a baked-on enamel finish.

The spectacle holding platform, one of the component parts, shall be designed for holding spectacles or a single lens in the instrument while being measured and analyzed. The platform shall be 6 to 6-3/4 inches in length and adjustable vertically for holding lens from 30 to 85 mm diameter. Holding the spectacle to the platform shall be used with a spring loading device.

The diopter power wheel shall be conveniently located on the outside of the instrument with engraved lens power scale on the perimeter. This scale shall be in 1/8 diopter steps from zero through ± 3.00 diopters, and in 1/4 diopter steps through ± 20 diopters.

The cylinder axis wheel shall be calibrated in one degree increments from zero to 180 deg. Reading is done externally.

FSC 6540

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

AMSC N/A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

The target shall consist of two sets of three parallel lines (American-style cross-line target).

The prism measuring system shall consist of a holder, scale and a reticle. The prism scale shall be calibrated in 5 deg. increments, from zero through 180 deg. The lens reticle shall be designed with a base-apex line and prism power rings, or other calibrations, indicating up to five prism diopters.

The lens marking device shall consist of three marking pens and an ink reservoir. The pens shall be capable of marking the optical center and the 180°axis of the lens in the same operation.

The light source shall provide sufficient illumination for proper operation of the instrument. Low light illumination shall be achieved by using dark filter or two-position switch. One power on-off switch shall be conveniently located on the instrument.

The eye piece shall have an external reference scale which allows different individuals to reset quickly without visual refocusing.

The electric cord shall be six feet nominal with 3-prong plug to conform to ANSI C73.11.

Provide two spare lamps and one spare ink pad with each unit.

Service data. Provide two copies each of a technical manual which contains complete instructions for installation, operation and maintenance, and a list of component parts.

Workmanship. The instrument shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One lens measuring instrument with accessories as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6540-00-181-8037

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6540-1171

Location S3-9-16

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.