A-A-51336
28 January 1986

COMMERCIAL ITEM DESCRIPTION

TIP, PIPET, Plastic, Disposable

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a plastic, disposable, pipet tip for use with the Medical Laboratory Automation, Inc., Model Electra 750, Plasma Coagulation Timer Unit (NSN 6630-01-126-9960).

Salient characteristics:

Shall be a plastic, disposable, pipet tip suitable for dispensing precise volumes of coagulation reagents when attached to the precision pipets furnished with the Plasma Coagultion Timer Unit (NSN 6630-01-126-9960).

The pipet tip shall fit the incubator well of the Plasma Coagulation Timer Unit (6630-01-126-9960).

The pipet tip shall be made of polypropylene material having non-wetting characteristics to assure complete dispensing of liquid reagents, and an optical clarity that is sufficient for visually checking the volumetric level of a liquid that has been drawn into the bore of the tip.

The volumetric capacity of the pipet tip shall be 200 microliters, designed so that when filled with liquid and positioned vertically along the diametral axis, the liquid does not make contact with the nozzle of the pipet.

Workmanship. The pipet tip shall be free from defects which detract from its appearance, or impair its serviceability.

Unit. Package (PG). One package containing 1000 tips, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

AMSC N/A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspection and tests. Sampling for inspection and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one package.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

For information only. The following product, as of the date of this document, is purported by the following company to meet the salient characteristics of the item described herein.

Medical Laboratory Automation, Inc. Catalog No. 9025/9620

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

The following Federal Stock Number is covered by this document:

6640-00-502-8231

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

uCustodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

51814

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6640-1350
Loc. CID, S3 8-7

Source: https://assist.dla.mil Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.