A-A-51338
27 January 1986

COMMERCIAL ITEM DESCRIPTION

COMPARATOR SET, COLOR
Residual Chlorine and pH Determination

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a comparator set suitable for use in performing pH and residual (free) chlorine determinations of water in the field.

Salient characteristics:

Shall be a comparator set with two color comparators, for use in performing pH and residual (free) chlorine determinations of water in the field.

The comparator set shall consist of the following components:

1. One color comparator for pH determinations in the range of 3 to 10 pH, in graduations of 3, 4, 5, 6, 7, 8, 9 and 10 pH.

2. One color comparator for residual (free) chlorine determinations by the DPD method in the range of 0.2 to 10.0 parts per million (ppm), in graduations of 0.2, 0.4, 0.6, 0.8, 1.0, 2.0, 5.0 and 10.0 ppm.

3. Two (2) 5 milliliter (ml) calibrated glass sample cells with caps, for use with the color comparator for pH determinations.

4. Two (2) 10 ml calibrated glass sample cells with caps, for use with the color comparator for residual (free) chlorine determination.

5. Twenty-five (25) ml of wide range pH reagent.

6. One-hundred (100) foil wrapped DPD No. 1 tablets.

7. Instructions for using the comparators.

8. A carrying case.

Each color comparator shall be constructed with eight (8) permanent color standards mounted in its housing, and in such a manner that a sample cell can be inserted in either of two openings in the housing, for comparison with four color standards at once.

Each color comparator housing shall be fabricated of impact resistant plastic, and constructed to include a light diffusion filter and permanent markings that identify the value of each color standard.

Workmanship. The comparator set shall be free from defects which detract from its appearance or impair its serviceability.

FSC 6630

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

AMSC N/A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Unit. Set (SE). One comparator set, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one comparator set.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of th contract.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document:

6630-01-151-6417

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVILIAN AGENCY COORDINATING ACTIVITY:

VA-OSS

51573

Project No. 6630-0423

Location: S3-7-9

U.S. GOVERNMENT PRINTING OFFICE: 1986 - 605-036/44572

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.