A-A-51339
22 April 1986

COMMERCIAL ITEM DESCRIPTION

BEAKER-FLASK, LABORATORY, (Borosilicate Glass)

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-B-37484.

This commercial item description covers one type of borosilicate glass laboratory beaker-flask for use in laboratory procedures.

Salient characteristics:

Each size beaker shall be uniformally cylindrical with a spout capable of being made liquid-air tight by means of a cap. The beaker shall be fabricated from borosilicate glass conforming to Type 1, Class A of ASTM E -438. The beakers of each respective size shall have a well defined pour out lip and be suitable for laboratory procedures. The beakers shall have an accuracy of ± 5% of the total volume in the respective size. Markings shall be permanent, legible and withstand common laboratory use. Each beaker shall have the nominal dimensions and graduations listed in Table I. Glassware shall conform to ASTM Standard E-671. The cap shall be fabricated of black polypropylene with a neutral (or resin cured butyl) rubber insert, in such a way to effect the liquid-air type seal. The cap shall withstand laboratory procedures and current cleaning methods. The cap shall fit each respective size beaker such that it does not afford a disproportionate appearance or affect the capability of the seal.

TABLE I Dimensions

Size (ml)	Dimensions* Nominal Height inches (cm)	O.D. Base inches (cm)	O.D. Mouth inches (cm)	Graduations Range (ml)	Intervals of markings (ml)	Cap O.D. inches (cm)	Unit of Issue package
150	3.1 (7.87)	2.75 (6.98)	2.38 (6.04)	40 to 120	20	2.75 (6.98)	10
300	4.9 (12.45)	2.75 (6.98)	2.38 (6.05)	50 to 250	50	2.75 (6.98)	10

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

TABLE I Dimensions (contined)

Size: (ml)	Dimensions* Nominal Height: inches (cm)	O.D. Base: inches (cm)	O.D. Mouth: inches (cm)	Graduations Range: (ml)	Intervals of markings: (ml)	Cap O.D.: inches (cm)	Unit of Issue package
500	7.0 (17.78)	2.75 (6.98)	2.38 (6.05)	50 to 450	50	2.75 (6.98)	10
800	7.5 (19.05)	3.75 (9.53)	2.84 (7.22)	50 to 800	50	3.75 (9.53)	10
1000	8.75 (22.23)	3.75 (9.53)	2.84 (7.22)	50 to1000	50	3.75 (9.53)	10
1200	10.0 (25.4)	3.75 (9.53)	2.84 (7.22)	50 to1200	50	3.75 (9.53)	10

*All dimensions shall be ± 1/16 inch.

Workmanship. The beakers shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing the number of beakers with caps as specified in Table I for the stock number specified constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspections. Inspections, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. Unless otherwise specified, sampling for examination shall be conducted at inspection level S-2, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided or under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Numbers are covered by this document:

6640-01-013-3553	BEAKER-FLASK, LABORATORY, 150 ml.
6640-01-013-5280	BEAKER-FLASK, LABORATORY, 300 ml.
6640-01-013-3551	BEAKER-FLASK, LABORATORY, 500 ml.
6640-01-013-3555	BEAKER-FLASK, LABORATORY, 800 ml.
6640-01-013-3552	BEAKER-FLASK, LABORATORY, 1000 ml.
6640-01-013-3554	BEAKER-FLASK, LABORATORY, 1200 ml.

Ordering data: (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

Project No. 6640-1323

Location: CID-beaker-flask-3553

U.S. GOVERNMENT PRINTING OFFICE: 1986 - 605-036/45208

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.