A-A-51342
31 January 1986

COMMERCIAL ITEM DESCRIPTION

SHAKING MACHINE, LABORATORY, (Blood Collecting Tube)

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-S-37453.

This commercial item description covers a mechanical shaking machine, suitable for use in thoroughly mixing, by rotation, blood samples contained in blood collecting tubes, in medical laboratories.

Salient characteristics:

The shaking machine shall consist of a housing which contains the motor, a rotating wheel, and a drive shaft to connect the motor to the wheel.

The motor housing sides and top shall be fabricated of a suitable sheet metal with openings to provide adequate ventilation for the motor during operation. The sides shall be secured to the top and base to form a stable structure for the shaker with at least one side and the top removable to allow access to the motor for servicing. The front may be straight for a vertically mounted wheel, or slanted inward to provide for an angular mounting of the wheel. The base shall provide a rigid mounting surface for the motor, and shall have rubber feet secured to each of the four external corners to afford minimum vibration and prevent creepage during operation.

The wheel shall be fabricated of a suitable sheet metal or plastic, and shall be mounted on the motor drive shaft at the front of the motor housing. The wheel shall be designed to securely hold 12 evenly spaced, blood collecting tubes with a diameter range of 12 to 16 mm, and lengths that include 125 mm.

The wheel shall be turned at a speed between 15 and 24 revolutions per minute by a fractional horsepower electric motor which requires no lubrication. The motor shall operate on 115 volts, 60 Hz, AC and shall be controlled by an "on-off" toggle switch on the motor housing. The attached power cord shall be three conductor with a three prong male plug conforming to ANSI C73.11. One conductor and one prong shall provide a ground connection. The external power cord shall be not less than 5.75 feet in length.

The shaking machine shall be supplied with complete instructions for use and maintenance, and shall be UL approved.

Finish. The shaking machine shall be fabricated of corrosion-resistant materials or shall be protected by appropriate plating or painting applied in accordance with standard commercial methods and practices.

Identification marking. A nameplate shall be permanently attached to the shaking machine legibly marked with the electrical characteristics and name or registered trademark of the manufacturer.

FSC 6640

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Workmanship. The shaking machine shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One shaking machine, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspection and tests. Sampling for inspection and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one shaking machine.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

For information only. The following product, as of the date of this document, is purported by the following company to meet the salient characteristics of the item described herein:

Angular Wheel	Sepco	Cat. No. 60448 Tube Rotator
Vertical Wheel	Drummond Scientific	Blood Machine

The following National Stock Number is covered by this document:

6640-00-982-4243

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6640-1354

Location: S3-8-19

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.