A-A-51343
14 February 1986

COMMERCIAL ITEM DESCRIPTION

SURGICAL PACK, LOWER EXTREMITY

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers one type of surgical pack, disposable, lower extremity, nonwoven material, sterile, individually sealed.

Salient characteristics:

Shall be a sterile, disposable, surgical drape pack fabricated of nonwoven material for creating a sterile field specifically for lower extremity procedures during surgery. The pack shall consist of the following components with the approximate dimensions in inches:

a. One overwrap/tablecover, 50 by 90 inches (127 by 228.6 cm).

b. One overwrapped scrub nurse gown and hand towel.

c. Two surgical gowns, with hand towels.

d. One Mayo stand cover with pad, 23 by 58 inches (58.4 by 147.3 cm).

e. Two utility drapes, 20 by 29 inches (50.8 by 73.6 cm).

f. One absorbent hand towel, 19 by 24 inches (48.2 by 60.96 cm).

g. One drape sheet 41 by 57 inches (104.1 by 144.78 cm).

h. One impervious stockinette.

i. One split sheet, 72 by 84 inches (182.8 by 213.3 cm) with adhesive strip and fabric reinforcement.

j. One drape, 72 by 100 inches (182.8 by 254 cm), with adhesive strip and fabric reinforcement.

k. One incise drape, 13 by 16 inches (33 by 40.6 cm).

The overwrap/tablecover shall be fan-folded for ease in opening on the surgical back table. Shall be impervious to fluids and moisture.

FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

AMSC N/A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

The gowns shall be of large or extra-large size. One of the towels shall be individually wrapped in a memory-free overwrap with a hand towel. All gowns shall:

a. Be soft, comfortable, pliable, quiet and memory-free.

b. Have a high breathability facilitating comfort when wearing for long periods of time.

c. Be fluid repellent.

d. Have stockinette cuffs that shall conform to wrists.

e. Be large enough to cover the back of the individual gowned.

f. Have a tie closure that shall not rip or separate when tying. Shall be designed to minimize the possibility of contamination when tying.

The Mayo tray and stand cover shall be made of a plastic material at least 2.75 mils thick and be impervious to fluids. Shall be fan-folded with a 6 inch (15.2 cm) cuff on the distal end for sterile draping of the Mayo stand. The pad shall be large enough to cover the surface of the Mayo tray.

The utility drapes shall be water repellent.

The towels shall be highly absorbent, to sufficiently dry arms after a surgical scrub.

The 41 by 57 inch (104.1 by 144.7 cm) drape sheet shall be water repellent. Shall be cuffed and fan-folded for aseptic draping.

Shall have a split sheet, approximately 72 by 84 inches (182.8 by 213.3 cm) with a 4 inch by 20 inch (10.16 by 50.8 cm) split, with a pressure sensitive adhesive strip, and a fabric reinforcement area around the split approximately 36 by 18-1/2 inches (91.4 by 46.9 cm). The reinforcement area shall be impervious to fluids yet highly absorbent, eliminating strike-through while preventing pooling of fluids. Shall be cuffed and fan-folded to permit aseptic draping of the operative area.

Shall have a body drape, approximately 72 by 100 inches (182.8 by 254 cm) with a pressure sensitive adhesive strip and a fabric reinforcement area approximately 36 by 18-1/2 inches (91.4 by 46.9 cm). The reinforcement area shall be impervious to fluids yet highly absorbent, eliminating strike-through while preventing pooling of fluids. Shall be cuffed and fan-folded to permit aseptic draping of the operative area.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:54Z
Check the source to verify that this is the current version before use.

Shall have a transparent adhesive backed incise drape, approximately 13 by 16 inches (33 by 40.6 cm). Shall have a non-adhesive tab on each side of the length to aid in application.

Shall have a large size impervious stockinette, approximately 9 by 48 inches (22.8 by 121.9 cm). Shall be rolled to permit aseptic application.

All items except the Mayo cover and stockinette, shall be fabricated of a soft nonwoven fabric, with minimal memory for good drapeability. Shall resist tearing, pilling and linting.

"Directions for Proper Placement" shall be clearly imprinted on each drape.

If the pack is inadvertently opened and not used, shall be capable of sterilization by steam or ethylene oxide.

The pack shall be sterile and individually wrapped in a suitable tear resistant, heat sealed package which will allow aseptic handling and maintain sterility of the contents until opened or damaged.

Each pack shall also be wrapped in a heat sealed exterior dust cover.

All pack components shall be clearly listed on the exterior wrapper.

Workmanship. The surgical pack shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing three (3) individually packaged surgical packs, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:54Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents therof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one package.

NOTE: The following National Stock Number is covered by this document:

6530-01-075-6660

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PGC 59603

Project No. 6530-1819

Location H1-17-21

Source: https://assist.dla.mil --Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.