A-A-51345
7 FEBRUARY 1986

COMMERCIAL ITEM DESCRIPTION

LARYNGEAL STENTS, Silicone Rubber

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers molded rubber stents of value in the surgical repair of laryngeal fracture, laryngeal stenosis and subglottic stenosis.

Salient characteristics:

The stents shall be made of silicone elastomer and molded to conform to the inner contour of the larynx. The stents shall be made of a precise hardness to give support without injuring surrounding tissues. Skin or mucosal grafts may be sutured to the stent. A heavy silk suture shall be molded into the stent to form a loop at the top and facilitate removal.

The stent shall be supplied with two silicone rubber suture buttons 12 mm in diameter and 2 mm thick. The dimensions for the stents are as follows:

Maximum Width and Length		Minimum Width at Bottom
1.8 cm	3.8	1 cm
3 cm	4.8	1.5 cm

The stents and buttons shall be able to withstand repeated steam sterilization of 276°F (135.5°C). The stents and buttons shall comply with USP Class IV plastic requirements for biological study. All elastomer surfaces shall be smooth, glossy and free from burrs, feathers or other surface defects.

Workmanship. The stents shall be free from defects which detract from their appearance or impair their serviceability.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

MSC N/A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Unit. Each (EA). One stent, along with buttons, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one laryngeal stent.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:54Z
Check the source to verify that this is the currentversion before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-140-5115	STENT, LARYNGEAL, Female, Silicone Rubber, 3.8 cm long, with two buttons
6515-01-140-5116	STENT, LARYNGEAL, Male, Silicone Rubber, 4.8 cm long, with two buttons

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-ME

Agent:

DLA-DM

CIVIL COORDINATING AGENCY:

VA - OSS

PGC 06827

Project No. 6515-3490

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.