A-A-51348
16 June 1986

COMMERCIAL ITEM DESCRIPTION

CONNECTOR, ELASTIC TUBING, BRANCHED, PLASTIC, T-SHAPED

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers T-shaped connectors suitable for use with flexible tubing in laboratory and surgical procedures.

Salient characteristics:

The connector shall be fabricated of a rigid transparent or translucent plastic and meet the requirements of the current USP for Class II plastic. The connector shall have no more than 5 total parts per million extractable barium, calcium and lead. The connector shall be capable of autoclaving at 275°F for 20 minutes without distortion or embrittlement.

The connector shall be shaped in the configuration of a "T" and shall provide leakproof connection of flexible tubing sized 3/16, 1/4, 5/16, 3/8, and 7/16 inch inside diameter (I.D.) Each leg of the connector shall be uniformly, stepwise tapered from 1/4 inch outside diameter (O.D.) at the end to 5/8 inch maxiumum o.d. at the center of the T. The outside diameter of the steps shall be a maxiumum 1/8 inch larger than the tubing to be attached. The length of each leg as measured from the center of the T to each end shall be 2 $\pm$ 1/2 inches. Wall thickness shall be 0.06 $\pm$ 0.02 inches.

Workmanship. The connector shall be examined for symmetry of the legs. Flow channels shall be examined for smoothness and obstructions to fluid flow. The connector shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing fifteen connectors, as specified, constitutes one unit.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted in accordance with the USP for Class II plastics. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be S-2 with an AQL of 1.0, in accordance with MIL-STD-105. The unit of unit of product for test purposes shall be one connector.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:54Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-857-2267	CONNECTOR, ELASTIC, TUBING, BRANCHED, Plastic, T-Shaped, 9/16 Inch Outside Diameter, 10s

Ordering data (Intermediate/exterior package quantities and labeliing and marking must be specified in the contract and/or order)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

CIVILIAN COORDINATING ACTIVITIES:

VA-OSS
PHS

Preparing Activity:

DoD-MB

Agent:

DLA-DM

PGC 05636

Project No. 6515-3485

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.