INCH-POUND

A-A-51348A
20 March 1992
SUPERSEDING
A-A-51348
16 June 1986

COMMERCIAL ITEM DESCRIPTION

CONNECTOR, ELASTIC TUBING, BRANCHED, Plastic, T-shaped

The General Services Administrations has authorized the use of this commercial item description.

This commercial item description covers T-shaped connectors suitable for use with flexible tubing in laboratory and surgical procedures.

Salient Characteristics.

The connector shall be fabricated of a rigid transparent or translucent plastic and meet the requirements of the current USP Class II plastic. The connector shall have no more than 5 total parts per million extractable barium, calcium and lead. The connector shall be capable of autoclaving at 250°F for 20 minutes without distortion of embrittlement.

The connector shall be shaped in the configuration of a "T" and shall provide leakproof connection of flexible tubing sized 3/16, 1/4, 5/16, 3/8, and 7/16 inch inside diameter. Each leg of the connector shall uniformly, stepwise tapered from 1/4 inch outside diameter at the end to 5/8 inch maximum outside diameter at the center of the T. The outside diameter of the steps shall be a maximum 1/8 inch larger than the tubing to be attached. The length of each leg as measured from the center of the T to each end shall be 2 ± 1/2 inches. Wall thickness shall be 0.06 ± 0.02 inch.

Certificate of biocompatibility. A certificate of compliance, stipulating that offered items meet the biocompatibility test, shall be acceptable, provided that the contractor has supplied acceptable items to the Government within the past three years and that there have

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

been no changes to material formulation and the method of processing since actual testing was performed. The certificate shall include the date the testing was performed, the test method and results, the name and title of the person evaluating the results, the current contract number and National Stock Number (NSN), a statement certifying that material processing and formulation have not changed since the time that biocompatibility was performed and the signature and title of the person making the certification.

Workmanship. The connector shall be examined for symmetry of the legs. Flow channels shall be examined for smoothness and obstructions to fluid flow. The connector shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing fifteen connectors, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted in accordance with the USP for Class II plastics. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Sampling for tests. Sampling for test shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one connector. Sampling shall be inspection level S-2 with an Acceptable Quality Level (AQL) of 1.0.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-234-6823	CONNECTOR, ELASTIC, TUBING, BRANCHED, Plastic, T-shaped, 9/16 Inch Outside Diameter, 15s

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5332
Location: word\nasa\A-A-51348A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-51348A	20 March 1992

3. DOCUMENT TITLE
CONNECTOR, ELASTIC TUBING, BRANCHED, Plastic, T-shaped

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS *(Include Zip Code)*	**d. TELEPHONE** *(Include Area Code)* **(1) Commercial** **(2) AUTOVON** *(If applicable)*
	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code)
Defense Personnel Support Center ATTN: DPSC-MSE	(1) Commercial: (215) 737-2870 (2) AUTOVON: 444-2870
c. ADDRESS *(Include Zip Code)* 2800 South 20th Street Philadelphia, PA 19145	**IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT** Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — *Previous editions are obsolete*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.