A-A-51350
21 February 1986

COMMERCIAL ITEM DESCRIPTION

SHIELD, GONADAL, X-RAY PROTECTIVE

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers an adjustable shadow-casting shield for patients during radiological examination in military medical treatment facilities.

Salient characteristics:

Shall be an adjustable protective device for shielding the gonads of male and female patient during radiological examinations.

When properly aligned in the primary x-ray beam and associated beam-defining light, the device shall permit the casting of a shadow over the patient's gonads.

Shall be either capable of permanent attachment to a suitable component of the x-ray unit or provide an adjustable mechanism for attachment to all diameters of tube end housings.

The thickness of material shall have the equivalent of at least 0.5 mm of lead.

Shall meet applicable standards and requirements of the Bureau of Radiological Health.

Workmanship. The x-ray protective shield shall be free from defects which detract from its appearance of impair its serviceability.

Unit. Each (EA). One x-ray protective shield, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

AMSC N/A FSC 6525

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6525-01-037-5182

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PGC 49321

Project No. 6525-1188

Location: CID/SHI-GON-XRAY-PROTECTIVE/H13

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.