A-A-51351
28 February 1986

COMMERCIAL ITEM DESCRIPTION

CENTRIFUGE, LABORATORY (Microhematocrit, Battery Powered)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a miniaturized microhematocrit centrifuge that is battery powered and is also operable from an AC supply line by use of an adapter.

Salient characteristics:

Shall be a miniaturized 9 volt (DC) powered centrifuge for hematocrit determinations.

Shall be supplied with a six (6) place rotor that accommodates heparinized microcapillary tubes 32 millimeter (mm) in length and graduated to contain 9 microliters of a blood sample. Shall also be supplied with a direct reading plate for hematocrit determinations for each microcapillary tube slot. The direct reading plate shall include a hematocrit measurement range capability of 10 to 80 percent.

Shall be internally powered by six (6) size C manganese-alkali battery cells (not required to be furnished with the centrifuge). Shall also be operable from a 115 volt, 50/60 Hz AC supply line by use of an adapter (not required to be furnished with the centrifuge).

Shall be operable in an ambient temperature range of 0 to 40 degrees (C). Shall also be capable of proper operation at the same ambient temperature range, after storage in an ambient temperature range of -20 to +55 degrees (C).

Shall be designed to have an operating speed of 11,500 rpm ±100 rpm, with a programmed centrifugation time of at least 3 minutes and 20 seconds. The centrifugal force (RCF) at 11,500 rpm shall be 5396 g.

The housing shall include a plastic cover that has sufficient optical clarity so that the hematocrit readings can be made without the lifting or removal of the cover.

MSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

The overall nominal dimensions of the centrifuge shall be 3-3/8 width x2-5/16 height x8-3/16 length (inches). The centrifuge housing shall include: a compartment for the internal battery power supply, an internal control circuitry for battery checks, an on-off switch, and a receptacle for connecting the AC adapter.

The centrifuge without batteries shall weigh approximately 15 ounces.

When connected to a suitable power source (batteries or AC adapter) the centrifuge shall operate without any noisy vibrations and shall automatically stop when the centrifugation time is complete.

Workmanship. The centrifuge shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One centrifuge, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Inspection and Tests. Inspections and tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Operational Test. Each centrifuge shall be tested to demonstrate its capability to operate properly.

Centrifugation Time. The centrifuge shall be connected to a suitable electrical source. The centrifuge shall be operated by turning the switch to the on position, and the time of programmed centrifugation shall be determined using a suitable timer. The sampling for testing shall be conducted in accordance with MIL-STD-105 and an inspection level of S-1. Any centrifuge that does not operate for at least 3 minutes and 20 seconds before stopping shall be cause for rejection. The acceptance number shall be zero.

Rotor Speed. The centrifuge shall be connected to a suitable electrical source. The centrifuge shall be operated and the rotor speed shall be determined by use of a strobic light or other suitable means. The sampling for testing shall be conducted in accordance with MIL-STD-105 and an inspection level of S-1. Any centrifuge whose rotor speed does not operate within a range of 11,400 to 11,600 rpm shall be cause for rejection. The acceptance number shall be zero.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document:

6640-01-205-2422

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MP

Agent

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

Project No. 6640-1333

CID-cen-lab-2422

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-ATTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51351	CENTRIFUGE, LABORATORY (Microhematocrit, Battery Powered)

3a. NAME OF SUBMITTING ORGANIZATION

b. ADDRESS *(Street, City, State, ZIP Code)*

4 TYPE OF ORGANIZATION *(Mark one)*

- ☐ VENDOR
- ☐ USER
- ☐ MANUFACTURER
- ☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.