A-A-51354
14 March 1986

COMMERCIAL ITEM DESCRIPTION

BEDPAN

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a disposable bedpan.

Salient characteristics:

The bedpan shall be disposable, plastic, and for single patient use, capable of supporting a 250 pound patient as evidenced by the ability to withstand, without damage, a 250 pound load, applied at any point on the pan, by a weight in contact with the bedpan along a contact area not to exceed 11 inches (280 cm) in diameter on a flat surface. The bedpan shall withstand this force without cracking or permanent indentation.

The bedpan shall be oval shaped with side pontoons only that permit cleaning in all commercial washers.

The bedpan shall be free of any cover, handle, sharp edges or burrs. Length shall be 14-1/2 ± 1/4 inch (36.83 ± .64 cm). Width shalll be 11-1/2 inch (29.21 cm). Front depth shall be between 3-1/4 and 4-1/4 inches (8.26 to 10.80 cm). Rear depth shall be 2-1/2 + 1/4 -0 inches (6.35 + .64 -.0 cm). Side depth shall be 2-1/2 inches (6.35 cm). Unless otherwise specified, the tolerance on the above dimensions shall be ± 1/2 inch (+ 1.27 cm).

The bedpan shall be capable of holding 2000 ml when inclined at any angle in any direction up to 10 degrees without any spillage.

The design shall provide stability, balance, and relative comfort during use.

Twenty bedpans shall be supplied in one package.

Workmanship. The disposable bedpan shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package of 20, as specified, constitutes one unit.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one bedpan.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-01-049-0428

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING AGENCY:

VA-OSS

Project No. 6530-1815

Location H1-17-7

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:54Z
Check the source to verify that this is the current version before use.