A-A-51359
12 March 1986

COMMERCIAL ITEM DESCRIPTION

GENERATOR, CARBON DIOXIDE-HYDROGEN, ANAEROBIC CULTURE APPARATUS

This commercial item description shall be used in preference to applicable portion of Military Specification MIL-A-37234.

This commercial item description covers a disposable carbon dioxide-hydrogen generator which produces an environment of anaerobic mircoorganisms.

Salient characteristics:

The generator shall consist of a flexible laminated plastic-aluminum foil envelope containing a hydrogen (H_2) producing tablet, a carbon dioxide (CO_2) producing tablet and a perfusion paper membrane to control the activation of the tablets.

Each generator shall contain a single charge of materials permitting controlled generation of hydrogen and carbon dioxide gas sufficient to produce an anaerobic atmosphere for the duration of the incubation cycle when used in conjunction with anaerobic apparatus. The envelope shall protect the contents from deterioration until opened for use. Direction for use shall be provided on the generator envelope.

The generator shall be suitable for use with anaerobic culture apparatus.

The generator shall be free of delamination, crushing, holes, tears, cracks or other damage. Tablets shall be free of crushing, cracks, powder or discoloration.

Workmanship. The generator envelope shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 100 generators, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Shelf life. The generator shall have at least a twenty-one month shelf life remaining at the time of delivery to the Government.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered byy this document has been deetermined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State,, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

For information only. The following items listed are used in conjunction with the carbon dioxide-hydrogen generator.

6640-00-001-6670	Lid, Anaerobic Culture Apparatus
6640-00-145-1150	Indicator, Oxygen, Anaerobic Culture Apparatus
6640-00-926-7655	Jar, Anaerobic Culture Apparatus

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

The following National Stock Number is covered by this document:

6640-00-935-6863	GENERATOR, CARBON DIOXIDE-HYDROGEN, ANAEROBIC CULTURE APPARATUS

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

56139

Project No. 6640-1177

Location: CID S3-1-44

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.