INCH-POUND

A-A-51361A
17 May 1992
SUPERSEDING
A-A-51361
14 March 1986

COMMERCIAL ITEM DESCRIPTION

SURGICAL PACK, GOWN AND TOWEL

The General Services Administration has authorized the use of this Commercial Item Description

This Commercial Item Description covers a sterile, disposable, operating gown and towel.

Salient Characteristics:

The gown and towel surgical pack shall be a disposable, non-woven, sterile surgical operating gown with towel, suitable for preventing bacterial migration into the sterile field. The surgical pack shall be supplied in the following sizes:

Size 1 - Large gown with towel.

Size 2 - Extra large gown with towel.

The gown shall conform to the following minimum dimensions in inches:

	Dimension	
Location	Size 1	Size 2
Chest	55	62
Front length	44	44-1/2
Sweep	60	65
Sleeve length (to top of cuff)	24-3/4	24-3/4
Belt length	13	13

AMSC N/A FSC 6532

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

The surgical operating gown shall be full length with a generous overlapping back. The gown shall have full length raglan sleeves set in by sewing. The sleeves shall have wristlets made of tubular knit polyester fabric. As an alternate, the gown may have set-in sleeves in lieu of raglan sleeves. The gown shall have four tie tapes, two on the outside, and two on the inside, approximately at waist level. The tie tapes shall be attached by sewing, or glueing.

The gown shall have a non-woven neck binding sewn to the neck area of the gown body. Plastic snap fasteners shall be at the neck area. These fasteners shall be applied to a non-woven ribbon made from the gown body material. The ribbon shall be attached to the gown by sewing, or by gluing. Three male snaps shall be on the left flap and two female snaps shall be on the right flap. As an alternate, the gown may have hook and pile closures (velcro) in lieu of the snaps.

The gown shall be fabricated from a soft, pliable, fluid resistant, non-woven material. The material shall have good resistance to dry-particle penetration. The material shall have high breathability, and shall be comfortable to wear for long periods. The color of the material shall be light pastel shades of blue, green, or blue-green.

Each gown shall be supplied with a scrim reinforced absorbent towel. The towel shall have minimum dimensions of 15 by 19 inches.

Each gown and towel shall be folded and overwrapped in a white or blue plastic, or non-woven fabric. The individually folded and wrapped gown and towel shall be packaged in a white Kraft, or plastic, sealed bag which is tear resistant and which will allow aseptic handling and maintain sterility of the contents until opened or damaged.

Each surgical pack shall be sterile in accordance with the USP or AAMI requirements. Sterility assurance level shall be 0.000001.

Workmanship. The gown and towel surgical pack shall be free from defects which detract from its appearance or impair its serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Unit. Package (PG).

For size 1. One package containing twenty-six or twenty-eight individually packaged sterile surgical packs (each pack containing a gown and towel), as specified, constitutes one unit.

For size 2. One package containing twenty-six individually packaged sterile surgical packs (each pack containing a gown and towel), as specified, constitutes one unit.

The individual package shall be a sealed, peel-open type container capable of maintaining sterility of contents unless the package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 and an inspection level of II. The unit of product for examination purposes shall be one gown and towel surgical pack.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 4.0 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one gown and towel surgical pack.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document.

6532-00-405-6009 (Size 1)
6532-01-018-3286 (Size 2)

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA - OSS

Project Number: 6532-1224

Location: PACKC13B.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-51361A	17 May 1992

3. DOCUMENT TITLE
SURGICAL PACK, GOWN AND TOWEL

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 96790

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.