A-A-51366
31 March 1986

COMMERCIAL ITEM DESCRIPTION

SUCTION APPARATUS, SURGICAL

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-S-37014.

This commercial item description covers a mobile high volume, low pressure pleural cavity pump, suitable for maintaining negativity in the pleural space by rapid removal of air during post-operative procedures.

Salient characteristics.

The suction apparatus shall be a mobile high volume, low pressure pleural cavity pump suitable for maintaining negativity in the pleural space by rapid removal of air during post-operative procedures.

The suction apparatus shall consist of a cabinet containing an electric motor and air pump, a variable transformer or solid state speed control and pressure gage, suitably interconnected and mounted on a mobile stand.

The pump shall have sufficient capacity to remove 50 liters of air per minute through the system when set at minus 10 centimeters of water (-10 cm, H_2O), and 70 liters of air per minute when set at -20 cm H_2O. Range of available negative pressures shall extend from zero up to 60 cm H_2O.

The selection of desired negative pressure shall be made by adjusting the variable transformer or solid state speed control supplied with the unit.

The suction apparatus shall have an aneroid gage to verify pump pressures at all times during operation.

The negative pressure system shall be of the type that protects the patient against dangerous build-up in the event tubing becomes obstructed.

Input voltage requirement should be 115 volt, 50/60 Hz. The electrical line cord shall be a 3 conductor type with 3 prong male connector conforming to NFPA 99, Standard for Health Care Facilities. The system shall be supplied with a pilot light to indicate motor operation

The mobile stand shall be primarily constructed of corrosion-resisting metal with rubber casters having a minimum wheel outside diameter of 3 inches.

The mobile stand shall be equipped with a shelf or tray capable of holding two trap bottles and have space available for one additional bottle.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

The trap bottles shall be one gallon (3.8 liters) or 3/4 gallon (2.8 liters). The primary trap bottle shall be provided with a graduated scale to allow visual reading of the approximate volume of drainage.

There shall be two 5/8" connectors in the primary bottle cap assembly to which the drainage tubes from the chest catheters may be joined and to which the tubes inside the bottle may be joined. These connections shall provide an unbroken drainage system from the pleural space to a protective underwater seal. A third connector in the cap assembly shall join the primary and secondary bottles.

The secondary cap assembly shall provide a connector for the primary and secondary bottles and the suction pump.

The connectors shall have a nominal 5/8 inch inside diameter. Tubing (from bottle to bottle and bottle to pump) shall have nominal dimensions of 7/8 inch (1.90 cm) inside diameter, wall thickness of 1/8 inch and a durometer of approximately 80. Drainage tube connectors (2) shall be 5/16 inch.

Drainage tubes from the patient chest catheters to the primary bottle and the bottle tubes shall have nominal dimensions of 3/8 inch (0.95 cm) inside diameter, wall thickness of 1/16 inch (0.15 cm) and a durometer of approximately 80

Drainage tubes to the chest catheters shall be approximately 4 feet (1.20 m) long and the tubes inside the primary-bottle approximately 8-1/2 inches (24.13 cm) in length. All tubes shall be of a clear polyvinyl plastic.

The complete unit shall be approximately 33 inches (83.82 cm) high by 20 inches (50.80 cm) wide by 18 inches (45.72 cm) deep. Net weight shall be approximately 50 pounds (22.50 kg).

The suction apparatus shall be supplied with service data containing complete instructions for operation, maintenance and trouble shooting.

Workmanship. The suction apparatus shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One suction apparatus, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document

6515-00-168-6832

Ordering data. Intermediate exterior package quantities and labeling and marking must be specified in the contract and or order.

MILITARY INTERESTS.

Custodians.

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITY.

VA-OSS

PREPARING ACTIVITY

DoD-MB

Agent.

DLA-DM

44773

Project No. 6515-2581

Location CID Hi-2-30

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.