COMMERCIAL ITEM DESCRIPTION

DEMINERALIZER, WATER, ION EXCHANGE

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-D-37079.

This Commercial Item Description covers an ion exchange water demineralize used for producing Up to 5 gallons (18.93 liters) an hour of demineralized water by gravity flow. Shall be Crystalab Inc. Model CL-5 or equal, having the following salient characteristics.

Salient characteristics:

The water demineralize shall be a table model for use in a laboratory and shall consist of the following:

1. A shatterproof and transparent plastic tower.
2. A base housing.
3. A direct reading quality check [salinity indicating) meter.
4. Outlet tubing with shut-off clamp.
5. Power cord.

The base shall support the tower, and shall house the quality check meter and electrical circuitry. The assembled water demineralize shall have an overall height of 17-3/4 inches (45.09 cm.) nominal and a maximum weight of five (5) pounds (2.25 kg.). The water demineralize shall be free of wobble and rocking when placed on a flat level surface.

Tower. The tower (cylinder) shall have a minimum capacity of 1 gallon (3.79 liters). The inside of the tower shall have a shelf opening to accommodate an inserted ion-exchange resin, cartridge (Crystalab Model FC-10 or equal). The top (open end) of the tower shall be provided with a rubber cushion to support an inverted l-gallon bottle (not to be supplied with water demineralize), thereby increasing the capacity to 2 gallons (7.75 liters). The tower shall have provisions to accommodate the outlet tubing, and electrodes for the quality check meter electrical circuitry.

Base. The base shall have nominal dimensions of 4-1/4 inches (10.8 cm) high by 6 inches (15.24 cm.) wide by 8-1/4 inches (20.95 cm.) wide. The top of the base shall have a cutout with a rubber cushion to support the tower. The base shall have additional cutouts for the quality check meter, push-botton switch, power cord and outlet tubing. The base shall be capable of being disassembled to allow access to and removal of the tower, quality check meter, and electrical circuitry,

Quality check meter. The quality check meter shall be calibrated in both arts per million (ppm) and the equivalent resistance in megohms. The meter scale shall include a range of O to 50 ppm. (infinity to 0.01 megohms). A push-button type switch, that is accessible on the base housing, shall be provided to activate the meter when readings are required.

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimilted.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Outlet tubing. The outlet tubing shall be connected to the lower section of the tower to control the discharge of deionized water. The free end of the outlet tubing shall be exposed through a cutout in the base. The outlet tubing shall be supplied with a shut-off clamp to stop or permit the discharge of the deionized water as required.

Power cord. A 3-conductor cord for the quality check meter shall be supplies with a non-detachable connector plug. The connector plug shall be a 3-blade male plug in accordance with the American National Standards Institute (ANSI) Standard C73.11. The issue of (ANSI) C73.11 in effect on the date of solicitation shall be used to determine compliance with stated requirements. The overall exposed length of cord and plug shall not be less than 5-1/2 feet (167 cm).

Power feature. The quality check meter shall operate from 115 volt, 60 hz, AC . power supply lines.

Performance. The inserted ion-exchange resin cartridge, when properly secured in the shelf opening of the tower, shall effect a watertight seal around the shelf cutout. The water demineralize shall operate by filling the upper chamber (i.e. space above shelf opening) of the tower with water to a level above the top of the inserted cartridge so that the water passes through the cartridge into the lower chamber of the tower. The water shall not exceed 105° Fahrenheit (40.5° Centigrade). The water passing through the cartridge shall be deonized by the ion exchange resins. By gravity flow, the deionized water in the lower chamber of the tower shall discharge through the outlet tubing. The quality check meter shall be used to check the salinity of deionized water that has accumulated in the lower chamber of the tower.

Workmanship. The water demineralize shall be free from defects which detract from its appearance or will impair its serviceability.

Unit. Each (EA.) One demineralize, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, he contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall maintained by the contractor and made available to the Government upon the government's request, at any time,. or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Inspections. Inspections as used herein, is defined as both examination (such as or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspection and testing. The water demineralize shall be inspected and tested to determine compliance with specification requirements.

Sampling for inspection and testing. Sampling for inspection and testing shall be conducted in accordance with commercial practice.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this descrip tion and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements,

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

For information only. The following product, as of the date of this document, is purported by the following company to meet the salient characteristics of the item described herein: CRYSTALAB Incorporated Model CL-5.

Note: The following National Stock Number is covered by this document:

4610-00-976-2420

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

COMMERCIAL ITEM DESCRIPTION A-A-51367

The following NSN, covers the cartridge, water demineralize, that is to be used with this item: 4610-00-921-5622.

Ordering data, (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVILIAN AGENCY COORDINATING ACTIVITY:

VA-OSS

PREPARING ACTIVITY:

DoD -MB

Agent:

DLA-DM

Project No. 4610-0108

Location: S3-9-5

☆U.S. GOVERNMENT PRINTING OFFICE: 1986 — 605-036/48076

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-ATTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51367	DIMINERALIZER, WATER, ION EXCHANGE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.