A-A-51369
25 MARCH 1986

COMMERCIAL ITEM DESCRIPTION

PROBE, GENERAL OPERATING

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers three sizes of general operating probes suitable for exploration of a wound or surgical cavity during general surgical procedures.

Salient characteristics:

Style and design. The probe shall have a bulbous tip on one end and an eye on the opposite end. The probe shall be straight and of the style and dimensions shown in Figure 1 for the size specified.

Material. The probe shall be manufactured from Class 3 Type 304 or Type 316 austenitic corrosion-resisting steel of the composition given in American Society for Testing and Materials (ASTM) Standard F899.

Finish. The probe shall have a mirror or satin finish, as defined in Federal specification GG-I-526.

Hardness. The finished instrument shall have Rockwell hardness of Rb 75-Rb 90. Finished instruments shall be tested for hardness in accordance with ASTM Standard E18 or ASTM Standards E92 and E140.

Identification markings. The probe shall bear the manufacturer's name or registered trademark and the word "Stainless." Instruments of foreign manufacture shall bear the name of the country of manufacture.

Resistance to corrosion. The finished instrument shall meet the requirements of the Resistance to Boil Test as specified in Federal Specification GG-I-526.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

AMSC N/A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Workmanship. The bulbous tip shall be smoothly and completely rounded. The finish shall be uniform and free from cracks and burrs. There shall be no defects which detract from the appearance or impair the serviceability of the probe.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with the requirements specified.

Unit. Each (EA). One probe, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one probe. Sampling shall be S-1 with an AQL of 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-119-0017	PROBE, GENERAL OPERATING, 8 Inch (20.3 cm)
6515-01-119-0018	PROBE, GENERAL OPERATING, 5 Inch (12.7 cm)
6515-01-119-0787	PROBE, GENERAL OPERATING, 10 Inch (25.4 cm)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:


DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:


VA-OSS

Project No. 6515-3557

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.



SIZE	1	2	3
A	5	8	10



PROBE, GENERAL OPERATING

FIGURE 1

U.S. GOVERNMENT PRINTING OFFICE: 1986 - 605-036/42697

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.