INCH-POUND

A-A-51370C
25 January 1993
SUPERSEDING
A-A-51370B
30 April 1992

COMMERCIAL ITEM DESCRIPTION

DRAPES, SURGICAL

(Self-Adhering)

The General Services Administration has authorized the use of this Commercial Item Description

This Commercial Item Description covers a sterile, self-adhering surgical drape.

Salient Characteristics:

The surgical drape shall be a sterile, self-adhering, translucent, plastic, soft, pliable, disposable drape which shall be repellent to low surface tension fluids and suitable for use in surgical procedures. The following types of drapes are described herein:

Type	Overall Dimensions (Inches)	Adhesive Area (Inches)
A	17-5/8 by 35-3/8	17-5/8 by 23-1/2
B	47-1/8 by 51-1/8	* 1-3/4
C	15-5/8 by 15-5/8	3-7/8 by 4-7/8
D	9 by 18	** 9 by 10-1/8
E	11-3/4 by 17-5/8	1-1/8 by 17-5/8
F	17-5/8 by 23-1/2	1-1/8 by 17-5/8
G	35-3/8 by 43-1/4	1/2 by 35-3/8
H	35-3/8 by 33-3/8	23-1/2 by 33-3/8
J	47-1/8 by 51-1/8	4-7/8 by 3-7/8
K	23-1/2 by 35-3/8	*** 23-1/2 by 1-3/8
L	32 by 50	4 by 5
M	15-5/8 by 15-5/8	**** 3-7/8 by 4-7/8

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

* Adhesive area shall be around a U-slot which shall be 6-3/8 by 19-1/4 inches.
** There shall be an aperture of 3-1/4 by 4-1/4 inches.

*** Cotton adhesive strip 1-1/4 inches wide at top. Circular aperture 1-7/8 inches in diameter.

**** Drape has a 2-1/4 inch diameter center circle aperture.

Tolerances on the above overall dimensions and adhesive area dimensions:

0 to 2 inches - plus or minus 1/8 inch.
Over 2 inches to 12 inches - plus or minus 1/4 inch.
Over 12 inches to 24 inches - plus or minus 1/2 inch.
Over 24 inches to 36 inches - plus or minus 2 inches.
Over 36 inches - plus or minus 3 inches.

Variations 10 percent beyond the above tolerances are permissible in the plus direction only for overall dimensions. Adhesive area dimensions may be under or over by 10 percent.

The drape shall be a sheet of plastic film with a thickness of not less than 0.0012 inch and a minimum ultimate tensile strength of 1.5 pounds per square inch. One side of the film sheet shall be coated with a translucent, non-rubber co-polymer adhesive (adhesive area dimensions) which shall be pressure-sensitive and chemically inert. A protective liner shall cover the adhesive, shall be easy to remove, and shall not lift off any of the adhesive upon removal.

The drape shall be sufficiently transparent to permit a clear view of the operative and surrounding areas and shall be suitable for use with x-ray equipment without removal. The drape shall be easily applied, adhere uniformly, and permit incision directly through the sheet. The drape shall be non-irritating and shall peel-off easily without leaving residue. The drape shall be treated to reduce build-up of static electricity during handling or application and shall comply with NFPA 99, Health Care Facilities Code.

The components and internal surfaces of each drape package shall be sterile and in accordance with the USP or AAMI requirements. Sterility assurance level shall be 0.000001. Each drape shall be individually packaged in a suitable container to permit easy opening under septic conditions and shall maintain sterility until opened.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Labeling shall be that usually furnished by the supplier as commercial practice. The inspection/test dating period shall not be less than 36 months. Not less than 33 months of the inspection/test dating period shall remain at the time of delivery to the Government.

Workmanship. The surgical drape shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing ten individually packaged sterile drapes of one type only of types A through G, L and M or forty individually packaged sterile drapes of one type only of types H, J, and K, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one surgical drape.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one surgical drape.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document.

6530-00-269-3597 - Type A
6530-01-047-0175 - Type B
6530-01-047-0176 - Type C
6530-01-047-9696 - Type D
6530-01-047-9697 - Type E
6530-01-047-9698 - Type F
6530-01-047-9699 - Type G
6530-01-151-1830 - Type H
6530-01-151-1831 - Type J
6530-01-151-6423 - Type K
6530-00-269-3592 - Type L
6530-00-269-3593 - Type M

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

A-A-51370C

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA - OSS

Project Number: 6530-2418
Location: DRPMC13A.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-51370C	2. DOCUMENT DATE (YYMMDD): 25 January 1993

3. DOCUMENT TITLE: DRAPES, SURGICAL (SELF-ADHERING)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 98/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.