MIL-A-36080C(DM)
25 March 1965
SUPERSEDING
MIL-A-36080B(DM)
11 February 1964

MILITARY SPECIFICATION

ANESTHESIA SET, ENDOTRACHEAL

1. SCOPE

1.1 This specification covers an endotracheal anesthesia set, suitable for use in the administration of anesthesia.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
QQ-A-225/5b	Aluminum Alloy Bar, Rod and Wire; Rolled, Drawn or Gold Finished, 2017
QQ-B-626	Brass, Leaded and Non-Leaded: Rods, Shapes, Forgings and Flat Products with Finished Edges (Bars, Flat Wire and Strips).
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-Up, Paperboard.

Military

MIL-P-116	Preservation, Methods of.
MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:04Z
Check the source to verify that this is the current version before use.

STANDARDS

Federal

Fed. Std. No. 595	Colors.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.
MS16686	Medical Services Symbol.

DRAWINGS

Defense Medical Supply Center

1249	Lamp, Incandescent, 2.5 volt, 0.28 Ampere, Bulb T-1-1/2, Frosted.
2165	Catheter Stylet and Stop for Catheter Stylet.
2166	Adapter, Y-Piece, Endotracheal Anesthesia Set.
2168	Tubes and Elbows, Endotracheal.
20182	Lamp, Incandescent, 2.7 Volt, 0.31 Ampere, Bulb T-1, Frosted.
20685	Adapter, Right Angle Elbow, Endotracheal Anesthesia Set.

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publication. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

AMERICAN SOCIETY FOR TESTING MATERIALS

ASTM Standard EL Thermometers.

(Application for copies should be addressed to the American Society for Testing Materials, 116 Race Street, Philadelphia, Pa. 19103).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:04Z
Check the source to verify that this is the current version before use.

U.S. PHARMACOPEIAL CONVENTION, INC.

Pharmacopeia of the United States.

(Application for copies should be addressed to the Mack Publishing Company, Easton, Pa.)

American Standards Association, Incorporated

Z79.2-1961 Endotracheal-Tube Connectors and Adapters

(Application for copies should be addressed to the American Standards Association, Inc., 10 East 40th Street, New York, N. Y. 10016).

3. REQUIREMENTS

3.1 Material.

3.1.1 Connectors. Connectors shall be fabricated of 17S aluminum alloy in accordance with the requirements of Specification QQ-A-225/5b for T4 temper material.

3.1.2 Y-piece adapter. The Y-piece adapter shall be fabricated from either 2017 aluminum alloy in accordance with the requirements of Specification QQ-A-225/5b, for T4 temper material; or the material shall be in accordance with Specification QQ-B-626 for chemical composition 22; and plating requirements and corrosion test requirements shall be in accordance with class 4, composition N of Specification GG-I-526, except that the Y-section may be fabricated of class 4, composition P material of Specification GG-I-526. As an alternate, the Y-piece connector shall be fabricated from brass, and shall be chromium plated (the outside shall be mirror finish, the inside shall be satin finish).

3.1.3 Adapter (Right angle elbow) and catheter stylet stop. Adapter and catheter stylet stop shall be fabricated from material in accordance with Specification QQ-B-626 for chemical composition 22. The plating requirements and corrosion test requirements shall comply with class 4, composition N material of Specification GG-I-526.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:04Z
Check the source to verify that this is the current version before use.

3.1.4 Laryngoscope. Laryngoscope shall be fabricated of brass, and shall be in accordance with the requirements of Military Medical Purchase Description in effect on date of invitation for bids for FSN 6515-616-5052 (see 3.2.9 and 6.2).

3.1.5 Catheter stylet. The catheter stylet shall be fabricated of the highest quality commercial copper tubing, suitable in all respects for the purpose intended.

3.1.6 Tubes.

3.1.6.1 Magill tubes. Magill tubes shall be fabricated of material of the best quality, and suitable in all respects for the purpose intended. (See 3.2.4).

3.1.6.2 Plastic tube. The tube shall be fabricated of a medical-grade polyvinyl material. Tubing shall contain not more than 5 parts per million (p.p.m.) of lead when tested as specified in 4.4.3.

3.1.7 Tube cuffs. The endotracheal tube cuffs shall be fabricated of natural latex of the highest quality throughout, and suitable in all respects for the purpose intended.

3.1.8 Case. The case shall be fabricated of aluminum suitable in all respects for the purpose intended.

3.1.9 Other material. All other material hereinafter specified shall be of the highest quality throughout, and suitable in all respects for the purpose intended.

3.2 Description. The endotracheal anesthesia set shall consist of a case containing all of the following components in the specified quantities:

Part	Quantity
Connector, endotracheal tube, 6 mm	1
Connector, endotracheal tube, 8 mm	1
Connector, endotracheal tube, 10 mm	3
Connector, endotracheal tube, 11 mm	1
Connector, endotracheal tube, 12 mm	1

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:04Z
Check the source to verify that this is the current version before use.

Part	Quantity
Tube (plastic) and elbow, endotracheal, 14 Fr.	1
Adapter, Y-piece, Endotracheal Anesthesia Set	1
Adapter, right angle elbow, Endotracheal Anesthesia Set	2
Stop, catheter, stylet	1
Stylet, catheter, flexible, copper	1
Tube, Magill, endotracheal, 5.5 mm id; 24 cm long	2
Tube, Magill, endotracheal, 7.0 mm id; 30 cm long	2
Tube, Magill, endotracheal, 8.0 mm id; 32 cm long	2
Tube, Magill, endotracheal, 9.0 mm id; 34 cm long	4
Tube, Magill, endotracheal, 9.5 mm id; 35 cm long	1
Tube, Magill, endotracheal, 10.0 mm id; 36 cm long	2
Cuff, endotracheal tube, 6.5 mm id	2
Cuff, endotracheal tube, 8.0 mm id	3
Cuff, endotracheal tube, 9.5 mm id	3
Laryngoscope, infant-child-adult MacIntosh type	1
Lamps, incandescent, laryngoscope (spare lamps)	5
Forceps, endotracheal, tube, Magill	1

3.2.1 Endotracheal tube connectors. The endotracheal tube connectors shall be fabricated of material as specified in 3.1.1 and shall be furnished in the sizes and quantities as specified in 3.2. The straight connectors shall be suitable for connecting the Y-piece adapter or the Rovenstine-Woods elbow specified in 3.2.2 and 3.2.3, respectively, to the Magill tubes specified in 3.2.4. The straight connectors shall be in accordance with the applicable portions of paragraphs 4 and 6, for the size indicated, in the American Standards Association ASA Z79.2-1961, and shall be in accordance with figure 2 and a serrated finger grip portion may be supplied.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:04Z
Check the source to verify that this is the current version before use.

3.2.2 Y-piece adapter The Y-piece adapter shall be fabricated of material as specified in 3.1.2. One Y-piece adapter shall be furnished with this set and shall be suitable for connecting the Rovenstine-Woods elbow specified in 3.2.3, or the endotracheal tube connectors specified in 3.2.1, to the gas circuit, that is, hoses of the anesthesia apparatus, Stock Nos. 6515-301-0400 and 6515-301-0430 (see 6.2). Y-piece adapter shall be constructed as shown in Drawing 2166.

3.2.3 Elbows. The elbows shall be of the Rovenstine-Woods design, and shall be fabricated of material as specified in 3.1.3. Two elbows shall be furnished with this set. The elbows shall be suitable for use as a connecting elbow with provision for insertion of an aspirating tube, between the endotracheal tube connectors specified in 3.2.1, and the Y-piece adapter specified in 3.2.2, in order to change the direction of tube attitude in relation to the Y-tube and the patient. The Rovenstine-Woods adapter shall be constructed as shown in Drawing 20685.

3.2.4 Magill endotracheal tubes. Tubes shall be furnished in the sizes and quantities specified in 3.2. Tubes shall be in accordance with the requirements of Military Medical Purchase Description in effect on date of invitation for bids for FSNS 6515-299-8543, 6515-316-2150, 6515-961-5519, 6515-961-5520, 6515-961-5521, 6515-961-5523 and 6515-961-5525. Two tubes each shall be supplied, except, that one tube shall be supplied for 6515-316-2150. Tubes need not be sterile, but shall be clean.

3.2.5 Plastic tube (14 Fr) and elbow. The plastic tube and elbow shall consist of a 14 french plastic tube and metal elbow and shall be suitable for insufflation and suction during endotracheal anesthesia. One plastic tube and elbow shall be furnished with this set.

3.2.5.1 Elbow. The metal elbow shall be fabricated of material as specified in 3.1.3. The elbow shall be fabricated with a finger valve and shall be constructed as shown in Drawing 2168.

3.2.5.2 Plastic tube (14 Fr). The tube shall be a 14 Fr size and shall be capable of being directly attached to metal elbow specified in 3.2.5.1, in a tight, leakproof manner. The finished tube shall be symmetrical in shape, and the surface shall have a smooth finish, free from irregularities or imperfections of any kind and shall not collapse when used for suction. Tubes shall be pyrogen free when tested for freedom from pyrogens as specified in 4.4.1. The tube shall pass the animal implantation test and shall show no evidence of toxicity, or adverse tissue reaction upon gross

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:04Z
Check the source to verify that this is the current version before use.

and microscopic examination, when tested as specified in 4.4.2. The tubes shall pass the resterilization test and shall not become distorted, cracked nor shall be adversely affected so as to impair their use when tested as specified in 4.4.4. French size shall be legibly and permanently marked on the tube. The tube shall be in accordance with Drawing 2168.

3.2.6 Cuff, endotracheal tubes. Cuffs shall be furnished in the sizes and quantities as specified in 3.2. Cuffs shall be in accordance with requirements of Military Medical Purchase Description in effect on date of invitation for bids for FSNS 6515-321-5000, 6515-321-5025 and 6515-321-5050.

3.2.7 Catheter stylet. The catheter stylet shall be fabricated of material as specified in 3.1.5. The stylet shall be a hollow, flexible stylet with olivary tips and shall be suitable for directing the Magill catheters during intubation. Stylet shall be constructed as shown in Drawing 2165.

3.2.8 Catheter stylet stop. The catheter stylet stop shall be fabricated of material as specified in 3.1.3. The stop shall be suitable for use in regulating the stylet length to a desired position, preventing its introduction beyond such position during the intubation of a catheter. The stop shall be constructed as shown in Drawing 2165.

3.2.9 Laryngoscope(and lamps). Laryngoscope shall be fabricated from material as specified in 3.1.4. Shall be in accordance with FSN 6515-616-5052, except case shall not apply and 3 additional lamps FSN 6240-552-9672 shall be supplied in lieu of two (see 6.2).

3.2.10 Forceps. Forceps shall be in accordance with the requirements of Military Medical Purchase Description in effect on date of invitation for bids for FSN 6515-332-3300 (see 6.2).

3.2.11 Case. The case shall be fabricated of material as specified in 3.1.8. The interior compartment shall be constructed of flocked material suitable to properly retain the components. The catches and handle shall be constructed of suitable material, either brass or stainless steel.

3.3 Fitting of components. All connections made by the components of the set shall be tight and leakproof.

3.4 Case finish. Case and hardware shall be finished in a lustreless olive drab shade conforming to color 34087 of Federal Standard No. 595 (see 6.3).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:04Z
Check the source to verify that this is the current version before use.

3.5 Identification marking. Except as specified above, each component shall be permanently and legibly marked with the contractor's name or registered trademark. The outside surface of the cover of the case shall be stenciled with a caduceus between the letters "U" and "S" as shown for type I, in Standard MS16686, plus the Stock No. underneath the caduceus. The caduceus shall be approximately 1-1/2 inches in height; the letters and stock number shall be of approximately relative size. The caduceus and the letters shall be stenciled with a suitable black paint. The above specified marking shall be centrally located on the outside surface of the case. The inside surface of the cover of the case shall be permanently and legibly marked with the contractor's name or registered trademark.

3.6 Certification for freedom from pyrogens, and animal implantation tests. A certificate (s) listing all lot or control numbers, of endotracheal and plastic tube shall state that representative sample (s) from each lot have been tested and have been found to be free from pyrogens and have passed the animal implantation test. This certificate (s) shall include the specific tests performed and the test results. The record of the animal implantation test and its results shall be certified by the Pathologist who performed or supervised performance of the animal implantation test. When inspection is made at source, the certificate (s) shall be submitted to cognizant inspector for transmittal to Chief, Quality Control and Maintenance Division, Directorate of Technical Operations, Defense Medical Supply Center, 3rd Avenue and 29th Street, Brooklyn, New York 11232. When inspection is made at destination, the certificate (s) shall be submitted by the contractor directly to Chief, Quality Control and Maintenance Division, Directorate of Technical Operations, Defense Medical Supply Center, 3rd Avenue and 29th Street, Brooklyn, New York 11232 and a copy shall be forwarded with shipment to the consignee.

3.7 Manufacturer's record. The manufacturer shall place a code or lot number on the paper label, or on the printed carton, which can be verified against the manufacturer's records of all tests made on animal implantation test, and pyrogen-free tubes. These records shall be made available for inspection by any duly authorized Government representative. No shipment shall be made until all tests have been completed, verified, and recorded in the permanent records of the manufacturer.

3.8 Workmanship. Workmanship shall be first class throughout. The endotracheal anesthesia set, case, and components shall be free from defects which detract from their appearance or may impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:04Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with Standard MIL-STD-105.

4.2.2 For tests.

4.2.2.1 For pyrogen. Sampling procedures for pyrogen test shall be conducted as indicated in the current edition of the Pharmacopeia of the United States, for transfusion and infusion assemblies.

4.2.2.2 For implantation test. Two samples shall be cut from every manufacturer's lot of tubing for test. If more than two tissue sites per animal show a reaction, the lot represented shall be rejected.

4.2.2.3 For lead test. Sample shall be taken for each lot of plastics as specified in 4.4.4.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:04Z
Check the source to verify that this is the current version before use.

4.2.2.4 For thickness of coatings, salt spray, boil and sterilization tests. Sampling for each test shall be conducted in accordance with Standard MIL-STD-105, inspection level S1, except that the acceptance number shall be zero (0) for all sample sizes, and a minimum of four samples shall be taken regardless of lot size.

4.2.3 Raw materials. Specimens of material for chemical analysis shall be obtained from the samples selected for thickness of coating, salt spray and boil tests.

4.2.4 Where possible, the same sample shall be used for the determination of two or more test characteristics.

4.3 Examination. The endotracheal anesthesia sets shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Pyrogen. Tubes shall be tested for freedom from pyrogens, in accordance with the current edition of the Pharmacopeia of the United States, for transfusion and infusion assemblies.

4.4.2 Animal implantation test. The samples selected shall have been processed (including sterilization) so as to be representative of the finished product. Care shall be exercised in the preparation of samples and specimens to prevent contamination with micro-organisms and other foreign matter. Clean cutting devices by appropriate method (e.g., successive cleaning with acetone and methylene chloride) prior to subdividing a sample. Clean all other equipment by thorough detergent scrubbing and prolonged rinsing with tap and distilled water. Render containers and devices used in transfer and administration of test material sterile and dry by a suitable process. If ethylene oxide, or similar agent, is used as sterilizing agent, sufficient time for complete degassing shall be allowed. Each sample (cut from manufacturer's lot of tubing) shall be divided into four specimens. The specimens shall be comprised of solid strips measuring at least 1 millimeter (mm.) by 10 mm. The edges of the strips shall be smooth to avoid mechanical trauma upon implantation. Any foreign material, such as lint and particles, shall be removed by placing the specimens in a beaker or cylinder and adding approximately 70 milliliters (ml.) of water for

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:04Z
Check the source to verify that this is the current version before use.

injection, and then agitating for approximately 30 seconds. This cleaning procedure shall be performed twice. (Note: The specimens shall not be cleaned with a cloth or by washing with an organic solvent, a surface active agent, etc.) The test shall be performed in a clean area using aseptic techniques. Healthy, adult rabbits weighing 4 to 6 pounds whose paravertebral muscles are sufficiently large in size to allow for embedding of the test specimens, shall be utilized. No longer than 20 hours before testing, the fur of the animals shall be clipped on both sides of the spinal column. A vacuum shall be used for removal of the clipped hairs so as to prevent their introduction into the implantation site and thereby causing a tissue reaction. If necessary, the animals shall be anesthetized with a commonly used anesthetic agent to a degree deep enough to prevent muscular movements such as twitching. The specimens, of the specified size, shall be implanted by means of a sterile hypodermic needle (e.g., a 15 gage, 3/4 inch needle with I.V. point) and a sterile stylet. The needle may be presterilized and the sterile specimen aseptically inserted therein, or the needle with the clean specimen inserted in the cannula, and with the hub and cannula properly covered, shall be appropriately sterilized (allowing for proper degassing if agents such as ethylene oxide are used). Four specimens shall be implanted into the paravertebral site (e.g., spinus rectus or sacrospinalis muscle) of each of two rabbits. Each needle shall be inserted, at an angle of approximately 30 degrees, 2 1/2 to 5 centimeters (cm.) from the midline and parallel to the spinal column, and approximately 2 1/2 centimeters apart from each other. The sterile stylet shall be inserted into the needle to hold the specimen in the tissue while withdrawing the needle. If excessive bleeding is observed after implantation of a specimen, a duplicate shall be placed. After 14 days, the rabbits shall be sacrificed using common laboratory techniques (e.g., with an overdose of an anesthetic agent). To prevent the animals from bleeding, the autopsy shall be performed after a sufficient time has lapsed. The muscle tissue surrounding the implants shall then be examined, both grossly and microscopically, by a pathologist certified by the American Board of Pathology or the American College of Veterinary Pathologists. If no tissue reaction is noted, test is completed and material is acceptable (see 4.2.2.2). If tissue reaction is noted, a bacterial stain shall be utilized to help rule out contamination. If three or more of the tissue sites per animal show a reaction (not ruled out as contamination), test is completed and material is not acceptable (see 4.2.2.2). If one or two of the tissue sites per animal show a reaction (not ruled out as contamination), then retest on three more animals shall be performed (see 4.2.2.2 for acceptance criteria).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:04Z
Check the source to verify that this is the current version before use.

4.4.3 Test for lead. Ten (10) Gms. of plastic shall be taken, ignited to char, and then further ignited in a muffler at 700° C. to a white ash. The ash shall be boiled in 15 milliliters (ml.) of 10 percent nitric acid filtered into a volumetric flask and made up to 100 ml. An aliquot portion, equivalent to one gram of sample, shall be taken and placed in a separator. The dithizone test for lead, as given in the current edition of the Pharmacopeia of the United States, shall be applied using a control of 5 p.p.m. or 12 p.p.m. of lead solution as given in the U.S.P. test. If the sample solution is darker than the limit, smaller aliquots shall be used until the color of the control is matched by the sample solution and the concentration of lead in p.p.m. is determined.

4.4.4 Sterilization.

4.4.4.1 Steam sterilization. Tubes shall be steam autoclaved for 20 minutes at 250° Fahrenheit (F.) and 15 pounds per square inch gage (p.s.i.g.), then allowed to cool to room temperature (68° F. to 72° F.).

4.4.4.2 Boil sterilization. Tubes shall be boiled for 30 minutes and then allowed to cool to room temperature (68° F. to 72° F.).

4.4.4.3 Cold sterilization. Tubes shall be immersed in a chemical disinfectant for 20 minutes at room temperature (68° F. to 72° F.), and then followed by five (5) water rinses.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue.

5.1.1.1 Anesthesia set - One (1) anesthesia set, as specified, constitutes one unit of issue.

5.1.1.2 Connector - One (1) connector, as specified, constitutes one unit of issue.

5.1.1.3 Adapter (Y-piece) - One (1) adapter (Y-piece), as specified constitutes one unit of issue.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:04Z
Check the source to verify that this is the current version before use.

5.1.1.4 Adapter (right angle elbow) - One (1) adapter (right angle elbow), as specified, constitutes one unit of issue.

5.1.1.5 Stop-One (1) stop, as specified, constitutes one unit of issue.

5.1.1.6 Stylet-One (1) stylet, as specified, constitutes one unit of issue.

5.1.1.7 Tube and elbow-One (1) tube and elbow, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table I. Packaging Quantities

Item	Unit Package	Intermediate package	Exterior Container
Anesthesia set	1 unit	Not required	6 units
Connector	1 unit	12 units	576 units
Tube and elbow	1 unit	12 units	288 units
Adapter (Y-piece)	1 unit	12 units	144 units
Adapter (right angle elbow)	1 unit	12 units	576 units
Stop	1 unit	12 units	576 units
Stylet	1 unit	12 units	576 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptance to a common carrier, which will insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:04Z
Check the source to verify that this is the current version before use.

5.1.3.1.1 Anesthesia set. Component parts shall be cushioned within the case in such manner as to prevent shifting or damage. Case shall be packaged in a carton of appropriate size constructed in accordance with Specification PPP-B-636, type CF, class domestic. Closure shall be effected in such manner as to prevent spillage of contents under normal handling. Each unit shall be further packaged in accordance with method IA-2 or IA-8 of Specification MIL-P-116.

5.1.3.1.2 Connector, adapter (right angle elbow), or stop (when procured separately). Each unit shall be packaged in an envelope of suitable size and design. Envelope shall be adequately secured to prevent accidental opening.

5.1.3.1.3 Tube and elbow, adapter (Y-piece), or stylet (when procured separately). Each unit shall be packaged in a carton of appropriate size constructed in accordance with Specification PPP-B-566 or PPP-B-676, except that commercial colors shall be acceptable. Closure shall be adequate to prevent spillage of contents under normal handling. Carton shall be further packaged in accordance with method IA-8 of Specification MIL-P-116.

5.1.3.2 Intermediate package.

5.1.3.2.1 Connector, tube and elbow, adapter (Y-piece and (right angle elbow), Stop, or Stylet (when procured separately). Intermediate package shall be a carton of appropriate size constructed in accordance with Specification PPP-B-566 or PPP-B-676, except that commercial colors shall be acceptable. Closure shall be adequate to prevent spillage of contents under normal handling.

5.1.4 Level B.

5.1.4.1 Unit package.

5.1.4.1.1 Anesthesia set. Component parts shall be cushioned within the case in such manner as to prevent shifting or damage. Case shall be packaged in a carton of appropriate size constructed in accordance with Specification PPP-B-636, type CF, class domestic. Closure shall be effected in such manner as to prevent spillage of contents under normal handling.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:04Z
Check the source to verify that this is the current version before use.

5.1.4.1.2 Connector, adapter (right angle elbow), or stop (when procured separately). Unit package shall be as specified in 5.1.3.1.2.

5.1.4.1.3 Tube and elbow, adapter (Y-piece), or stylet (when procured separately). Each unit shall be packaged in a carton of appropriate size constructed in accordance with Specification PPP-B-566, or PPP-B-676, except that commercial colors shall be acceptable. Closure shall be adequate to prevent spillage of contents under normal handling.

5.1.4.2 Intermediate package.

5.1.4.2.1 Connector, adapter (right angle elbow), or stop (when procured separately). Intermediate package shall be as specified in 5.1.3.2.1.

5.1.4.2.2 Tube and elbow, adapter (Y-piece), or stylet (when procured separately). Intermediate package shall be a carton of appropriate size and design, or a complete wrapping of kraft paper of suitable weight. Closure shall be adequate to prevent spillage of contents under normal handling.

5.1.5 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for the shipment and use of the product for its intended purpose within the continental limits of the United States of America.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:04Z
Check the source to verify that this is the current version before use.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4, of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic and method I closure shall be utilized.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit carton and barrier, when used, and each intermediate package shall be marked in accordance with Standard MIL-STD-129, except that when labels are utilized, water-proofing shall be required only when applicable container is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown. In addition, and as applicable, lot (control) numbers as indicated in 3.6 shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with Standard MIL-STD-129. In addition, and as applicable, lot (control) numbers as indicated in 3.6 shall be shown.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:04Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 Copies of Military Medical Purchase Descriptions for stock numbers enumerated in 3.1.4, 3.2.2, 3.2.4, 3.2.6, 3.2.9, 3.2.10, and 5.1 may be obtained upon application to the procuring activity or as directed by the contracting officer.

6.3 Availability of color ships. The individual 3-by-5-inch chips of the color referenced in 3.4 are available from General Services Administration, Business Service Center, Region 3, Seventh and D Streets, S.W., Washington, D. C. 20407.

6.4 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:

Defense Supply Agency - DM

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

U.S. GOVERNMENT PRINTING OFFICE: [illegible]

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:04Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:04Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity *(as indicated on reverse hereof)*.

SPECIFICATION

ORGANIZATION *(Of submitter)*	CITY AND STATE

CONTRACT NO.	QUANTITY OF ITEMS PROCURED	DOLLAR AMOUNT
		$

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES.

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO IF "YES", IN WHAT WAY?

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity)*	DATE

DD FORM 1 APR 63 1426 REPLACES NAVSHIPS FORM 4863, WHICH IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:04Z
Check the source to verify that this is the current version before use.

FOLD

DEPARTMENT OF THE NAVY

POSTAGE AND FEES PAID
NAVY DEPARTMENT

OFFICIAL BUSINESS

Commander
Defense Personnel Support Center
Attn: DPSC-ATTS
Material Standards Branch
Technical Operations Division
Directorate of Medical Material
2800 South 20 th Street
Philadelphia, Pa. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:04Z
Check the source to verify that this is the current version before use.