MILITARY SPECIFICATION

BACK REST, LITTER, NONADJUSTABLE

1. SCOPE

1.1 This specification covers a nonadjustable litter back rest suitable for use with Litter, Folding, Rigid Pole, Aluminum Pole (see 6.1). The back rest shall be suitable for use in military ambulance evacuation aircraft; and in field and evacuation hospitals, to support litter casualties.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

V-T-276	Thread, Cotton.
QQ-A-601	Aluminum Alloy Sand Castings.
TT-E-527	Enamel, Alkyd, Lustreless.
WW-T-789	Tube, Aluminum Alloy, Round, Square, Rectangular, and Other Shapes, Drawn, Seamless, 6061 and 6062.
CCC-C-419	Cloth, Cotton, Duck, Unbleached, Plied-Yarns (Army and Numbered).
CCC-T-191	Textile Test Methods.
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-Up, Paperboard.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

Military

MIL-W-530	Webbing, Textile, Cotton, General Purpose, Natural or in Colors.
MIL-T-3530	Treatments, Mildew-Resistant and Water Repellent for Thread and Twine.
MIL-C-5541	Chemical Films for Aluminum and Aluminum Alloys.
MIL-P-8585	Primer Coating, Zinc Chromate, Low-Moisture-Sensitivity.
MIL-L-10547	Liners, Case, Waterproof.

STANDARDS

Federal

Fed. Test Method Std. No. 151	Metals; Test Methods.
Fed. Std. No. 595	Colors.
Fed. Std. No. 751	Stitches, Seams, and Stitchings.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.
MS16686	Medical Services Symbol.

DRAWING.

Military Medical Supply Agency

10013 Back Rest, Litter, Nonadjustable.

PURCHASE DESCRIPTION

Quartermaster Corps

IP/DES S-243-1	Dyeing and Aftertreating Processes for Cotton Fabrics, CCC-D-950(1).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

(Copies of specifications, standards, drawing, and purchase description required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publication. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids shall apply.

COLOR ASSOCIATION OF THE UNITED STATES, INC.

United States Army Standard Color Card for the Official Standardized Shades of Sewing Threads.

(Application for copies should be addressed to the Color Association of the United States, Inc., 200 Madison Avenue, New York 16, N.Y.)

3. REQUIREMENTS.

3.1 Materials. Except as hereinafter specified, materials shall conform to the applicable portions of the specifications, standards, drawing (with bill of material), and purchase description specified herein and shall meet the requirements and be subjected to the tests specified therein. In addition, the metals shall conform to the chemical requirements of their applicable specifications when tested as specified in 4.4.1. When materials to be used are not specifically designated, those used shall be entirely suitable for the purpose intended.

3.1.1 Back rest frame clamp. Back rest frame clamp shall be formed of an aluminum alloy casting, in strict conformance with class 17M, alloy number 40E of Specification QQ-A-601.

3.1.2 Back rest frame. Back rest frame shall be formed from 6061-T6 aluminum alloy tubing, 1 inch in diameter by 0.065 inch wall, in strict conformance with Specification WW-T-789.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

3.1.3 Cover, back rest, frame.

3.1.3.1 Duck, cotton. The cotton duck for the cover, before finishing, shall be ply yarn cotton duck conforming to Specification CCC-C-419 for type I, No. 10, hard texture, as shown in table I of the same specification.

3.1.3.1.1 Color and treatment. Cotton duck shall be vat dyed olive drab No. 7, water-repellent and mildew-resistant, conforming to type I, class B of Interim Purchase Description IP/DES S-243-1, except that the cloth shall be made mildew-resistant by treatment limited to two mildew-inhibitors only: (a) 2,2' methylene-bis (4-chlorophenol), or (b) copper-8-quinolinolate, in accordance with the requirements as shown in table III of the same specification (see 4.4.2).

3.1.3.2 Binding. Except as specified herein, binding for the cover shall conform to Specification MIL-W-530, shall be type I, lightweight webbing, 0.30 ounce per linear yard, 3/4 inch wide, and shall conform to class 4 of the same specification, except that the maximum amount of copper-8-quinolinolate shall be 1.28 percent. The binding shall be vat dyed olive drab No. 7 (see 4.4.3).

3.1.3.3 Thread. Except as specified herein, the thread shall conform to Specification V-T-276, and shall be type III, heavy thread, 3-ply, ticket No. 10/3, as shown in tables V and VI. The thread shall be vat dyed olive drab shade S-1, conforming to U.S. Army Standard Color Card. All references to "approved shade standard", "approved standard shade sample", and "approved standard shade" made in Specification V-T-276 shall not apply. Thread shall be treated for mildew-resistance in accordance with Specification MIL-T-3530, using type I, class 1, mildew-inhibiting agent (see 4.4.4). The finished thread shall show good colorfastness to laundering, chlorine bleaching, and weather. Paragraph titled: "Certification" in Specification V-T-276 shall not apply (see 4.1.2).

3.2 Construction. Construction of the back rest shall conform to the requirements as specified herein and as shown in the drawing. The approximate overall weight of the back rest, including canvas cover and clamps, shall be 3-1/2 pounds.

3.2.1 Back rest frame and clamps. Back rest frame and clamps shall be formed as shown in the drawing. All sharp edges shall be removed and corners shall be slightly rounded. The back rest frame clamps (3.1.1) are designed in such a manner as to fit over the rails of the litter (see 1.1).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

3.2.2 Back rest frame cover. Back rest frame cover shall be fabricated of cotton duck as specified in 3.1.3 and shall conform in all details to applicable portions of the drawing. The cover shall be so constructed that when placed on the inverted "U" frame, the warp shall run parallel to the long axis of the frame. The frame shall provide a back support for the patient. Turn unders and seams shall be neatly made. No loose or raveled edges shall be exposed.

3.2.3 Stitching. All stitching shall be machine stitching conforming to stitch type 301 of Federal Standard No. 751. There shall be a minimum of 8 stitches to the inch. When machine stitching is not a continuous thread, it shall be backstitched not less than 1-inch at each break. All stitching shall be backstitched at the ends to prevent raveling. Thread tension shall be properly maintained so that the lock will be embedded in the center of the materials sewed.

3.2.4 Marking.

3.2.4.1 Insignia. On the back of the back rest cover centrally located, as shown in the drawing, there shall be placed a caduceus 3 inches in height centered between the letters "U" and "S" of appropriate size as shown for type I in Standard MS16686. Marking shall be done with black waterproof ink or black waterproof stencil paint.

3.2.4.2 Identification marking. The name or registered trade-mark of the manufacturer shall be marked in an inconspicuous place on the back rest cover using black waterproof ink or black waterproof stencil paint.

3.2.5 Finish.

3.2.5.1 Degreasing. All aluminum surfaces shall be thoroughly cleaned and chemically degreased prior to chemical treatment and then finished with a chemical film conforming to Specification MIL-C-5541.

3.2.5.2 Primer. After degreasing and chemical treatment, a primer coat shall be applied to all aluminum surfaces. The primer shall be zinc-chromate conforming to Specification MIL-P-8585. The primer coat shall form a grounding bond and base for the finishing enamel.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

3.2.5.3 Enamel. The primer coat shall be followed by one coat of lustreless olive drab enamel conforming to Specification TT-E-527. The resultant finish shall be hard, smooth, and highly resistant to abrasion, chipping, and cracking. The color of the final finish shall be lustreless olive drab conforming to color 34087 of Federal Standard No. 595 (see 6.3).

3.2.6 Workmanship. Workmanship shall be first class throughout. Complete back rests shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Supplier responsibilities for inspection. The supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own or any other inspection facilities and services acceptable to the Government. Inspection records of the examination and tests shall be kept complete and available to the Government as specified in the contract or order. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.1.3 Government furnished material shall not be sampled or tested unless specifically requested by the Government (see 6.1).

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with Standard MIL-STD-105.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with Standard MIL-STD-105.

	Inspection level	AQL (percent defective)	Minimum sample size
Component material testing	I2	*	3

*Acceptance number shall be zero (0) for all sample sizes.

4.2.3 For metals. Four ounces of each lot of each metal used in the construction of the back rest shall be selected for chemical analysis as specified in 4.4.1.

4.2.4 Component materials. Sufficient samples of the component materials shall be selected from each lot of material used in the manufacture of the back rests to perform tests as required by the applicable subsidiary specifications or by tests specified herein. When sample size for component materials covered by the subsidiary specifications is not specified, the sample size for each test shall be as indicated in 4.2.2. The unit of product for testing shall be the quantity necessary to perform all required tests one time each, taking into consideration the fact that, where possible, the same material may be used to perform two or more tests. Units of product are expressed in multiples of the following:

Duck - Yards or fractions of a yard.
Webbing - - - - - - - - - - - - - - - - - - - Yards or fractions of a yard.
Thread - Cones or spools.

4.3 Examination. The back rests shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted, as necessary, to determine compliance with specification requirements.

4.4.1 Chemical analysis of the metals. The samples selected in accordance with 4.2.3 shall be analyzed in conformance with methods 111 or 112 of Federal Test Method Standard No. 151. In case of dispute, method 111 shall be used for referee analysis.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

4.4.2 Cloth, cotton, duck. The unfinished duck shall be tested in accordance with the test procedures of Specification CCC-C-419 (see 3.1.3.1). The finished duck shall be tested in accordance with test procedures of Interim Purchase Description IP/DES S-243-1, except that performance of the tests for mildew-inhibitor content and mildew-resistance shall be conducted under the following conditions:

(a) If the material conforms to the requirements for mildew-inhibitor content, the mildew-resistance tests shall not be performed, and the material shall be considered as complying with mildew-resistance requirements.

(b) If the mildew-inhibitor content exceeds the maximum requirement, the material shall be rejected.

(c) If the mildew inhibitor content is found to be below the minimum requirement, the material may be accepted with respect to mildew-resistance provided it passes the mildew-resistance test, method 5762, as specified in the test section of Interim Purchase Description IP/DES S-243-1. When the mildew-resistance test is performed, the leaching process shall be omitted, and the breaking strength shall be determined by method 5100 of Specification CCC-T-191.

4.4.3 Webbing. The webbing shall be tested in accordance with the test procedures of Specification MIL-W-530, except that the performance of the tests for mildew-inhibitor content and mildew-resistance shall be conducted under the following conditions:

(a) If the material conforms to the requirements for mildew-inhibitor content, the mildew-resistance tests shall not be performed, and the material shall be considered as complying with mildew-resistance requirements.

(b) If the mildew-inhibitor content exceeds the maximum requirement, the material shall be rejected.

(c) If the mildew-inhibitor content is found to be below the minimum requirement, the material may be accepted with respect to mildew-resistance provided it passes the mildew-resistance test.

4.4.3.1 Mildew-resistance. The finished webbing shall be tested in accordance with method 4762 as specified in the test section of Federal Specification CCC-T-191. The leaching process shall be omitted.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

4.4.3.2 Water absorption. The finished webbing shall be tested for water absorption by method 5500 as specified in the test section of Specification MIL-W-530, except that the size of the specimen and the size of the blotting paper shall conform to method 5502 of Specification CCC-T-191.

4.4.4 Thread. The untreated thread shall be tested in accordance with the test procedures of Specification V-T-276 (see 3.1.3.3). The treated thread shall be tested as specified in the test section of Specification MIL-T-3530.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) back rest, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in procurement document.

Table I. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not required	18 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be packaged in a carton of appropriate size constructed in accordance with Specification PPP-B-566, PPP-B-636, type I, class 1, or PPP-B-676. Closure shall be adequate to prevent spillage of contents under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose within the continental limits of the United States of America.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load and constructed in accordance with Specification PPP-B-585, style 3, for class 3 use; or PPP-B-621, table III.

5.2.1.2 Case liner. Each level A box shall be lined with a waterproof case liner conforming to Specification MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto.

5.2.1.3 Strapping. Strapping, when required, shall be in accordance with applicable box specifications.

Note: Strapping will not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be constructed in accordance with Specification PPP-B-621, table I, or PPP-B-636, type I, class 1.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Unit package. Each unit package shall be marked with stock number, item identification, unit, quantity, contract number, and name of contractor.

5.3.2 Exterior container. Exterior container shall be marked in accordance with Standard MIL-STD-129, except as follows: The letter "L" shall precede the Federal stock number. One space shall be allowed between the letter "L" and the first digit of the stock number. The terms "Stock number" or "FSN" shall not be indicated. Service color and service symbol markings are required only on direct shipments from the contractor's plant to overseas activities.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Intended use. The back rest covered by this specification is intended to be used with Stock No. 6530-783-7905, Litter, Folding, Rigid Pole, Aluminum Pole.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) When sampling or testing of Government supplied material is required (see 4.1.3).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 Availability of color chips. The individual 3- by 5-inch chips of the color referenced in 3.2.5.3, are available from General Services Administration, Business Service Center, Region 3, Seventh and D Streets, S.W., Washington 25, D.C. Price 5 cents.

6.4 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

6.5 Transportation description. Transportation description applicable to this item is:

Back rests, invalids, folded in boxes.

Carload minimum weight 20,000 pounds subject to Rule 34 of the Uniform Freight Classification.

Truckload minimum weight 22,000 pounds per National Motor Freight Classification.

Notice. When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

Custodians:
Army-Med
Navy-BuMed
Air Force-AFCSG

Preparing activity:
Defense Medical Supply Center

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.