MIL-C-36044A
20 October 1967
SUPERSEDING
MIL-C-36044
21 April 1961

MILITARY SPECIFICATION

CHISEL, BONE (Smith-Petersen)

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers chisels (osteotomes) and case. Chisels shall be suitable for use in orthopedic surgery; case shall be suitable for storing chisels, for use in protecting the cutting edges of chisels, and to permit sterilization of the chisels.

1.2 Classification. Chisels shall be of the following types and sizes, as specified (see 6.1):

Type I. Straight
- Size 1. 1/4 inch wide
- Size 2. 1/2 inch wide
- Size 3. 3/4 inch wide
- Size 4. 1-1/4 inches wide

Type II. Curved
- Size 1. 3/8 inch wide
- Size 2. 5/8 inch wide
- Size 3. 1 inch wide

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:01Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, (for Carton Sealing).

Military

MIL-P-116	Preservation, Methods of

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWINGS

Defense Personnel Support Center (DPSC)

20005 Bone Chisels, Smith-Petersen and Case

(Copies of specifications, standards, and drawings, required by contractors in connection with specific procurement functions should be obtained as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material.

3.1.1 Chisels. Chisels shall be fabricated from class 4, type 420, 420F, 440A, 440B or 440C material, conforming to the requirements of GG-I-526. Type 420 or 420F shall have a carbon content of 0.30 percent minimum.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:01Z
Check the source to verify that this is the current version before use.

3.1.1.1 Hardness. After proper stress relieving the chisels shall have Rockwell hardness of C50 to C55.

3.1.2 Case. The case shall be fabricated from class 3, type 302, 303 or 304 material conforming to the requirements of GG-I-526.

3.1.2.1 Hardness. Rockwell hardness for the case shall be B75 minimum.

3.2 Style, design, and dimensions.

3.2.1 Chisels. Chisels shall be of the Smith-Petersen design, suitable for orthopedic surgery conforming to Drawing 20005.

3.2.2 Case. Case shall be designed to contain the chisels and permit sterilization of the chisels contained in the case. Case shall conform to Drawing 20005.

3.3 Finish.

3.3.1 Chisels. Chisels shall have satin finish as specified in GG-I-526.

3.3.2 Case. Case shall have a No. 4 or 2B commercial finish.

3.4 Identification marking. Each chisel and case made in U.S.A. shall be permanently marked with the name or registered trademark of the contractor, and the word "stainless". In addition, the handle of the chisel shall be stamped with its size (width of blade) in inches. Chisels and cases of foreign (other than U.S.) manufacture shall also be marked with the country of origin.

3.5 Additional requirements. Except as specified herein, the chisels and case shall comply with applicable requirements and tests of GG-I-526.

3.6 Workmanship. Workmanship shall be first class throughout. Chisels and case shall be free from defects which detract from their appearance or impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:01Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to insure that supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in the following table:

Table I. Sampling Criteria

Component	Characteristic	Req.	Test Procedure	Unit of Product	Insp level	AQL
Chisel and Case	Boil test (For all material)	GG-I-526	GG-I-526	1-chisel and 1-case	S-2*	**
Chisel	Copper Sulfate test (For class 4 material)	GG-I-526	GG-I-526	1-chisel	S-2*	**
Chisel and Case	Hardness	3.1.1.1 and 3.1.2.1	GG-I-526	1-chisel and 1-case	S-2*	**
Chisel and Case	Dimensions	3.2.1 and 3.2.2	-------	1-chisel and 1-case	S-2	4.0

*A minimum of three units shall be tested regardless of lot size.

**Acceptance number shall be zero for all sample sizes.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:01Z
Check the source to verify that this is the current version before use.

4.2.3 Raw materials. Four ounces of each lot of each metal used in construction of the chisel or case shall be selected for chemical analysis.

4.2.4 Test samples. Where possible, the same sample unit shall be used for the determination of two, or more test characteristics.

4.3 Examination. The chisel or case shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted as necessary, to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One chisel or one case, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in the following table shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table II. Packaging Quantities

Item	Unit package	Intermediate package	Exterior container
Chisel	1 unit	6 units	96 units
Case	1 unit	Not required	36 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A

5.1.3.1 Unit package (Chisel). Each unit shall be cleaned throughly and packaged in accordance with MIL-P-116, method 1A-15.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:01Z
Check the source to verify that this is the current version before use.

5.1.3.2 Unit package (Case). Each unit shall be packaged in a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2-inch wide tape conforming to PPP-T-76. Closing of bottom flap with adhesive conforming to MMM-A-250 or metal stitches is permitted.

5.1.3.3 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2-inch wide tape conforming to PPP-T-76. Closing of bottom flap with adhesive conforming to MMM-A-250 or metal stitches is permitted.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-636, type CF, class weather-resistant. Closure and strapping shall be in accordance with appendix of applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, with appendix of applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, constructed so as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:01Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Military requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:01Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:01Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements.

SPECIFICATION

MIL-C-3604A

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 68 1426

REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:01Z
Check the source to verify that this is the currentversion before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
AUG 66

☆ U.S. GOVERNMENT PRINTING OFFICE: 1968-301-515 3575

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:01Z
Check the source to verify that this is the current version before use.