MIL-C-36069A(DM)
3 November 1964
SUPERSEDING
MIL-C-36069(BuMed)
17 July 1961

MILITARY SPECIFICATION

CONDUCTOR, BONE CUTTING WIRE SAW, BAILEY

1. SCOPE

1.1 This specification covers one type of conductor (Bailey) for use in introducing a Gigli bone wire saw.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal form a part of this specification to the extent specified herein.

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.

Military

MIL-P-12902	Packaging of Surgical, Dental and Veterinary Instruments, Diagnostic Equipment, and Related Items.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Military Medical Supply Agency
20446 Conductor, Bone Cutting Wire Saw, Bailey

(Copies of specifications, standards and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:03Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Material. The conductor shall be fabricated of class 1, composition E steel conforming to the requirements of Specification GG-I-526, except that it shall be nickel plated only and that the hardness shall be Rockwell 45-N 54 to 58.

3.2 Style, design, and dimensions. The conductor shall be of the Bailey design, suitable for use in introducing a gigli bone wire saw. Style, design, and dimensions shall be in accordance with Drawing 20446.

3.3 Finish. The conductors shall have a bright finish.

3.4 Additional requirements. Except as specified herein, the conductors shal comply with the applicable requirements and tests of Specification GG-I-526.

3.5 Workmanship. Workmanship shall be first class throughout. Conductors shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the co[illegible]c or purchase order, the supplier is responsible for the performance of all inspe[illegible]io requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Governmen The Government reserves the right to perform any of the inspections set forth in th specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performanc of such testing by the contractor.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:03Z
Check the source to verify that this is the current version before use.

4.2 Sampling for lot acceptance.

4.2.1 Inspection lot. The inspection lot shall be determined in accordance with MIL-STD-105.

4.2.2 Sampling for examination. A random sample of conductors shall be selected from each lot offered to the Government in accordance with MIL-STD-105. Examination for dimensions shall be in accordance with MIL-STD-105, and as indicated in table I.

4.2.3 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in table I.

Table I. Sampling Criteria

Component	Characteristic	Rqmt.	Test procedure	Unit of product	Insp. level	AQL (percent defective)
Conductor	Hardness	GG-I-526	GG-I-526	1-conductor	S-2	*
	Boil test	GG-I-526	GG-I-526	1-conductor	S-2	*
	Salt spary test	GG-I-526	GG-I-526	1-conductor	S-2	*
	Dimensions	3.2	———	1-conductor	S-2	4.0

* Acceptance number shall be zero for all sample sizes.

4.2.4 Raw materials. Four ounces of each lot of metal used in the construction of the conductors shall be selected for chemical analysis.

4.2.5 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The conductors shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging and packing. Domestic or overseas.

5.1.1 Unit of issue. One (1) conductor, as specified, constitutes one unit of issue.

5.1.2 Level of protection. The required level of protection (level II, domestic or level III, overseas), for packaging and packing shall be as specified in the procurement document for the required number of units of issue.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:03Z
Check the source to verify that this is the current version before use.

5.2 Procedures. Packaging and packing for the specified level of protection shall be in accordance with the requirements of Military Specification MIL-P-12902, dated 20 July 1953 and Amendment-1, dated 25 August 1961, for the applicable procedure code number as indicated in table II. Commercial colors shall be acceptable on paperboard cartons when furnished as specified. In addition, the following revisions shall be effected.

In 3.4.8.2 of Military Specification MIL-P-12902, at end of paragraph, add 'As an alternate, exterior container shall be a carton of appropriate size construct in accordance with Federal Specification PPP-B-636, type CF, class weather-resistant. Strapping and closure of each fiberboard box shall be as specified in the appendix of Federal Specification PPP-B-636. In addition, waterproofing shall be as specified in 30.4.

In 3.4.8.2.1 of Military Specification MIL-P-12902, at end of paragraph, add 'Waterproof barrier bag shall not be required for fiberboard boxes.'

5.2.1 Quantities. The number of units of issue to be contained in each unit or intermediate package and exterior container shall be as indicated in table II.

Table II.

Level of protection	Procedure code No.	Packaging and packing quantities		
		Unit package	Intermediate package	Exterior container
II Domestic (Level B)	20	1 unit	12 units	576 units
III Overseas (Level A)	30			

5.3 Marking.

5.3.1 Interior packages. Interior packages shall be marked in accordance with Military Standard MIL-STD-129. When labels are utilized on cartons, waterproofing shall be required only when applicable cartons are fabricated from water-resistant material. Contract or purchase order number and name of contractor shall be shown. In addition, and as applicable, marking as required by 3.6.1.1 and 3.6.1.2 of Military Specification MIL-P-12902 shall also be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with Military Standard MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:03Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1).

6.2 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

Custodians:
Army. - MD
Navy. - MS
Air Force. - 03

Preparing activity:
Defense Medical Supply Agency. - DM

Review activities:
Army. - MD
Navy. - MS
Air Force. - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:03Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

FORM APPROVED BUDGET BUREAU NO. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use of the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity.

SPECIFICATION MILITARY SPECIFICATION Dated,
CONDUCTOR, BONE CUTTING WIRE SAW, BAILEY, MIL-C-36069A(DM), 3 November 1964

ORGANIZATION	CITY AND STATE

CONTRACT NO.	QUANTITY OF ITEMS PROCURED	DOLLAR AMOUNT
		$

MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?
   ☐ YES ☐ NO IF "YES" IN WHAT WAY?

REMARKS (ATTACH ANY PERTINENT DATA WHICH MAY BE OF USE IN IMPROVING THIS SPECIFICATION. IF THERE ARE ADDITIONAL PAPERS, ATTACH TO FORM AND PLACE BOTH IN AN ENVELOPE ADDRESSED TO PREPARING ACTIVITY)

SUBMITTED BY (PRINTED OR TYPED NAME AND ACTIVITY) DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:03Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold Under

DEFENSE SUPPLY AGENCY
DEFENSE MEDICAL SUPPLY CENTER
3RD AVENUE AND 29TH STREET
BROOKLYN, N.Y., 11232

OFFICIAL BUSINESS

POSTAGE FEES
DEFENSE

Defense Supply Agency
Defense Medical Supply Center
ATTN: Directorate of Technical Operations (Code T-H5)
3rd Avenue and 29th Street
Brooklyn, N.Y., 11232

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:03Z
Check the source to verify that this is the current version before use.