MIL-C-36136B
13 September 1968
SUPERSEDING
MIL-C-36136A
11 January 1968

MILITARY SPECIFICATION

CLAMP, UMBILICAL CORD

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a sterile, disposable, self-locking, plastic, umbilical cord clamp.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.

Military

MIL-L-10547	Liners, Case and sheet overwrap; Water-vaporproof, flexible.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:07Z
Check the source to verify that this is the current version before use.

STANDARDS

Federal

FED-STD-101	Preservation, Packaging and Packing, Materials. Test Procedures.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWINGS

Defense Personnel Support Center (DPSC)

20597 Clamp Umbilical Cord, Plastic, Disposable, 2-$\frac{1}{4}$", 125s

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publication. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

U.S. Pharmacopeial Convention, Incorporated

Application for copies should be addressed to

Pharmacopeia of the United States
Mack Publishing Company, Easton, Pa.

3. REQUIREMENTS

3.1 Material.

3.1.1 Clamp. The clamp shall be fabricated from an autoclavable, inert, non-toxic, non-irritating, nylon plastic.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:07Z
Check the source to verify that this is the current version before use.

3.1.2 Clipper. The clipper shall be fabricated from chromium over nickel-plated carbon steel or as an alternate, the clipper shall be fabricated of acetal plastic, with a nickel plated brass eyelet and stainless steel blade.

3.1.3 Immediate container. The immediate container shall be fabricated from a trifluorochloroethylene plastic, or plastic consisting of an oriented polyester film having a heat-sealable, polyethylene-type surface, firmly and adherently bonded to one side thereof, without the use of any intervening adhesive layer, or of paper coated with polyethylene plastic, or any combination thereof (see 3.2).

3.2 Style, design, and construction. The umbilical cord clamp shall be a sterile, disposable clamp, sealed in an individual tamperproof container which shall maintain the sterility of its contents until opened. The sealed container shall be so designed as to permit ready dispensing of the clamp by sterile techniques, and shall withstand the leakage test specified in 4.4.1 without leaking. Each unit package of 125 clamps shall be supplied with one clipper (see 3.2.2).

3.2.1 Clamp. Style, design, and dimensions of the clamp shall be in accordance with DPSC Drawing No. 20597. The clamp shall be self-locking, and once closed shall open only when cut at the hinge. When closed, the teeth on the jaws of the clamp shall intermesh to provide positive sealing of the umbilical cord of a new born infant, and remain securely locked and under tension as the umbilical cord shrinks. The finished clamp shall be free from flash, roughness, sharp edges, sprue or mold marks, or other detrimental defects. The clamp shall weight $2\text{-}\frac{1}{2} \pm \frac{1}{4}$ grams.

3.2.2 Clipper. Each unit package of 125 clamps shall be supplied with one clipper. The clipper shall be toenail nipper, or other similar device, that will readily cut through the clamp, at the hinge, using hand pressure.

3.3 Manufacturer's record. The manufacturer shall place a lot control number on the clamp unit of issue (125 clamps), which can be verified against the manufacturer's records, of all tests made available for inspection by any duly qualified medical officer of the Armed Forces, National Institutes of Health, or any official of the Federal Food and Drug Administration. No shipment shall be made until all of the tests have been completed, verified, and recorded in the permanent records of the manufacturer.

3.4 Sterilization. When tested as specified in 4.4.2 the clamps shall not become distorted, cracked, or adversely affected so as to impair their use.

3.5 Sterility. The clamp shall comply with sterility test requirements of the current edition of the Pharmacopeia of the United States (USP), when tested as specified in 4.4.3.

Source: https://assist.dla.mil - Downloaded: 2026-10-21T08:07Z
Check the source to verify that this is the current version before use.

3.5.1 Certificates of sterility, listing all sterility lot numbers, stating that samples from each sterility lot have been tested for sterility and found to be sterile, shall be furnished by the contractor. Certificates shall specify the tests performed and the results obtained. The certificates of sterility shall be submitted by the contractor directly to Technical Operations Division, Directorate of Medical Materiel, ATTN: Quality Control Branch, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101.

3.6 Marking.

3.6.1 Identification marking. Each clamp shall be marked with the contractor's name or registered trademark.

3.6.2 Immediate container. The immediate container shall be marked "sterile", together with directions for removal and use of the clamp.

3.7 Instructions. Each unit package shall be provided with instructions, including illustrations, describing the proper technique for application and removal of the umbilical cord clamp.

3.8 Workmanship. Workmanship shall be first class throughout. Clamps shall be free from any defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:07Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 Unit of product. The unit of product for inspection purposes only shall consist of one clamp.

4.2.2 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection level and acceptability level (AQL) shall be as indicated in table I. Unit of product shall be as indicated in 4.2.1.

Table I Sampling for Examination

Component	Type defect	Inspection level	AQL(percent defective)
Clamp	Major A	S-4	1.0
	Major B	S-1	1.0
	Minor A	S-4	6.5
	Minor B	S-1	4.0

4.2.3 For tests.

4.2.3.1 For sterility test. Sampling for sterility test shall be as indicated for Liquids and Suspensions under Sterility Tests in the current edition of the U.S.P.

4.2.3.2 For other tests. Sampling for other tests shall be conducted in accordance with MIL-STD-105 and Table II.

Table II. Sampling for Tests

Component	Characteristic	Reqt.	Test procedure	Unit of product	Insp. level	AQL
Clamp	Leakage test	3.2	4.4.1	1-clamp	S-1	*
	Sterilization	3.4	4.4.2	1-clamp	S-1	*

*Acceptance number shall be zero for all sample sizes.

4.2.4 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The clamps shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted as specified in tables I and III. Inspection, however, shall not be restricted to defects listed in table III. Any clamp in the sample containing one or more defects shall be rejected and, if the number of defective clamps in any sample exceeds the acceptance number for that sample, the lot represented by the sample, shall be rejected.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:07Z
Check the source to verify that this is the current version before use.

Table III. Classification of defects in accordance with MIL-STD-105

Categories	Defects
CRITICAL:	None defined
Major A	
101	Clamp does not lock, does not lock properly, or breaks or cracks upon locking.
102	Clamp opens readily after being locked.
103	Poor meshing of clamps jaw serrations.
104	Clamp not free from flash, roughness, sharp edges, sprue or mold marks or other detrimental defects.
105	Immediate container fails to open readily.
106	Immediate container not free of holes, tears, cuts, punctures, or open seals.
107	Marking on immediate container missing, not legible, not correct, not complete.
Major B	
151	Dimensions of clamp not as specified.
152	Weight of clamp not as specified.
Minor A	
201	Contractor's name or registered trademark on clamp missing, not legible, not correct, not complete.
Minor B (Unit of product - Unit package and contents)	
251	Average count per unit package not a minimum of 125 clamps.
252	Clipper missing or inoperative.
253	Clipper finished poorly (i.e.) sharp edges, roughness, incomplete plating, flaking of plating.

4.4 Tests. Tests shall be conducted, as necessary, to determine compliance with specification requirements (see 4.2.3).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:07Z
Check the source to verify that this is the current version before use.

4.4.1 Immediate container leakage test. The immediate container shall be submitted to the vacuum chamber technique test as specified in Test Method 241 of Federal Test Method Standards No. 101.

4.4.2 Sterilization test. Clamp shall be steam autoclaved for 20 minutes at 250 degrees Farenheit (^{o}F) and 15 pounds per square inch gage (psig), then allowed to cool to room temperature (68°F to 72°F).

4.4.3 Sterility test. Sterility test shall be performed as described under Sterility Tests in the current edition of the U.S.P.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One package (125 clamps and 1 clipper), as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in the following table shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	4 units	8 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit packaging.

5.1.3.1.1 Immediate container. Each clamp shall be individually packaged as specified in 3.1.3 and 3.2. Clipper (see 3.1.2) shall be packaged in a carton of appropriate size and design.

Source: https://assist.dla.mil - Downloaded: 2026-10-21T08:07Z
Check the source to verify that this is the current version before use.

5.1.3.1.2 Unit package. Each unit shall be packaged in a set up box of appropriate size and design. Instructions for use as specified in 3.7 shall be furnished. Closure shall be adequate to prevent spilling of contents under normal handling.

5.1.3.2 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class domestic, Closure shall be adequate to prevent spilling of contents under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table III, using type I, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure and strapping of wood boxes shall be as specified in the appendix of the applicable box specification. Each fiberboard box shall be closed and waterproofed as specified in the appendix of PPP-B-636.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table II, using type III, class 1 plywood; or shall be designed for type 2 load and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; of PPP-B-636, type CF, class domestic. Closure shall be in accordance with appendix of applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:07Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Immediate container. Immediate container shall be marked as specified in 3.5.2. In addition, packing shall include the legend: "Sterility guaranteed until package is opened or damaged" or similar statement.

5.3.2 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown. Lot (control) number shall be shown.

5.3.3 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Lot (control) number shall be shown.

6. NOTES

6.1 Intended use. The clamp is used on the umbilical cord of a new born infant and will remain securely locked and under tension until the umbilical cord shrinks.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packing (See 5.1 and 5.2).

6.3 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Military Government requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Category N

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:07Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements.

SPECIFICATION

MIL-C-36136B

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 54, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:07Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL CODE: ATS
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
AUG 66

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:07Z
Check the source to verify that this is the current version before use.