MIL-D-36105(DMSC)
21 March 1962

MILITARY SPECIFICATION

DARKROOM, X-RAY, PORTABLE, TENT TYPE, LIGHTWEIGHT, FIELD

1. SCOPE

1.1 This specification covers an X-ray darkroom tent set for use in the field.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

T-R-605	Rope; Manila and Sisal.
T-R-616	Rope; Mildew-Resistant.
V-F-106	Fasteners, Slide Interlocking.
V-T-276	Thread, Cotton.
QQ-A-267	Aluminum Alloy Bars, Rods, and Shapes, Extruded, 2024
QQ-A-268	Aluminum Alloy Bars, Rods, and Wire; Rolled, Drawn, or Cold-Finished, 2024.
QQ-A-270	Aluminum Alloy; Bars, Rods and Shapes, Extruded; 6061 and 6062.
QQ-A-318	Aluminum Alloy Plate and Sheet 5052.
QQ-A-601	Aluminum Alloy, Sand Castings.
QQ-W-461	Wire, Steel, Carbon, (Round, Bare and Coated).
TT-E-527	Enamel, Alkyd, Lustreless.
TT-E-529	Enamel, Alkyd, Semi-Gloss.
UU-P-271	Paper, Wrapping, Waterproofed Kraft.
CCC-C-419	Cloth, Cotton, Duck, Unbleached, Plied-Yarns (Army and Numbered).
CCC-D-950	Dyeing and Aftertreating Processes for Cotton Fabrics.
CCC-T-191	Textile Test Methods.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-00640	Boxes, Corrugated, Triple Wall, 350-Pound Maximum Weight.
PPP-C-96	Cans, Metal, 28 Gage and Lighter.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, Water Resistant (for Carton Sealing).

FSC 6545

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

Military

MIL-A-101	Adhesive, Water-Resistant, for Sealing Fiberboard Boxes.
MIL-F-495	Finish, Chemical, Black, for Copper Alloys.
JAN-C-496	Clips, End.
MIL-W-530	Webbing, Textile, Cotton, General Purpose, Natural or in Colors.
MIL-B-543	Buckles, Tongueless and Web Strap.
JAN-P-735	Primer, Paint, Zinc-Chromate, Alkyd Type.
MIL-S-1734	Slips, Tent Line.
MIL-P-2383	Pins, Tent, Wood.
MIL-T-3530	Treatments, Mildew Resistant and Water Repellent for Thread and Twine.
MIL-R-5001	Rubber; Latex Foam Sponge.
MIL-C-5540	Cement, Neoprene.
MIL-T-7807	Thread, Nylon.
MIL-A-8625	Anodic Coatings, for Aluminum and Aluminum Alloys.
MIL-F-10884	Fasteners, Snap.
MIL-S-11699	Snap, for Equipage and Leather Items.
MIL-G-16491	Grommet, Metallic.
MIL-C-8514	Coating Compound, Metal Pretreatment, Resin-Acid.

STANDARDS

Air Force - Navy Aeronautical

AN257	Hinge-Continuous.
AN526	Screw, Machine, Truss Head.

Federal

Fed. Test Method Std. No. 151	Metals; Test Methods.
Fed. Std. No. 595	Colors.
Fed. Std. No. 751	Stitches, Seams, and Stitchings.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.
MS16686	Medical Services Symbol.
MS20470	Rivet-Universal Head, Aluminum Alloy.

DRAWINGS

Armed Services Medical Procurement Agency

Darkroom, X-Ray, Portable, Tent Type, Lightweight, Field, consisting of:

10134 (3 sheets), covers: Tent Frame.
10135 (2 sheets), covers: Tent Covering.
10136 (6 sheets), covers: Exhaust Fan.

(Copies of specifications, standards, and drawings required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids shall apply.

COLOR ASSOCIATION OF THE UNITED STATES, INC.

United States Army Standard Color Card for the Official Standardized Shades of Sewing Threads.

United States Army Color Card of Standard Shades for Slide Fastener Tapes.

(Application for copies should be addressed to the Color Association of the United States, Inc., 200 Madison Avenue, New York 16, N.Y.)

MILITARY CLOTHING AND TEXTILE SUPPLY AGENCY

Interim Purchase Description IP/DES S-243-1 titled: Dyeing and Aftertreating Processes for Cotton Fabrics CCC-D-950(1).

(Application for copies should be addressed to Executive Director, Military Clothing and Textile Supply Agency, Philadelphia Quartermaster Center, U. S. Army, 2800 South 20th Street, Philadelphia 1, Pa. Attn: Specifications Branch, Storage Division.)

UNDERWRITERS' LABORATORIES, INC.

Standard 507 Fans, Electric.

(Application for copies should be addressed to the Underwriters' Laboratories, Inc., 161 Sixth Avenue, New York 13, N.Y., 207 East Ohio Street, Chicago 11, Illinois, or 1655 Scott Lane, Santa Clara, Calif.)

3. REQUIREMENTS

3.1 Fire and casualty hazards.

3.1.1 The contractor shall submit to the contracting agency proof, prior to shipment, that the equipment he will supply under this specification conforms to the applicable requirements of the Underwriters' Laboratories, Inc. Standard 507. Such proof shall consist of a laboratory report or listing by the Underwriters' Laboratories, Inc.

3.1.2 As an alternate, the contractor may submit proof satisfactory to the contracting agency that his equipment conforms to the published standards or applicable requirements, including methods of tests of the Underwriters' Laboratories, Inc. Such proof shall include the following:

(a) Laboratory report by a qualified independent laboratory acceptable to the contracting agency which shall include test results and methods used to determine conformance with published Underwriters' Laboratories Standards or applicable requirements established and used by the Underwriters' Laboratories in testing the same or similar items.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

(b) The Government reserves the right to require a facilities report showing the qualifications of the testing laboratory and to conduct a facilities survey to determine its acceptability. Laboratory reports of tests shall be submitted to the contracting agency and approval obtained prior to shipments.

3.1.3 Compliance with requirements regarding fire and casualty hazards does not absolve the contractor from complete compliance with the other requirements of this specification in order to assure the acceptance of his commodity.

3.2 Component parts. Each darkroom set consists of a tent frame, a tent covering, an exhaust fan, a canvas carrying case for the tent frame and tent covering, a metal carrying case for the exhaust fan, four guy ropes and four tent pins for the tent frame, and a patching kit for the tent covering.

3.3 Materials. Except as specified herein, materials shall conform to the applicable portions of the specifications, standards, and drawings (with bill of materials) listed in section 2, and shall be capable of meeting the requirements shown therein. In addition, the metals shall conform to the chemical requirements of their applicable specifications when tested as specified in 4.4.7. Where materials to be used are not specifically designated, those used shall be of the highest quality suitable for the purpose intended. Where commercial components are specified, it shall be interpreted to mean that similar components produced by other manufacturers may be substituted, provided that the item furnished is interchangeable with the specified item, and that all the established functional requirements will be met.

3.3.1 Tent frame.

3.3.1.1 Aluminum alloy.

3.3.1.1.1 Pipe. Pipe shall be 3/4 inch Iron Pipe Size, (ips) and shall be aluminum-alloy 63S-temper T6.

3.3.1.1.2 Castings. Aluminum alloy castings shall conform to class 17M of Specification QQ-A-601.

3.3.1.1.3 Rod. Aluminum rod shall be aluminum-alloy 6061-temper T6, conforming to Specification QQ-A-270.

3.3.1.1.4 Sheet. Floor plates shall be of aluminum-alloy sheet conforming to Specification QQ-A-318, and shall be temper H34.

3.3.1.2 Steel.

3.3.1.2.1 Retaining screws and coupling pins. The pole retaining screws and coupling pins shall be fabricated from 18-8 corrosion-resisting steel.

3.3.1.2.2 Carriage bolts and self-locking nuts. Carriage bolts and self-locking nuts shall be fabricated from 18-8 corrosion-resisting steel.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

3.3.2 Carrying case, canvas.

3.3.2.1 Body fabric. The body fabric, before finishing, shall be ply yarn cotton duck conforming to Specification CCC-C-419, for type I, No. 6, hard texture, as shown in table I of the same specification.

3.3.2.1.1 Color and treatment. Cotton duck shall be vat dyed olive drab No. 7, conforming to type I, class B of Interim Purchase Description IP/DES S-243-1, except that the cloth shall be made mildew-resistant by treatment limited to two mildew-inhibitors only: (a) 2,2' methylene-bis (4-chlorophenol), or (b) copper-8-quinolinolate, in accordance with the requirements as shown in table III of the same specification (see 4.4.1.1).

3.3.2.2 Webbing. Except as specified herein, webbing for the shoulder strap and handles shall conform to Specification MIL-W-530 and shall be type II, 1 inch wide, or type IIb, 2 inches wide, as applicable, and shall be class 4 in conformance with the requirements of the same specification, except that the maximum amount of copper-8-quinolinolate shall be 1.28 percent. The webbing shall be vat dyed olive drab No. 7 (see 4.4.1.2).

3.3.2.3 Thread, cotton. Except as specified herein, the thread for the stitching on the canvas case shall conform to Specification V-T-276, and shall be type IIIb, ticket No. 10/4 as shown in table VI, and shall be vat dyed olive drab shade S-1, conforming to U.S. Army Standard Color Card. All references to "approved shade standard", "approved standard shade sample", and "approved standard shade" made in Specification V-T-276 shall not apply. Thread shall be treated for mildew-resistance in accordance with Specification MIL-T-3530, using type I, class 1 mildew-inhibiting agent (see 4.4.1.3). The finished thread shall show fair colorfastness to laundering, chlorine bleaching, and weather. Paragraph titled: "Certification" in Specification V-T-276 shall not apply.

3.3.2.4 Grommets with washers. Grommets with washers shall conform to Specification MIL-G-16491 for type III, class 3, size 5.

3.3.2.5 Clips, end. End clips shall be ball type, brass, 1 inch, conforming to type I of Specification JAN-C-496. Finish shall be as specified in Specification MIL-F-495.

3.3.2.6 Buckles. Buckles shall conform to type II, style 4, class 1 or 2, 1 inch size, as specified in Specification MIL-B-543. Buckles shall be given a bronze finish as specified in 3.4.1 of the same specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

3.3.2.7 Keeper. Keeper shall be made from steel wire 0.1483-inch-diameter. Steel wire shall be Nos. 1015 through 1020, conforming to Specification QQ-W-461. The keeper shall be given a finish consisting of electro-plated zinc followed by a phosphate surface treatment and bake-enameled with one coat of lusterless black enamel as specified in 3.3.5.9. Keeper shall be baked in accordance with the paint manufacturer's instructions.

3.3.2.8 Snap. Snap shall conform to class 2 of Specification MIL-S-11699.

3.3.2.9 Fasteners, snap. Snap fasteners shall conform to style 2, construction "A", finish 2 of Specification MIL-F-10884.

3.3.3 Guy ropes.

3.3.3.1 Rope. Rope shall conform to Specification T-R-605, type S, and shall be 3/8 inch in diameter. The rope shall be vat dyed olive drab No. 7, and mildew-resistant.

3.3.3.1.1 Mildew-resistance. The rope shall be treated for mildew-resistance in accordance with Specification T-R-616. The mildew-resistance shall be obtained by the use of copper-8-quinolinolate as specified in 7.2.1.1 of Specification T-R-616 (see 4.4.2).

3.3.3.1.2 Slips, tent. Tent slips shall conform to type I of Specification MIL-S-1734.

3.3.3.1.3 Hooks, snap. Snap hooks shall be North and Judd No. 125, or equal, and shall have a 5/8-inch eye. Hooks shall have commercial black enamel finish.

3.3.3.2 Tent pins. Tent pins shall conform to type II of Specification MIL-P-2383.

3.3.3.3 Metal pretreatment coating. Metal pretreatment coating shall conform to Specification MIL-C-8514. As an alternate material, the coating may be a polyvinyl chloride vehicle with zinc-chromate in a two-part system to produce an acid-type material. It shall meet the performance requirements specified in the following paragraphs of Specification MIL-C-8514, namely, 3.3 Manufacture; 3.5.5 Compatibility; 3.5.6 Mixing and application properties; 3.5.7 Metal and coating anchorage; and 3.5.8 Smoothness and appearance. The material shall be tested as specified in section 4 of Specification MIL-C-8514 (see 6.4).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

3.3.4 Tent covering.

3.3.4.1 Tent fabric. The tent fabric shall consist of five layers, as follows:

Layer 1	Neoprene compound, pigmented lusterless black.
Layer 2	Nylon cloth, bias.
Layer 3	Neoprene compound.
Layer 4	Nylon cloth, straight.
Layer 5	Neoprene compound, pigmented lusterless black.

Before coating, the nylon cloth shall have a nominal weight of 1.35 ounces per square yard, and shall have a minimum breaking strength of 85-pounds-per-inch in both warp and fill direction. The nylon shall be boiled off and heat set. The coated, finished fabric, shall have a weight of 8 (± 10 percent) ounces per square yard and shall have an average minimum breaking strength, employing the grab method, of 180 pounds in warp and 140 pounds in fill. The ply adhesion shall be not less than 4-pounds-per-inch strip, and the maximum permeability to hydrogen shall be 10 liters. The tear strengh using the strip method, shall be not less than 7 pounds in the warp and 7 pounds in the filling. The diaphragm bursting strength of the fabric shall be not less than 300-pounds-per-square-inch. The neoprene coating shall have sufficient flexibility so that it will not crack, break, or become tacky when the fabric is subjected to test temperatures of minus 67° Fahrenheit (F.), and plus 170° F. When tested as specified, the fabric shall withstand a static head of 20 inches of water for 10 minutes, with leakage of no more than 5 milliliters (ml.) of water through the abraded portion. There shall be no signs of leakage in the unabraded portion. When exposed to accelerated weathering, the fabric shall lose not more than 25 percent of its original breaking strength, and shall not crack when folded sharply on itself. The fabric shall be opaque and free from rips, tears, and holes before and after accelerated weathering. Tests on fabric shall be performed as specified in 4.4.4.

3.3.4.2 Tape, seam-sealing. The seam-sealing tape shall consist of four layers, as follows:

Layer 1	Neoprene compound, pigmented, 1.2 oz. sq. yd.
Layer 2	Nylon cloth, bias as specified in 3.3.4.1.
Layer 3	Neoprene compound, pigmented.
Layer 4	Neoprene compound, uncured.

The combined weight of layers 3 and 4 shall be 2.5 oz. sq. yd. The pigmented neoprene compound shall be as specified in 3.3.4.1. The uncured neoprene compound shall be as required to make strong joints when the tape is cemented over the seams, that is, shall be capable of curing when in contact with the cement.

3.3.4.3 Cement, seam-sealing. Cement used for sealing seams shall be entirely suitable for the purpose, and shall be capable of producing strong joints, which shall not be adversely affected when the tent is used as intended. Cement for sealing seams shall conform to type I or type III of Specification MIL-C-5540.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

3.3.4.4 Collar. The fabric for the collar shall be the same as specified in 3.3.4.1 for the tent fabric.

3.3.4.5 Cement, patching. The patching cement shall be an air-curing neoprene cement having a long shelf life, and suitable for cementing layers of tent fabric as specified in 3.3.4.1 together in a strong joint, which shall not be adversely affected when the tent is used as intended. The cement shall conform to type II of Specification MIL-C-5540.

3.3.4.6 Thread, nylon. Thread for sewing the tent covering shall be nylon in accordance with Specification MIL-T-7807 for type II, class 1, size E. Thread shall be black in color.

3.3.4.7 Fiber, vulcanized. Vulcanized fiber shall be commercial type. The fiber shall be encased in the tent covering material for use as the arms of the airports.

3.3.4.8 Fasteners, slide, interlocking.

3.3.4.8.1 Lighttight fastener. The lighttight fastener shall be B. F. Goodrich, pressure-sealing, heavy-duty fastener, closed both ends, type 400, or equal. Slider shall be on the cotton tape side only. The tape shall be vat dyed olive drab shade S, conforming to U.S. Army Color Card of Standard Shades for Slide Fasteners, and shall be water-repellent and mildew-resistant. Mildew-resistance shall be accomplished as specified in 3.6.2.1 of Specification MIL-W-530, except that the maximum amount of copper-8-quinolinolate shall be 1.28 percent (see 4.4.3.2). Brass parts shall be given a black oxidized finish.

3.3.4.8.1.1 Water-repellence. The finished tape shall be tested as specified in 4.4.3.1 and shall not absorb more than 25 percent of water based on weight of tape.

3.3.4.8.1.2 Salt spray resistance. Slide fasteners shall meet the salt spray resistance requirements specified in Specification V-F-106.

3.3.4.8.2 Conventional slide fasteners. The conventional slide fasteners shall conform to Specification V-F-106 for type I, style 2, size MH, except as follows:

(a) Elements or scoops shall be of nylon and shall be injection-molded onto the tape.
(b) Tape shall be twill or plain woven monofilament nylon fabric with selvage on each edge. A nylon bead shall run along one edge and shall be either woven integrally or securely sewn on.
(c) Slider may be an enamelled zinc-alloy die casting.
(d) Crosswise strength shall be 150 pounds minimum. The tape shall be vat dyed olive drab No. 7, and the scoops and slider shall be colored to blend with the tape.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

3.3.4.9 Lift fasteners. Lift fasteners shall be large curtain-type fastener, and shall be United-Carr Fastener Co., Inc.'s "Mud-Proof" Lift fasteners, cap No. 16503, socket No. 16204, stud No. 16343, washer No. 744A, or equal. Material shall be brass with black chemical finish conforming to Specification MIL-F-495.

3.3.4.10 Fasteners, snap. Snap fasteners shall conform to style 2, construction "A", finish 2 of Specification MIL-F-10884.

3.3.4.11 Webbing, cotton. Webbing shall be as specified in 3.3.2.2, except that webbing shall be type IIa, 1 inch wide and type IIa, 1-1/4 inches wide, as applicable, in the locations shown on Drawing 10135.

3.3.4.12 Grommets with washers. Grommets with washers shall conform to Specification MIL-G-16491 for type III, class 3, size No. 4.

3.3.4.13 Clips, end. End clips shall conform to Specification JAN-C-496, and shall be type I, 1-inch brass, or 1-1/4-inch brass, as required. Finish shall be as specified in Specification MIL-F-495.

3.3.4.14 Buckles. Buckles shall be 1-inch steel web strap buckles, North and Judd No. 3105, or equal. Buckles shall be given a commercial black enamel finish.

3.3.4.15 Rubber, latex foam. Foam rubber shall conform to Specification MIL-R-5001 for type II, medium class, grade C.

3.3.4.16 Rope. Rope shall be as specified in 3.3.3.1, except that it shall be 1/4-inch diameter. Three 12-1/2-foot lengths of rope shall be provided with each covering. One end of each length shall be eye-spliced, the other end shall be suitably treated or tied to prevent fraying.

3.3.5 Exhaust fan.

3.3.5.1 Aluminum, sheet. Aluminum-alloy sheet shall conform to Specification QQ-A-318 and shall be temper H32 or H34, as specified.

3.3.5.2 Aluminum, rod. Aluminum rod shall conform to Specification QQ-A-267 or QQ-A-268, and shall be temper T4.

3.3.5.3 Wire cloth. Wire cloth shall be steel, galvanized after weaving, 2-by-2 mesh.

3.3.5.4 Sealer. Sealer shall be black weatherstrip adhesive, Minnesota Mining and Mfg. Co.'s, or equal.

3.3.5.5 Fasteners and hardware. Screws, rivets, latch, hinges, etc., shall be as shown on Drawing 10136.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

3.3.5.6 Electrical components. The fan motor, fan blade, extension cord with motor plug and receptacle, male caps, and line switch, multiple outlet cord with cap, and four outlet cluster receptacle shall conform to the commercial references given on Drawing 10136, or equal. The motor shall be a totally enclosed unit-bearing motor, 1500 revolutions per minute, 115 volt, 50/60 cycle, 9 watt. The electrical power cord shall be a 2-conductor, No. 18 wire, type SJ.

3.3.5.7 Metal pretreatment. Metal pretreatment coating shall be as specified in 3.3.3.3.

3.3.5.8 Primer. Primer shall be as specified in 3.3.3.4.

3.3.5.9 Enamel, lusterless. Olive drab enamel shall be as specified in 3.3.3.5. Lusterless black enamel shall conform to Specification TT-E-527. The color of the final finish shall be lusterless black conforming to color 37038 of Federal Standard No. 595 (see 6.3).

3.3.5.10 Enamel, black, semi-gloss. Semi-gloss enamel shall conform to class B of Specification TT-E-529. The color of the final finish shall be semi-gloss enamel, conforming to color 27038 of Federal Standard No. 595(see 6.3).

3.4 Construction.

3.4.1 Tent frame, carrying case, guy ropes, and pins.

3.4.1.1 Tent frame. The frame shall be constructed in strict accordance with the requirements shown herein and to those shown on Drawing 10134. Construction shall be sound and durable in all respects.

3.4.1.1.2 Finish.

3.4.1.1.2.1 Aluminum alloy parts. The aluminum-alloy parts of the tent frame shall be finished with enamel as specified herein. All aluminum alloy surfaces shall be thoroughly cleaned and chemically degreased.

3.4.1.1.2.2 Anodizing. Prior to assembly of couplings, aluminum-alloy parts of couplings shall be anodized in accordance with Specification MIL-A-8625.

3.4.1.1.2.3 Metal pretreatment coating. All aluminum alloy surfaces, except as specified in 3.4.1.1.2.2, shall be given a film of pretreatment coating as specified in 3.3.3.3., and applied in accordance with the manufacturer's instructions. The dry film shall be 0.2 to 0.4 mils thick.

3.4.1.1.2.4 Primer. After pretreatment, a primer coat having a minimum dry film thickness of 1.0 mil shall be applied. Primer shall comply with 3.3.3.4. The primer shall be sprayed on and oven-baked in accordance with the paint manufacturer's recommendations.

3.4.1.1.2.5 Enamel. The primer coat shall be followed by one coat of lusterless olive drab enamel as specified in 3.3.3.5, minimum dry film thickness of 1.0-1.5 mils, sprayed on, and oven-baked by a heat treatment equivalent to baking at 250° F., for not less than 45 minutes. The resultant finish shall be hard, smooth, and highly resistant to abrasion, cracking, and chipping.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

3.4.1.1.2.6 Other metal parts. Corrosion-resisting pole retaining screws shall be given a touch up coat of enamel as specified in 3.4.1.1.2.5.

3.4.1.2 Carrying case. The carrying case shall be "duffel-bag" style, partitioned in the center and constructed in strict accordance with the requirements shown herein and to those shown on Drawing 10134, except that the construction of the case shall be modified by moving the longitudinal seam from the side of the bag as shown in section A-A, sheet 3 of Drawing 10134 to the opposite side.

3.4.1.2.1 Stitches, seams, and stitching. All seams and stitchings shall be straight, conform to Federal Standard No. 751, and consist of stitch type 301 stitches formed by machine, with 7 to 8 evenly spaced stitches per inch. The ends of each continuous row of stitches (including ends due to thread breaks), shall be anchored by backstitches for a length of not less than 1 inch, except where anchorage is provided by another seam or stitching. Thread tension shall be properly maintained so that there will be no loose stitching, and so that the lock will be imbedded in the middle of the material sewed.

3.4.1.2.2 Setting of grommets. Holes punched to receive the grommets shall be smaller than the outside diameter of the grommet barrel so that the barrel must be forced through the hole. The grommets shall be tightly clinched without cutting the material.

3.4.1.3 Guy ropes. Each frame shall be provided with four (4) guy ropes. Each rope shall have a net length of 12 feet overall. One end of the rope as specified in 3.3.3.1, shall be eye-spliced to the swivel eye of the snap hook as specified in 3.3.3.1.3, the other end shall be properly threaded through the tent slip as specified in 3.3.3.1.2, and eye-spliced to the eye of the tent slip.

3.4.1.4 Tent pins. Each frame shall be provided with four (4) tent pins as specified in 3.3.3.2.

3.4.2 Tent covering. Construction of the tent covering shall be strict accordance with the requirements shown herein and to those shown on Drawing 10135 except that on sheet 2 of Drawing No. 10135, section FF, delete "33-1/2 inch" dimension and substitute "32-1/2 inch"; Section GG, delete "37-1/2 inch" dimension and substitute "36-1/2 inch". Completed tent covering assembled on the tent frame including the exhaust fan as specified in 3.4.4.1, shall form a lighttight tent suitable for processing uncovered X-ray film when the exterior of the tent is exposed to full daylight. Lighttightness shall be tested as specified in 4.4.6.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

3.4.2.1 Stitches, seams, and stitching. All seams and stitchings shall be straight, conform to Federal Standard No. 751, shall consist of stitch type 301 stitches formed by machine, using the smallest practical needle size so as to minimize the hole size. There shall be 7 or 8 evenly spaced stitches per inch. The ends of each continuous row of stitches (including ends due to thread breaks), shall be anchored by backstitches for a length of 1 inch, except where anchorage is provided by another seam or stitching. Thread tension shall be maintained so that there will be no loose stitching, and so that the lock will be imbedded in the middle of the material sewed. Each seam and stitching shall be covered with seam-sealing tape as specified in 3.3.4.2, shall be cemented in place by seam-sealing cement as specified in 3.3.4.3, and applied so as to exclude air bubbles and prevent loose edges. Reinforcing patches of tent covering material shall also be cemented in place by seam-sealing cement as specified in 3.3.4.3, and applied so as to exclude air bubbles and prevent loose edges.

3.4.2.2 Snap fasteners and grommets. The snap fasteners specified in 3.3.4.10, and the grommets specified in 3.8.4.12, shall be affixed in locations shown on the drawing, and shall be tightly clinched without cutting the fabric and so as not to be detached from the fabric in normal use. Washers of suitable olive drab colored webbing or duck shall be employed in conjunction with fasteners if more body to the material is required. When holes are made in the fabric before insertion of the fastener tubes or grommet barrel, such holes shall be small enough to require that the tubes or barrels be forced through. The male snap fasteners shall be accurately aligned with their respective female snap fasteners so as to afford a neat, secure closure.

3.4.2.3 Web straps. Web straps for securing the foam-rubber collars to the passbox or exhaust fan, and web strap towel loops, shall be made of webbing as specified in 3.3.2.8, and shall be constructed as shown on the drawing. Each strap shall have a buckle as specified in 3.3.4.14, or an end clip as specified in 3.3.4.13, as applicable. End clips shall be firmly attached without cutting the webbing, which shall be inserted the full depth and shall be flat at the point of entry. All raw ends of webbing shall be turned under and double stitched.

3.4.3 Patching supplies. With each tent covering, there shall be supplied one piece of tent fabric as specified in 3.3.4.1, not less than 30 inches wide by 2 yards long, and two 1-pint cans of patching cement as specified in 3.3.4.5. Two (2) cans of accelerator, each of adequate capacity to hold the necessary accelerator, for 1-pint of patching cement shall be furnished. The cans shall conform to Specification PPP-C-96 for type V, class 4, rectangular shape with metal screwcap closure. The can shall be 1-pint size, and shall measure 3-13/16 by 1-11/16 inch-base by 5-7/8 inches high. Tolerance shall be ± 1/8 inch. Minimum base weight tinplate or terneplate shall be 85-pounds-per-base-box. The tinplate shall have a nominal tin coating weight of 0.50 pound or greater. Paragraph on "Interior coatings" shall apply; Plan "B" of "Exterior nonmetallic coatings" shall apply, except that the color shall conform to No. 24087 as shown in Federal Standard No. 595 (see 6.3). The smaller cans for the accelerator shall be 2, 4, or 6 ounce nominal liquid capacity size, dimensions optional, as required. Each can shall be properly marked to identify the contents as specified in 7.4.1.2 of Specification PPP-C-96. Cans made of terneplate shall bear the legend as specified in the same paragraph. Complete instructions for use of the patching supplies shall be stored in the compartment provided in the exhaust fan carrying case as specified in 3.4.5. There shall also be supplied a suitable small metal or plastic measuring device, marked so as to accurately measure the amounts of material necessary to mix the two parts of the cement.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

3.4.4 Exhaust fan and carrying case.

3.4.4.1 Exhaust fan. The exhaust fan shall be constructed in strict accordance with the requirements specified herein and to those shown on Drawing 10136, except that an alternate method of constructing the body of the housing for the exhaust fan, part 1, sheet 2, Drawing 10136 may be employed, whereby the housing may be made of two equal pieces of the specified aluminum-alloy sheet instead of one piece as shown. The assembled housing shall have two seams, each of the type shown on the drawing, and located on opposite sides of the housing.

3.4.4.1.1 Assembly. Assembly of the exhaust fan shall be as follows:

(a) Construct assembly consisting of fan housing, internal light baffles (spotwelded in place), and hinged hanger (spotwelded in place).

(b) Fabricate all remaining exhaust fan metal components.

(c) Finish all metal surfaces as follows:

All surfaces shall be cleaned and chemically degreased.

All surfaces shall be given a metal pretreatment coating as specified in 3.3.3.3, and applied in accordance with the manufacturer's instructions. The dry film thickness shall be 0.2 - 0.4 mils.

After pretreatment, one coat of primer, as specified in 3.3.3.4, having a dry film thickness of 0.6 - 0.8 mils, shall be applied. Primer shall be sprayed on and oven-baked in accordance with paint manufacturer's recommendations.

The primer coat shall be followed by one coat of lusterless black enamel, as specified in 3.3.5.9, minimum dry thickness 0.6 - 0.8 mils, sprayed on and oven-baked at 250°F., for not less than 45 minutes. Resultant finish shall be smooth, hard, and highly resistant to abrasion, cracking, and chipping.

(d) Continuously seal the joint between the light baffles and the fan housing to eliminate all light leaks. Use the sealer specified in 3.3.5.4.

(e) Rivet in place the front baffle, and attach front cover (without motor plug). Do not install motor and fan blade.

(f) Finish all exterior surfaces with semi-gloss black enamel as specified in 3.3.5.10, baked in accordance with the manufacturer's instructions. Dry film thickness shall be 0.6 - 0.8 mils.

(g) Remove front cover and install motor, fan blade, motor plug, and connect wiring from motor to motor plug. Attach front cover.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

3.4.5 Exhaust fan carrying case. The exhaust fan carrying case shall be constructed in strict accordance with the requirements specified herein and to those shown on Drawing 10136, except that an alternate method of constructing the body of the exhaust fan carrying case, part 15, sheet 6, Drawing 10136, may be employed whereby the body may be made of two equal pieces of the specified aluminum-alloy sheet instead of one piece as shown. The assembled body shall have two seams, each of the type shown on the drawing, and located on opposite sides of the body.

3.4.5.1 Finish. Finish of the carrying case shall be cleaned and degreased, given a metal pretreatment coat, primed, and finished with lusterless olive drab enamel as specified in 3.4.4.1.

3.5 Identification marking.

3.5.1 Tent covering. Centrally located on the visible surface of the flap designed as "Flap 2" on the drawing, there shall be permanently and legibly marked in a suitable dark red color, in block letters 3/4 inch high, the following:

(a) 6545-664-4811
(b) Darkroom, X-Ray, Portable, Tent Type, Lightweight, Field.
(c) Manufacturer's name or registered trade-mark.
(d) Contract number or order number.

On the same surface there shall also be located a 2 inch diameter Medical Services Symbol as shown for type I in Standard MS16686.

Centrally located on the visible surface of the flap, above the entrance, on the outside of the tent, there shall be permanently and legibly marked in red, in block letters 1 inch high, the following legend:

"Darkroom, X-Ray".

3.5.2 Carrying bag, canvas. On the exterior surface of the canvas carrying bag, centrally located with respect to the two body straps and the two carrying handles, in block letters 3/4 inch high, on one side of the bag, there shall be permanently and legibly marked in black, the following:

6545-664-4811

Darkroom, X-Ray, Portable, Tent Type, Lightweight, Field.

Part 1 of 2

Inconspicuously located on the same surface there shall be marked the manufacturer's name or registered trade-mark, and the contract number or order number.
On the opposite side of the bag, centrally located as described above, there shall appear a 3-inch diameter Medical Services Symbol as shown for type I in Standard MS16686.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

On the interior of the carrying bag, the two flaps covering the compartments of the bag shall be marked as follows:

With the flaps in the closed position, the upper surface of the flap which may be closed by means of snap fasteners, shall be marked with the legend: "Tent Covering" in block letters ½ inch high; the upper surface of the other flap shall be marked with the legend: "Frame" in the block letters ½ inch high.

3.5.3 Exhaust fan carrying case. Centrally located on the exterior surface of the hinged lid of the exhaust fan carrying case, there shall be permanently and legibly marked in black (3.3.5.9), the following:

A 2 inch diameter Medical Services Symbol as shown for type I in Standard MS16686.

In block letters, ½ inch high:-

6545-664-4811

Darkroom, X-Ray, Portable, Tent Type, Lightweight, Field.

Exhaust Fan and Repair Kit.

Part 2 of 2.

On the inner surface of the lid there shall be located the manufacturer's name or registered trade-mark, and the contract number or order number in block letters ¼ inch high.

3.6 Welding. Welding shall be done by qualified welders using appropriate welding equipment in a qualified process. Spotwelds shall be capable of developing the tensile strength of the parent metal. All exposed welds shall be smoothed sufficiently to blend with the adjacent surfaces after the finish shall have been applied. All other welds shall be finished in a clean and workmanlike manner. All welds shall be clean, nonporous, and sound.

3.7 Instructions. A complete set of instructions for erecting and for taking down the tent frame and the tent covering (complete with exhaust fan), shall be provided with each X-ray darkroom. The instructions shall be clearly and legibly printed on good quality paper, and shall be permanently cemented with a waterproof cement to the inner surface of the lid of the exhaust fan carrying case.

3.8 Assembly. Prior to final preparation for delivery as specified in section 5, the X-ray darkroom, shall be assembled as follows:

(a) In carrying case, canvas. The completed tent covering shall be tightly rolled, after the 3 lengths of rope intended to be used for lashing the tent covering to the frame has been laced through the grommets in the covering to prevent loss, and inserted into the compartment marked "Tent Covering". The flap shall then be secured.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

(b) In the compartment marked "Frame", insert the frame, guy ropes (complete), and the tent pins.

(c) The exhaust fan carrying case shall contain the exhaust fan, extension cords, and repair kit.

3.9 Workmanship. Workmanship shall be first class throughout. All completed component parts (frame, tent covering, exhaust fan, carrying case, etc.) shall be clean, well made, and free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Supplier responsibilities for inspection. The supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own or any other inspection facilities and services acceptable to the Government. Inspection records of the examination and tests shall be kept complete and available to the Government as specified in the contract or order. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with Standard MIL-STD-105.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with Standard MIL-STD-105, unless otherwise specified below:

	Inspection level	AQL (percent defective)	Minimum sample size
Component material testing	I2	*	3
End item testing			
Lighttightness	-	**	-

* Acceptance number shall be zero (0) for all sample sizes.
** Each completed tent shall be tested for lighttightness as specified in 4.4.6.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

4.2.3 For chemical analysis of metals. Four ounces of each lot of each metal used in the construction of the components for the tent frame, exhaust fan, and carrying case for the fan shall be selected for chemical analysis as specified in 4.4.7.

4.2.4 Component materials. Sufficient samples of the component materials shall be selected from each lot of material used in the manufacture of the darkroom to perform tests as required by the applicable subsidiary specifications or by tests specified herein. When sample size for component materials covered by the subsidiary specifications is not specified, the sample size for each test shall be as indicated in 4.2.2. The unit of product for testing shall be the quantity of material necessary to perform all required tests one time each, taking into consideration the fact that, where possible, the same material may be used to perform two or more tests. The units of product are to be expressed in multiples of the following:

Canvas	Yards or fractions of a yard.
Tent fabric	Yards or fractions of a yard.
Seam sealing tape	Yards or fractions of a yard.
Webbing	Yards or fractions of a yard.
Thread	Cone or spool.
Rope	Yards or fractions of a yard.
Grommets and washers	Each.
End clips	Each.
Buckles	Each.
Keepers	Each.
Snaps	Each.
Snap fasteners	Each (male and female elements).
Tent slips	Each.
Snap hooks	Each.
Cement	1 pint of patching cement plus accelerator.

4.3 Examination. The darkroom shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted as specified herein, and as necessary, to determine compliance with specification requirements.

4.4.1 Material testing.

4.4.1.1 Cloth, cotton, duck. The unfinished duck shall be tested in accordance with the test procedures of Specification CCC-C-419 (see 3.3.2.1). The finished duck shall be tested in accordance with the test procedure of Interim Purchase Description IP/DES S-243-1, except that performance of the tests for mildew-inhibitor content and mildew-resistance shall be conducted under the following conditions:

(a) If the material conforms to the requirements for mildew-inhibitor content, the mildew-resistance tests shall not be performed, and the material shall be considered as complying with mildew-resistance requirements.
(b) If the mildew-inhibitor content exceeds the maximum requirement, the material shall be rejected.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

(c) If the mildew-inhibitor content is found to be below the minimum requirement, the material may be accepted with respect to mildew-resistance provided it passes the mildew-resistance test, method 5762, as specified in the test section of Interim Purchase Description IP/DES S-243-1. When the mildew-resistance test is performed, the leaching process shall be omitted, and the breaking strength shall be determined by method 5100 of Specification CCC-T-191.

4.4.1.2 Webbing. The webbing shall be tested in accordance with the test procedures of Specification MIL-W-530, except that the performance of the tests for mildew-inhibitor content and mildew-resistance shall be conducted under the following conditions:

(a) If the material conforms to the requirements for mildew-inhibitor content, the mildew-resistance tests shall not be performed, and the material shall be considered as complying with mildew-resistance requirements.

(b) If the mildew-inhibitor content exceeds the maximum requirement, the material shall be rejected.

(c) If the mildew-inhibitor content is found to be below the minimum requirement, the material may be accepted with respect to mildew-resistance provided it passes the mildew-resistance test.

4.4.1.2.1 Mildew-resistance. The finished webbing shall be tested in accordance with method 4762 as specified in the test section of Specification CCC-T-191. The leaching process shall be omitted.

4.4.1.2.2 Water absorption. The finished webbing shall be tested for water absorption by method 5500 as specified in the test section of Specification MIL-W-530, except that the size of the specimen and the size of the blotting paper shall conform to method 5502 of Specification CCC-T-191.

4.4.1.3 Thread. The untreated thread shall be tested in accordance with the test procedures of Specification V-T-276 (see 3.3.2.3). The treated thread shall be tested as specified in the test section of Specification MIL-T-3530.

4.4.1.3.1 Colorfastness to laundering. The finished thread shall be tested for colorfastness to laundering by method 4610 of Specification CCC-T-191.

4.4.2 Rope. The untreated rope shall be tested in accordance with the test procedures of Specification T-R-605. The treated rope shall be tested as specified in the test section of Specification T-R-616. If the rope conforms to the requirements for mildew-inhibitor content, the tests specified in 4.4.2.1 shall not be required and material shall be considered as complying with the mildew-resistance requirements. The material shall be rejected if the mildew-inhibitor content is found to exceed the maximum requirements. If the mildew-inhibitor content is found to be below the minimum requirements, the material may be accepted with respect to mildew-resistance provided it passes the tests as specified in 4.4.2.1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

4.4.2.1 Mildew-resistance of rope. The finished rope shall be tested by either the direct innoculation, sterilized specimen method specified in Specification T-R-616 or by the 14-day soil-burial method 4762 as specified in Specification CCC-T-191. The finished rope shall show not more than 15 percent loss in breaking strength. The leaching process shall be omitted.

4.4.3 Slide fasteners.

4.4.3.1 Water-repellence. The finished tape shall be tested for water-repellence by method 5500 of Specification CCC-T-191. The specimens shall conform to 2.3 and the blotting paper to 3.1.1 of method 5502 of the same specification.

4.4.3.2 Mildew-inhibitor content. The copper content of the finished slide-fastener tape shall be determined by either the electrolytic method or the polarographic method specified in test section of Specification MIL-W-530. If the material conforms to the requirements for mildew-inhibitor content, the test specified in 4.4.3.3 shall not be required, and the material shall be considered as complying with mildew-resistance requirements. The material shall be rejected if the mildew-inhibitor content is found to exceed the maximum requirements. If the mildew-inhibitor content is found to be below the minimum requirements, the material may be accepted with respect to mildew-resistance provided it passes the test requirements in 4.4.3.3.

4.4.3.3 Mildew-resistance. The finished slide-fastener tape shall be subjected to the 14-day soil-burial test by method 4762 of Specification CCC-T-191, except that the leaching process shall be omitted. The test specimen shall include the chain. At the end of the test, the loss in breaking strength of the slide fastener crosswise shall not exceed 10 percent of the initial breaking strength.

4.4.3.4 Salt spray resistance. The slide fasteners shall be tested in conformance with 4.4.14 of Specification V-F-106.

4.4.4 Tests on tent fabric. The methods of testing the tent covering fabric shall be as follows:

Test	Test Method
Weight	Method 5041, Specification CCC-T-191
Breaking strength	Method 5100, Specification CCC-T-191
Tearing strength	Method 5134, Specification CCC-T-191
Flexibility	
(a) Elevated temperature	Method 5872, Specification CCC-T-191
(b) Low temperature	A 1-by-4 inch specimen with the long dimension fillingwise, shall be exposed for 4 hours at a temperature of 67° F., and then bent sharply over a 1/8 inch rod.
Ply adhesion	Method 5950, Specification CCC-T-191
Permeability	Method 5460, Specification CCC-T-191
Bursting strength	Method 5122, Specification CCC-T-191

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

Waterproofness	Method 5516, Specification CCC-T-191, except water-temperature shall be 80° F. ± 10° F. Before the specimen is placed in the apparatus, one side shall be abraded with grade 1/0 garnet paper cemented to an 8 ± .1 ounce metal block 2 inches square. The specimen shall be laid on a flat surface, face up, and the garnet paper and weight placed upon the face of the specimen. The abradant shall be grasped with the thumb and finger at the base of the block and moved gently across the specimen fillingwise, back and forth so as to abrade the specimen five (5) strokes in each direction. The abraded area shall be placed face up across the center line of the clamping head so that the center of the abraded area shall coincide with the center of the exposed part of the specimen. The water height shall be raised to 20 inches and leakage noted.
Accelerated weathering	Method 5804, Specification CCC-T-191, exposure time 150 ± 0.5 hours, minimum conditioning time after exposure, 24 hours. One specimen, warpwise and one fillingwise, shall be folded by hand sharply upon itself and visually inspected for cracking. Other specimens shall be tested for breaking strength in accordance with method 5100, Specification CCC-T-191. Test specimens for accelerated weathering test shall consist of six (6) specimens, 4 by 6 inches minimum length warpwise, and six (6) specimens, 4 by 6 inches minimum length fillingwise.
Opacity	Fabric shall be opaque and free from pinholes when examined by means of a 40 watt Fish-Schurman Corp. "Zirconarc" photomicrographic lamp, or equal.

4.4.5 Tests on tape, seam-sealing. The methods of testing the tape shall be as follows:

Test	Test Method
Weight	Method 5041, Specification CCC-T-191
Breaking strength	Method 5100, Specification CCC-T-191

4.4.6 Lighttightness. Each tent shall be set up outdoors. The exhaust fan shall be installed and operated while the outside of the tent is exposed to bright sunlight. X-ray film strips shall be partially uncovered within the tent and shall remain so for a period of 5 minutes, and then covered. Each strip shall be not less than 6 inches long by 1 inch wide, and approximately half of each strip shall be uncovered. Sufficient strips shall be employed to test the suitability of the entire tent enclosure. The film strips shall then be processed in the normal manner and examined for any evidence that uncovered portions have been struck by light.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

4.4.7 Chemical analysis. The samples selected in accordance with 4.2.3 [illegible] be analyzed in conformance with method 111 or 112 of Federal Test Method Standard No. 151. In case of dispute, method 111 shall be used for reference analysis.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) darkroom, as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Unit preparation. Exhaust fan, extension cords, and repair kit shall be cushioned within the exhaust fan carrying case in such manner as to prevent shifting or damage. Exhaust fan carrying case shall be wrapped in flexible corrugated fiberboard or packaged in a corrugated fiberboard carton of suitable size and design. Wrapping or carton shall be further wrapped in waterproofed paper conforming to class C-1 of Specification UU-P-271 and sealed with water-resistant adhesive.

5.1.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose within the continental limits of the United States of America.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. The exhaust fan carrying case and canvas carrying bag (1 unit) shall be overpacked in an exterior container designed for type 2 load and constructed in accordance with Specification PPP-B-585, style 3, for class 3 use; PPP-B-621, table III; or PPP-B-00640, class 2, style E. When boxes conforming to Specification PPP-B-00640 are furnished, top and bottom flaps shall be securely closed with a water-resistant adhesive conforming to the requirements of Specification MIL-A-101. The adhesive shall be applied to the entire area of contact between the flaps. Center seams shall be closed by means of tape conforming to the requirements of Specification PPP-T-76. The tape shall be not less than 3 inches wide and shall extend at least 2 inches onto each of the end panels.

5.2.1.2 Strapping. Strapping, when required, shall be in accordance with applicable box specification.

Note: Strapping will not be required for shipment forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, [illegible] the lowest rate, to point of delivery called for in the contract or purchase [illegible]er.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Exterior container. Exterior container shall be marked in accordance with Standard MIL-STD-129, except as follows: The letter "L" shall precede the Federal stock number. One space shall be allowed between the letter "L" and the first digit of the stock number. The terms "Stock number" or "FSN" shall not be indicated. Service color and service symbol markings are required only on direct shipments from the contractor's plant to overseas activities.

6. NOTES

6.1 Intended use. The portable, tent type, X-ray darkroom, is intended to provide darkroom facilities for developing X-ray films in the field.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 Availability of color chips. The individual 3-by 5-inch color chips of the colors referenced in 3.3.3.5, 3.3.5.9, 3.3.5.10, and 3.4.3, are available from General Services Administration, Business Service Center, Region 3, Seventh and D Streets, Washington 25, D. C. Price 5 cents.

6.4 The metal pretreatment coating of the Cloverleaf Paint and Varnish Corporation, Long Island City, New York, its No. 25 zinc-chromate primer and No. 25A reducer, is considered to meet the requirements of the alternate material in 3.3.3.3.

6.5 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

6.6 Transportation description. Transportation description applicable to this item is:

Reproduction equipment, photographic or transfer, military, in boxes.

Carload minimum weight 24,000 pounds subject to Rule 34 of the Uniform Freight Classification.

Truckload minimum weight 20,000 pounds in accordance with National Motor Freight Classification (A Series).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

Notice. When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

Custodians:
Army-Med
Navy-BuMed
Air Force-AFCSG

Preparing activity:
Defense Medical Supply Center

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.