MILITARY SPECIFICATION

FORCEPS, GALL DUCT, CURVED (LAHEY)

This Amendment forms a part of Military Specification MIL-F-36054A, dated 10 October 1972, and is approved for use by all Departments and Agencies of the Department of Defense.

Pages 4 and 5:

4.2.2 For tests. Delete Table II entirely and notes under table, and substitute:

"Table II. Sampling Criteria for Tests

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL
Forceps	Hardness	3.2	GG-I-526	S-2	1.5
	Dimensions	3.4	---	S-2	4.0
	Boil test	GG-I-526	GG-I-526	S-2	1.0
	Copper sulfate test	GG-I-526	GG-I-526	S-2	1.0
	Box lock stiffness	3.6	GG-I-526	S-2	2.5
	Distortion test	GG-I-526	GG-I-526	S-2	2.5

."

Add new paragraph:

"4.2.3 Raw materials. A sufficient quantity of each lot of metal used in the construction of the forceps shall be selected for chemical analysis."

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:02Z
Check the source to verify that this is the current version before use.

Page 8:

5.1.3.2 Preservative, Delete "Not required" and substitute: "Preservative shall not be applied".

6.3 Delete title and substitute: "National Stock Number (NSN) coverage." Delete "Federal Stock Number 6515-334-1400", and substitute: "National Stock Number 6515-00-334-1400."

Custodians:

Army - MD
Navy - MS
Air Force - 03

Review activities:

Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1882

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:02Z
Check the source to verify that this is the current version before use.