MIL-N-36129(DMSC)
11 October 1962

MILITARY SPECIFICATION

NEBULIZER, MEDICINAL, PLASTIC

1. SCOPE

1.1 This specification covers a plastic medicinal nebulizer, for use in the vaporization of a solution of epinephrine or similar medications in the treatment of asthma by aerosol therapy.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-Up, Paperboard.

Military

MIL-L-10547	Liners, Case, Waterproof.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:06Z
Check the source to verify that this is the current version before use.

DRAWING

Defense Medical Supply Center

20561 Nebulizer, Medicinal, Plastic.

(Copies of specifications, standards, and drawing required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material.

3.1.1 Body, vapor tube, and jet. Body, vapor tube, and jet shall be of either clear propionate or clear polystyrene plastic, suitable for the purpose intended.

3.1.2 Caps, stoppers, plugs, and straps. Caps, stoppers, plugs, and straps shall be of polyethylene or polystyrene plastic suitable for the purpose intended.

3.1.3 Gasket. Gasket shall be fabricated from sheet rubber or neoprene, suitable for the purpose intended.

3.1.4 Bulb. Bulb shall be fabricated from rubber or a plastic material, suitable for the purpose intended.

3.2 Style, design, and construction. Style, design, and construction shall be in accordance with the drawing. The nebulizer shall be suitable for use with Stock Nos. 6505-117-5200, EPINEPHRINE INHALATION, USP, 5 cc, and 6505-299-9661, ISOPROTERENOL HYDROCHLORIDE INHALATION, USP, 1:200 Solution, 10 cc. Stopper and vent bulb shall be joined by a plastic retainer strap. Vent plug shall be permanently affixed to the nebulizer in such a manner that it may be opened or closed, but cannot be removed. Nebulizer shall be supplied with an extra jet.

3.3 Instruction sheet. An instruction sheet which shall provide operating and cleaning instructions shall be supplied with each nebulizer.

3.4 Case. Case shall be of a suitable plastic and shall be so designed as to accommodate the nebulizer, extra jet, and instruction sheet.

3.5 Identification marking. Nebulizer shall be permanently and legibly marked with the manufacturer's name or registered trade-mark.

3.6 Workmanship. Workmanship shall be first class throughout. Nebulizers shall be free from defects which detract from their appearance or may impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Supplier responsibilities for inspection. The supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own or any other inspection facilities and services acceptable to the Government. Inspection records of the examination and tests shall be kept complete and available to the Government as specified in the contract or order. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling. Sampling for examination shall be conducted in accordance with Standard MIL-STD-105.

4.2.1 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The nebulizer shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted, as necessary, to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) nebulizer, as specified, constitutes one unit of issue.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table I. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	12 units	144 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit, with or without bulb attached, together with spare jet, carry case and operating and cleaning instructions (see 3.2, 3.3, and 3.4) shall be packaged in a container of appropriate size. Container design shall include suitable interior fittings to insure against shifting of, or damage to, the contents. Closure shall be adequate to prevent spillage of contents under normal handling.

5.1.3.2 Intermediate package. Intermediate package shall be a carton of appropriate size constructed in accordance with Specification PPP-B-566, PPP-B-636, or PPP-B-676. Closure shall be adequate to prevent spillage of contents under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose within the continental limits of the United States of America.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with Specification PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; or PPP-B-621, table III.

5.2.1.2 Case liner. Each level A box shall be lined with a waterproof case liner conforming to Specification MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:06Z
Check the source to verify that this is the current version before use.

5.2.1.3 Strapping. Strapping, when required, shall be in accordance with applicable paragraphs of appendix thereto.

NOTE: Strapping will not be required for shipment forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to Specification PPP-B-601, table I, using type III, class 1 plywood, or shall be designed for type 2 load, and constructed in accordance with Specification PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type I, class 1.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Intermediate package. Each unit package and each intermediate package shall be marked in accordance with Standard MIL-STD-129. Contract number and name of contractor shall be shown.

NOTICE. When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

Custodian:
Army - Med
Navy -BuMed
Air Force- AFCSG

Preparing activity:
Defense Medical Supply Center

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

6.2 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

6.3 Transportation description. Transportation description applicable to this item is:

Atomizers or Nebulizers, in boxes.

Carload minimum weight 24,000 pounds subject to Rule 34 of the Uniform Freight Classification.

Truckload minimum weight 24,000 pounds per National Motor Freight Classification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.