MIL-R-36128A
17 DECEMBER 1969
SUPERSEDING
MIL-R-36128 (DMSC)
4 OCTOBER 1962

MILITARY SPECIFICATION

RACK, HEMOSTATIC CLIP (MC KENZIE)

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers a hemostatic clip holding rack for use with McKenzie clips and clip introducing forceps.

2. APPLICABLE DOCUMENTS

* 2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

GG-I-526	Instruments, Dental and Surgical; General Specifications.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.

Military

MIL-P-116	Preservation, Methods of.
MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.

MIL-STD-129 Marking for Shipment and Storage.

DRAWINGS

Defense Personnel Support Center (DPSC)

20536 Rack, Hemostatic Clip, McKenzie

(Copies of specifications, standards and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

* 3.1 General requirements. The hemostatic clip rack shall be in accordance with the applicable requirements of GG-I-526 and as specified herein.

* 3.2 Material. The hemostatic clip rack shall be fabricated from the following material in accordance with GG-I-526: base and rack shall be class 6, type 430 stainless steel; cover shall be class 3, type 302 or 303 stainless steel; and the screws and machined parts shall be a suitable stainless steel.

* 3.3 Hardness. After proper stress relieving the Rockwell hardness for the base, rack and cover shall be B75 to B85 and the Rockwell hardness for the screws and machined parts shall be B90 to B100.

* 3.4 Style, design and dimensions. The style, design, and dimensions of the hemostatic clip rack shall conform to the Defense Personnel Support Center (DPSC) Drawing No. 20536 Revision A, dated 17 December 1969.

3.5 Finish. The finish shall be mirror, or mirror and stain in accordance with GG-I-526.

* 3.6 Corrosion resistance. Shall be in accordance with GG-I-526.

* 3.7 Identification marking. Shall be in accordance with GG-I-526. Instruments of foreign manufacture shall be marked "stainless."

* 3.8 Workmanship. Shall be in accordance with GG-I-526.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

* 4.1 Responsibility for inspection. Shall be in accordance with GG-I-526.

4.1.1 Inspection. Shall be in accordance with GG-I-526.

* 4.1.2 Certificates of quality. Certificates of quality supplied by the manufacturer of the stainless steel may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the item identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

* 4.2 Sampling.

* 4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in Table I. The unit of product for examination purposes shall be one hemostatic clip rack.

Table I Sampling for Examination

Type of Defect	Inspection Level	AQL (Percent Defective)
Major	II	1.0
Minor	II	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

Table II Sampling for Tests

Component (unit of product)	Characteristic	Requirement	Test Procedure	Insp. Level	AQL (Percent Defective)
1-Rack	Hardness	3.3 and GG-I-526	GG-I-526	S-2	*
1-Rack	Dimensions	3.4	---	S-1	*
1-Rack	Copper sulfate **	3.6 and GG-I-526	GG-I-526	S-2	*
1-Rack	Boil test	3.6 and GG-I-526	GG-I-526	S-2	*

* Acceptance number shall be zero for all sample sizes.
** Shall apply for Type 430 only.

4.2.3 Raw materials. Four ounces of each lot of each metal used in construction of the hemostatic clip rack shall be selected for chemical analysis.

4.2.4 Sample unit usage. Where possible, the same sample unit shall be used for determination of two or more test characteristics.

4.3 Examination. Examination shall be in accordance with GG-I-526.

4.3.1 Classification of defects. Examination shall be conducted in accordance with MIL-STD-105 and the classification of defects specified in Table III. Inspection, however, shall not be restricted to the classified possible defects listed in Table III.

Table III Classification of Defects

Categories	Defects
Critical	None defined.
Major	
101	Rack not complete.
102	Style and design not as specified.
103	Rack not free of sharp edges, rough areas, burrs, cracks, grind marks, or overlaps.
Minor	
201	Identification marking not present, correct, complete, or legible.
202	Finish not as specified.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

* 4.4 Tests. Tests shall be conducted to determine compliance with specification requirements in accordance with GG-I-526.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One rack, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not required	48 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

* 5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

* 5.1.3.2 Preservative. Not required.

* 5.1.3.3 Intimate wrap. Each unit shall be wrapped in neutral tissue or other suitable anticorrosive materials.

* 5.1.3.4 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method IA-15.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

* 5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3. Closure and strapping shall be as specified in the appendix of the applicable box specification.

* 5.2.1.2 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.2 Level B.

* 5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Unit package. Each unit package shall be marked in accordance with MIL-STD-129. Contract or purchase order number and name of contractor shall be shown.

* 5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

* 6.1 Intended use. The hemostatic clip rack shall be suitable for holding McKenzie clips and for use with introducing hemostatic clip forceps during surgical procedures.

* 6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
(c) Federal stock number.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

* 6.3 This specification is intended to cover the following item appearing in the Federal Supply Catalog: FSN 6515-350-6000, RACK, HEMOSTATIC CLIP, McKenzie. It does not include all types, sizes, or grades as indicated by the title or all types which are commercially available.

* 6.4 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

* 6.4 Supersession data. This specification includes the requirements of Purchase Description No. 6, dated 6 February 1968.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Reveiw information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:06Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS. This sheet is to be filled out by personnel, either Government or contractor, involved in the use of this specification in procurement of products to be purchased by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements.

SPECIFICATION

MIL-R-36128A FSN

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place number and date of document in block 1.)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1426

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:06Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE [illegible]
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.