MILITARY SPECIFICATION

RULE, MEASURING, OPHTHALMOLOGICAL, PLASTIC, 17 CM.

1. SCOPE

1.1 This specification covers an ophthalmological rule for use in measuring eyes and eyeglasses.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-Up, Paperboard.

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications and standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:07Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Material. Material shall be opaque white plastic material of the highest quality throughout, and suitable in all respects for the purpose intended.

3.2 Style, design and dimensions. The measuring rule shall be ophthalmological type and shall be in accordance with figure 1. One surface shall be permanently and legibly graduated in millimeters on one edge and graduated in inches on the opposite edge. The graduated surface shall be covered with a permanent, clear transparent plastic coating to prevent obliteration of the graduations.

3.3 Identification marking. The rule shall be clearly and legibly marked with the manufacturer's name or registered trademark.

3.4 Workmanship. Workmanship shall be first class throughout. Rules shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facility or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling. Sampling for examination shall be conducted in accordance with Standard MIL-STD-105.

4.3 Examination. The rules shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted, as necessary, to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:07Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) rule, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit, intermediate and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table I. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	6 units	288 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be packaged in an envelope of appropriate size and design. Envelope shall be adequately secured to prevent accidental opening.

5.1.3.2 Intermediate package. Intermediate package shall be a carton of appropriate size constructed in accordance with Specification PPP-B-566, or PPP-B-676. Commercial colors shall be acceptable when carton conforming to Federal Specification PPP-B-566 or PPP-B-676 is utilized. Closure shall be adequate to prevent spillage of contents under normal handling.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:07Z
Check the source to verify that this is the current version before use.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with Specification PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; or PPP-B-621, table III.

5.2.1.2 Case liner. Each level A box shall be lined with a waterproof case liner conforming to Specification MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto.

5.2.1.3 Strapping. Strapping, when required, shall be in accordance with applicable paragraphs of appendix thereto.

NOTE: Strapping will not be required for shipment forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to Specification PPP-B-601, table I, using type III, class 1 plywood, or shall be designed for type 2 load, and constructed in accordance with Specification PPP-B-585, style 3, for class 2 use; PPP-B-621, table III; or PPP-B-636, type I, class 1.

5.3 Marking.

5.3.1 Interior package. Each unit and intermediate package shall be marked in accordance with Standard MIL-STD-129. Contract or purchase order number and name of contractor shall be shown. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material.

5.3.2 Exterior container. Exterior container shall be marked in accordance with Standard MIL-STD-129. In addition, the letter "L" shall precede the Federal stock number. One space shall be allowed between the letter "L" and the first digit of the stock number. The terms "Stock number" or "FSN" shall not be indicated. Service color and service symbol markings are required only on direct shipments from the contractor's plant to overseas activities.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:07Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

6.3 Transportation description. Transportation description applicable to this item is:

Rulers or Rules, plastic, NOI, in boxes.

Carload minimum weight: 30,000 pounds, per Uniform Freight Classification.

Truckload minimum weight: 30,000 pounds, per National Motor Freight Classification.

Notice. When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

Custodians:
Army-Med
Navy-BuMed
Air Force-AFCSG

Preparing activity:
Defense Medical Supply Center

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:07Z
Check the source to verify that this is the current version before use.




FIG. 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:07Z
Check the source to verify that this is the current version before use.