# MILITARY SPECIFICATION

# SYRINGE AND CANNULA, LARYNGEAL, LUKENS

## 1. SCOPE

1.1 This specification covers one size of laryngeal syringe and cannula of the Lukens design.

## 2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

| | |
|---|---|
| PPP-B-566 | Boxes, Folding, Paperboard. |
| PPP-B-585 | Boxes, Wood, Wirebound. |
| PPP-B-601 | Boxes, Wood, Cleated-Plywood. |
| PPP-B-621 | Boxes, Wood, Nailed and Lock-Corner. |
| PPP-B-636 | Box, Fiberboard. |
| PPP-B-676 | Boxes, Set-Up, Paperboard. |

Military

| | |
|---|---|
| MIL-L-10547 | Liners, Case, Waterproof. |

STANDARDS

Military

| | |
|---|---|
| MIL-STD-105 | Sampling Procedures and Tables for Inspection by Attributes. |
| MIL-STD-129 | Marking for Shipment and Storage. |

DRAWING

Military Medical Supply Agency

| | |
|---|---|
| 20295 | Syringe and Cannula, Laryngeal, Lukens, 2 cc. |

(Copies of specifications, standards, and drawing required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:05Z
Check the source to verify that this is the current version before use.

## 3. REQUIREMENTS

### 3.1 Material.

3.1.1 Syringe. All metal parts of the syringe shall be fabricated from nickel-plated brass. The glass shall be suitable for the purpose intended and shall be capable of being sterilized by boiling (see 3.3).

3.1.2 Cannula. The cannula proper (tube) shall be fabricated from nickel silver tubing, silver lined inside. The hub shall be fabricated of brass. The entire cannula (including hub and cannula proper) shall be silver plated after assembly.

3.1.3 Wood case. The wood case shall be fabricated from a suitable hardwood

3.2 Style, design, and dimensions. The unit shall consist of a 2 cubic centimeter (cc.) laryngeal syringe and cannula of the Lukens design. Style, design, and dimensions shall be in accordance with the drawing.

3.2.1 Case. Each unit shall be supplied in a hardwood case equipped with interior fittings to insure against shifting or breakage of the syringe and cannula. The case shall be provided with a hinged lid and a suitable latch to prevent accidental opening. The case shall be smoothly finished with a suitable lacquer or varnish.

3.3 Boil test. When tested as specified in 4.4.1, the barrel shall show no cracks, chipping, or any signs of separation of the metal parts from the glass tube. In addition, the graduation marks shall not show any loss of pigment.

3.4 Fluid tightness test. When tested as specified in 4.4.2, there shall be no leakage in excess of a falling drop at any point in the syringe and cannula.

3.5 Accuracy. When tested as specified in 4.4.3, syringes shall deliver plus or minus 5 percent of the capacity indicated at each graduation mark tested.

3.6 Finish. Metal parts shall have a smooth, bright finish. The inside surface of the glass tube shall be finished to a very fine, uniform surface, free from pits, grind marks, or open air lines. The outside of the glass tube shall have a smooth clear surface.

3.7 Identification marking. Each syringe and cannula shall be permanently and legibly marked with the contractor's name or registered trade-mark.

3.8 Workmanship. Workmanship shall be first class throughout. Syringes and cannulae shall be free from defects which detract from their appearance or may impair their serviceability.



Source: https://assist.dla.mil -- Downloaded: 202610-21T08:05Z
Check the source to verify that this is the currentversion before use.

## 4. QUALITY ASSURANCE PROVISIONS

4.1 Supplier's responsibilities for inspection. The supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own or any other inspection facilities and services acceptable to the Government. Inspection records of the examination and tests shall be kept complete and available to the Government as specified in the contract or order. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling for lot acceptance.

4.2.1 Inspection lot. The inspection lot shall be determined in accordance with Standard MIL-STD-105.

4.2.2 Sampling for examination. A random sample of syringes and cannulae shall be selected from each lot in accordance with Standard MIL-STD-105. Examination for dimensions shall be in accordance with Standard MIL-STD-105, and as indicated in table I.

4.2.3 For tests. Sampling for tests shall be conducted in accordance with Standard MIL-STD-105, and as indicated in table I.

Table I. Sampling Criteria

| Component | Characteristic | Requirement | Test procedure | Unit of product | Insp. level | AQL(percent defective) |
|---|---|---|---|---|---|---|
| Syringe | Boil test | 3.3 | 4.4.1 | 1-syringe | L7* | ** |
| Syringe and cannula | Fluid tightness | 3.4 | 4.4.2 | 1-unit | L7 | 1.5 |
| | Accuracy | 3.5 | 4.4.3 | 1-unit | L5* | ** |
| | Dimensions | 3.2 | - - | 1-unit | L2 | 4.0 |

* A minimum of 10 units shall be tested regardless of lot size.
** Acceptance number shall be zero for all sample sizes.



Source: https://assist.dla.mil -- Downloaded: 202610-21T08:05Z
Check the source to verify that this is the current version before use.

4.2.4 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The syringes and cannulae shall be examined to determine compliance with applicable requirements contained in this specification.

4.4 Tests. Tests shall be conducted, as necessary, to determine compliance with specification requirements.

4.4.1 Boil test. After being thoroughly cleaned, the syringe barrel, with plunger removed, shall be immersed in boiling water for 30 minutes.

4.4.2 Fluid tightness test. This test shall be applied to those samples submitted to the boil test (see 4.4.1). The syringe, with cannula attached, shall be filled with water to the 2 cc. graduation mark. With the thread cap unscrewed and pushed back to the thumb knob, and with the syringe in a horizontal position, a pressure of 10 pounds per square inch gage shall be applied for 10 seconds.

4.4.3 Accuracy test. The syringes shall be tested for accuracy, with cannula attached, at the 1 cc. and 2 cc. graduation mark with a calibrated burette, or by weight, using distilled water.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) syringe and cannula, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table II shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table II. Packaging Quantities

| Unit package | Intermediate package | Exterior container |
|---|---|---|
| 1 unit | 12 units | 144 units |

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.



Source: https://assist.dla.mil -- Downloaded: 202610-21T08:05Z
Check the source to verify that this is the currentversion before use.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be packaged in a carton of appropriate size constructed in accordance with Specification PPP-B-566, or PPP-B-676. Closure shall be adequate to prevent spillage of contents under normal handling.

5.1.3.2 Intermediate package. Intermediate package shall be a carton of appropriate size and design, or a complete wrapping of kraft paper of suitable weight. Closure shall be adequate to prevent spillage of contents under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose within the continental limits of the United States of America.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 3 load, and constructed in accordance with Specification PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; or PPP-B-621, table IV.

5.2.1.2 Case liner. Each level A box shall be lined with a waterproof case liner conforming to Specification MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto.

5.2.1.3 Strapping. Strapping, when required shall be in accordance with applicable paragraphs of appendix thereto.

NOTE: Strapping will not be required for shipment forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to Specification PPP-B-621, table I, or PPP-B-636, type I, class 1, fabricated from not less than 275-pound test fiberboard.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with Standard MIL-STD-129. In addition, contract number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with Standard MIL-STD-129, except as follows: The letter "L" shall precede the Federal stock number. One space shall be allowed between the letter "L" and the first digit of the stock number. The terms "Stock number" or "FSN" shall not be indicated. Service color and service symbol markings are required only on direct shipments from the contractor's plant to overseas activities.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

6.3 Transportation description. Transportation description applicable to this item is:

Surgical instruments NOIBN in boxes.

Carload minimum weight 10,000 pounds subject to Rule 34 of the Uniform Freight Classification.

Truckload minimum weight 10,000 pounds per National Motor Freight Classification.

Notice. When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

Custodians:
Army - Med
Navy - BuMed
Air Force - AFCSG

Preparing activity:
Navy - MMSA



Source: https://assist.dla.mil -- Downloaded: 20261021T08:05Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2)

6.2 This specification does not include all types, classes, grades, sizes, tc., of the commodity indicated by the title of this specification, or which re commercially available, but is intended to cover the types, etc., which re normally procured to meet Military requirements.

6.3 Transportation description. Transportation description applicable o this item is:

Surgical instruments NOIBN in boxes.

Carload minimum weight 10,000 pounds subject to Rule 34 of the Uniform Freight Classification.

Truckload minimum weight 10,000 pounds per National Motor Freight Classification.

---

Notice. When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

Custodians:
Army—Med
Navy—BuMed
Air Force—AFCSG

Preparing activity:
Navy—MCSA



Source: https://assist.dla.mil -- Downloaded: 202610-21T08:05Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:05Z
Check the source to verify that this is the current version before use.