MIL-U-36112A
5 June 1972
SUPERSEDING
MIL-U-36112(MMSA)
14 November 1961

MILITARY SPECIFICATION

UREASE TEST BROTH, DEHYDRATED

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers packets of dehydrated powder which, when prepared, are suitable for the identification of bacteria on the basis of urease activity. (See 6.2).

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of this specification to the extent specified herein:

SPECIFICATION

Federal

PPP-C-00186a	Containers, Packaging and Packing for Drugs, Chemicals, and Pharmaceuticals.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inpsection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specification and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

FSC 6505

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:06Z
Check the source to verify that this is the current version before use.

FIGURE 1. Tube box assembly, typical. (01X02X02)

2.2 Other publication. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

U. S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE,
FOOD AND DRUG ADMINISTRATION

Federal Food, Drug, and Cosmetic Act and
Regulations Promulgated Thereunder.

(Application for copies should be addressed to the Food and Drug Administration, Washington, D. C. 20204.)

3. REQUIREMENTS

3.1 Material. Shall be a stable, dehydrated powder suitable for identification of bacteria on the basis of urea utilization (urease activity) and especially for the differentiation of Proteus from Salmonella and Shigella in the diagnosis of enteric infections. When prepared in accordance with the manufacturer's labeled directions, using purified water, shall contain in each 1000 cc of prepared medium, the following ingredients:

Urea	20.0	Grams
Monopotassium Phosphate	9.1	Grams
Disodium Phosphate	9.5	Grams
Yeast Extract	0.1	Gram
Phenol Red	0.01	Gram

The pH of the prepared medium, after sterilization in an autoclave at 121° C. (15 pounds pressure) for 15 minutes, shall be between 6.70 and 6.90 at 25° C., when determined potentiometrically, using the U.S.P. procedure.

The medium shall contain not less than 95.0 percent and not more that 105.0 percent of the specified amount of each ingredient for each 1000 cc of prepared medium.

Each immediate container (packet) shall contain sufficient ingredients (at least 9.7 grams) for the preparation of 250 cc of medium.

The medium shall be homogeneous in color, free from foreign matter, and free from caking.

3.2 Ingredients. All ingredients entering into the preparation shall be of the highest grade suitable for bacteriological use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

3.3 Delivery. Not more than 6 months shall have elapsed from the date of manufacture of the product, to the date of delivery to the Government.

3.4 Workmanship. Material and its containers shall be free from defects which detract from their appearance of may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Supplier responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 No company supplying any ingredient(s) to the contractor will be considered an acceptable facility for the performance of any inspection requirements specified herein.

4.2 Lot. For purposes of this specification, a lot, batch, or control is that single, uniform, and homogeneous quantity of dehydrated urease test broth produced from one formulation, subjected to the same compounding and manufacturing operation, and filled into final, immediate containers.

4.3 Sampling. Sampling shall be conducted in accordance with MIL-STD-105.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:06Z
Check the source to verify that this is the currentversion before use.

For visual examination	Inspection level*	AQL (percent defective)	Unit of product
Major A	S-2	1.0	Filled packet
Major B	II	1.0	Filled packet
Major C	II	1.0	Filled packet
Major D	II	1.0	Unit package
Major E	II	1.0	Intermediate package
Major F	II	1.0	Exterior package
Minor A	II	2.5	Filled packet

4.3.1 For end item testing. Acceptance number shall be zero (0) for all sample sizes.

Characteristic	Requirement paragraph	Inspection level*	Minimum sample
pH (prepared medium)	3.1	S-2	5
Net contents	5.1	per PPP-C-00186a	10
Differential capability	3.1	S-2	5
Yield	3.1	S-2	5
Leakage	5.1.1	---	10
Label adhesion	5.1	per PPP-C-00186a	10

*Where possible, the same sample shall be used for two (2) or more tests.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

4.4 Examination. The Urease Test Broth, Dehydrated, shall be examined to determine compliance with all requirements of this specification.

4.5 Classification of defects. Examination shall be conducted in accordance with the following classification of defects:

Classification of Defects**

Major A	
101.	Powder not free from caking.
102.	Color of powder not homogeneous.
103.	Powder not free of foreign matter.
Major B	
151.	Labeling not complete.
152.	Labeling not legible, not free from defacing marks and tears.
Major C	
161.	Immediate container not as specified
162.	Immediate container not properly sealed.
Major D	
171.	Labeling not complete.
172.	Labeling not legible, not free from defacing marks and tears.
Major E	
181.	Material and marking of intermediate package not as specified.
Major F	
191.	Material and marking of exterior container not as specified.
Minor A	
201.	Exterior of immediate container not free from contents or foreign material.

**Inspection is not restricted to the classified possible defects listed above.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

4.6 Tests. Tests and examinations shall be performed to determine compliance with all requirements of this document and all referenced documents.

5. PREPARATION FOR DELIVERY

5.1 Preparation for delivery shall be in accordance with all applicable requirements of Interim Federal Specification PPP-C-00186a, dated 15 May 1969, and Amendment-1, dated 27 October 1969, and as specified herein:

5.1.1 Immediate containers. Immediate containers shall be a suitable, flexible, moistureproof packet. The immediate container shall be fabricated of aluminum foil. The foil shall have an inner facing of polyethylene which permits heat sealing, and an outer paper overlay which shall bear the required label information. The layers of the laminated foil shall not separate in handling, storage, and use. The packets shall pass the leakage test specified in 4.3.1.5 of PPP-C-00186a.

5.2 Labeling. Labeling shall be in accordance with the requirements of the Federal Food, Drug, and Cosmetic Act, and shall include the information required below:

See page 7

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

5.2.1 Immediate containers. Each immediate container (packet) label shall bear the following information. However, the information is not required to appear in the sequence indicated:

(a) the item name designated as "UREASE TEST BROTH, DEHYDRATED"

(b) the quantity of contents designated as "9.7 Grams"

Note: The official abbreviation "g." may be used in lieu of the word "grams."

(c) the lot or control number

(d) the Federal Stock No. designated as "FSN 6505-102-4549" or "Stock No. 6505-102-4549"

(e) the name and address of the manufacturer. When the manufacturer is not the contractor, the name and address of the contractor shall also appear.

When both names are placed on the label, the following designations shall precede the names:

"MFR" for the manufacturer and
"CONTR" for the contractor.

(f) directions for preparing the medium for use.

(g) the formula of the product for 1000 cc of prepared medium

(h) the date of manufacture.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

5.2.2 Unit packages. Each unit package label shall bear the following information. However, the information is not required to appear in the sequence indicated:

(a) the item name designated as "UREASE TEST BROTH, DEHYDRATED"

(b) the quantity of contents per immediate container designated as "9.7 Grams"

Note: The official abbreviation "g." may be used in lieu of the word "grams."

(c) the Federal Stock No. designated as "6505-102-4549" or "Stock No. 6505-102-4549"

(d) the lot or control number

(e) the date of manufacture

(f) the name and address of the manufacturer. When the manufacturer is not the contractor, the name and address of the contractor shall also appear.

When both names are placed on the label, the following designations shall precede the names:

"MFR" for the manufacturer and
"CONTR" for the contractor.

(g) directions for preparing the medium for use

(h) the formula of the product for 1000 cc of prepared medium

(i) the statement "Store at controlled room temperature (59° - 86° F.)."

(j) the unit of issue designated as

"1 PG
(of 6 packets)"

The parenthetical phrase shall appear in smaller characters than the unit of issue designation.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

5.3 Packaging and packing.

5.3.1 Unit of issue. One package containing 6 packets, as specified, constitutes one unit of issue.

5.3.2 Unit package. Each unit shall be packaged as specified in 5.2.5 of PPP-C-00186a.

5.3.3 Procedure code. Procedure code No. 4 applies, except that in Table I, column 2, of PPP-C-00186a, delete "12" and substitute "6" and in column 4, delete "12/96" and substitute "6/72."

5.4 Marking.

5.4.1 Intermediate package. Each intermediate package shall be marked as specified in paragraph 5.5.3 of PPP-C-00186a, except that at end of paragraph add: "Date of manufacture shall be shown in lieu of date packed. Marking shall include the following legend:

"STORE AT CONTROLLED ROOM TEMPERATURE (59° - 86° F.)."

5.4.2 Exterior container. Each exterior container shall be marked as specified in 5.5.4 of PPP-C-00186a, except that at end of paragraph, add: "Date of manufacture shall be shown in lieu of date packed. Marking shall include the following legend:

"STORE AT CONTROLLED ROOM TEMPERATURE (59° - 86° F.)."

6. NOTES

6.1 Ordering data. Purchasers should exercise any desired options offered herein, and the procurement document should specify the following:

(a) Number and date of this specification
(b) Federal Stock No. (FSN)
(c) Selection of applicable levels of packaging and packing (See PPP-C-00186a).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

6.2 Federal Stock No. coverage. This specification covers the following item appearing in the Federal Supply Catalog:

Federal Stock No.	Item Identification	Unit
6505-102-4549	UREASE TEST BROTH, DEHYDRATED, 6s	Package

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6505-1288

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MILITARY SPECIFICATION MIL-U-36112A

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.

FOLD

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

DOD-304

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:06Z
Check the source to verify that this is the current version before use.