MILITARY SPECIFICATION

PSYCHROMETER, HAND, ELECTRIC, ASPIRATING; ML-450 ( )/UM AND CARRYING CASE, PSYCHROMETER, CY-1238 ( )/UM

This amendment forms a part of Military Specification MIL-P-17098A, 14 Oct. 1952, and has been approved by the Department of Defense and is mandatory for use by the Departments of the Army, the Navy, and the Air Force.

Page 2, Paragraph 3.4.2 *Delete* the entire paragraph and substitute the following: "Case.—The case shall be made of lightweight metal or a suitable plastic material such as to withstand continued hard usage aboard ship."

Page 2, Paragraph 3.5 *Change* title to: "DESIGN AND CONSTRUCTION"

Add "The psychrometer shall conform basically to the manufacturer's individual design subject to the requirements set forth hereunder."

Page 3, Paragraph 3.5.1 First sentence. Place a period after the phrase "mercury-in-glass type." *Delete* remainder of sentence and substitute new sentence: "The scales shall be graduated in whole degrees Fahrenheit over a minimum length of 4½ inches and not to exceed 6½ inches."

Page 3, Paragraph 3.5.2 Change "0.5° Fahrenheit" to "0.4° Fahrenheit".

Page 3, Paragraph 3.5.3 *Insert* after first sentence "Minimum air speed when produced by action of the ventilating fan shall be in excess of 15 feet per second."

Page 3, Paragraph 3.5.6 *Delete* entire paragraph and substitute the following: "3.5.6 Wick Moistener: A means for manually or automatically moistening the wick of the wet bulb thermometer shall be provided within the instrument, and shall hold at least 1 ounce of water. The design of an automatic wick-moistening arrangement may be any which will conveniently and reliably deliver directly the correct amount of water to saturate the wet bulb wick."

Page 3, Paragraph 3.5.7 *Delete*

Page 3, Paragraph 3.5.9.1 *Delete* "a water bottle of four ounces capacity" and substitute "four ounces of bottled water."

Page 3, Paragraph 3.5.10 *Delete* entire paragraph and substitute: "3.5.10 Nameplate. A nameplate satisfactory to the procuring agency shall be affixed to the carrying case and also affixed to, or imprinted into, the housing of the instrument. For Navy purchases, see paragraph 6.2 under Notes."

Page 3, Paragraph 3.5.11 *Delete* entire paragraph and substitute: "3.5.11 Instructions. A brief set of operating instructions in summary form together with appropriate psychometric tables or charts shall be provided within the carrying case, either in pamphlet form or in laminated sheet plastic form."

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:45Z
Check the source to verify that this is the current version before use.

Page 3, Paragraph 3.5.12 *Delete* "a water bottle of four ounce capacity" *Add* "a water bottle or bottles of four ounces total capacity."

Page 3, Add Paragraph "3.5.13 Dimensions. Dimensions shall be in a suitable proportion for easy handling, carrying and reading of thermometers and shall not exceed 10″ in length, 4½″ in height and 2″ in depth."

Page 3, Add Paragraph "3.5.14 Weight. The instrument, complete with batteries, shall be as light in weight as is consistent with other requirements of the specification and shall not exceed 3 pounds."

Page 6, Paragraph 6.1 After the words "Prospective suppliers" insert, "under Navy invitations to bid."

Page 6, Paragraph 6.2 *Delete* entire paragraph and substitute: "6.2 Nameplate and Nomenclature (Navy Purchases Only). For procurements by the Bureau of Aeronautics or for the Bureau of Aeronautics by the Navy Purchasing Office, Washington or the Aviation Supply Office, Philadelphia, nomenclature should be obtained and nameplate will be furnished in accordance with specification MIL–N–18307(Aer), "Nomenclature and Nameplates for Airborne Electronic and Associated Equipment" with the following modifications:

(a) All references to Bureau of Aeronautics (Aer–EL–419) should be deleted and Bureau of Aeronautics (Aer-AY–2) should be substituted.

(b) All general provisions of this specification will apply as appropriate, in the same manner as they would apply to equipment "associated" with airborne electronics equipment under the terms of the specification.

(c) Nameplate format should be in accordance with Figure 1 of the specification, deleting the data between the double horizontal lines at the center and substituting "Stock Number (*as assigned*)." Dimensions may be proportionately reduced as necessary to fit space available on the instrument and carrying case.

(d) On instruments with plastic housing, the imprinting or molding of nameplate data directly into the surface of the housing may be allowed subject to the approval of the Bureau of Aeronautics (AY–2)."

Page 6, Add:

Notice. When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation or conveying any rights or permission to manufacture, use or sell any patented invention that may in any way be related thereto.

Custodians:
Army—Signal Corps
Navy—Bureau of Aeronautics

Other interest:
Army—MT
Navy—Md.

Preparing activity:
Navy—Bureau of Aeronautics

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:45Z
Check the source to verify that this is the current version before use.