A-A-0053166(DM)
10 September 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

PORTABLE BLOOD WARMER

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a portable blood warmer used to control fluid infusion temperature when used with intravenous sets.

Salient characteristics:

General. Shall be a device which provides external heat to infusion blood tubing to warm up blood or fluid flowing during transfusion.

Material. The exterior cabinet shall be fabricated of high impact plastic and the heating plates of teflon coated aluminum.

Construction. The warmer shall be a rectangular shaped cabinet consisting of a front and back section, connected by a hinge, each having a flat heating plate. The back heating plate shall have depressed convaluted channels to allow blood infusion tubing to be contained in channels for heating of blood or fluids. The warmer shall accommodate standard 80 to 110 inch long I.V. sets. An inlet to the channels shall be provided at the top of the plate and two outlets at the bottom to accommodate various lengths of I.V. tubing. Priming shall not be required for this device. Clamping provisions shall be provided on the back of the cabinet to allow mounting to a standard I.V. pole. Instructions for use shall be permanently attached to the back of the cabinet. The warmer shall provide sufficient heating capacity to maintain the following fluid exit temperatures:

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:16Z
Check the source to verify that this is the current version before use.

Inlet water temperature	Flow Rate	Exit Temperature
4-6°C	5 ml/min	33°C
4-6°C	25 ml/min	29°C
4-6°C	100 ml/min	22 ± 2°C

The heating plate shall have sufficient electrical capacity to warm to 37°C within 3 minutes. The device shall have solid state integrated circuitry with the panel displaying digital heater temperature readout and containing power indicator light. An independent audio alarm and automatic power cut-off shall be provided to function if heating plate temperature exceeds 40°C. A push button shall be provided on the cabinet back to allow the alarm system to be manually tested. The device shall be approximately 8 inches wide by 6 inches high by 4 inches deep with a weight of 5 lbs.

The warmer shall operate on 110 to 115 VAC, 60 Hz and 220 VAC, 50 Hz electrical power. A 15 foot long power cable with a hospital grade plug shall be provided. The blood warmer shall be in compliance with all applicable requirements of the Bureau of Medical Devices and Diagnostic Products.

Workmanship. The blood warmer shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One blood warmer, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:16Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one blood blood warmer.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-150-7840	WARMER, BLOOD, Portable, 110/115 Volt, 60 Hz, AC, and 220 Volt, 50 Hz, AC, Solid State

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 51574

Project No. 6515-4186

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (DO NOT STAPLE), and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation was too rigid, restrictive, loose, ambiguous, or incorrect, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

Fold along this line

Fold along this line

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL

Commander
Defense [illegible] nel Support Center
ATTN [illegible] S
2800 [illegible] n Street
Philadel[illegible] Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

See Instructions – Reverse Side

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-0053166 (DM)	PORTABLE BLOOD WARMER

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION (Mark one)
	☐ VENDOR
	☐ USER
b. ADDRESS (Street, City, State, ZIP Code)	☐ MANUFACTURER
	☐ OTHER (Specify)

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER (Last, First, MI – Optional)	b. WORK TELEPHONE NUMBER (Include Area Code) – Optional
c. MAILING ADDRESS (Street, City, State, ZIP Code) – Optional	8. DATE OF SUBMISSION (YYMMDD)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

(FOLD ON THIS LINE)

(FOLD ON THIS LINE)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.