A-A-0053184(DM)
18 September 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

FIRST AID KIT, INDIVIDUAL

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers one type of individual First Aid Kit for use by Government Agencies.

Salient characteristics:

Shall be an assembly of medical supplies suitable for individual use in first aid treatment.

Each component item shall be in total conformance with the latest specification that is applicable.

All components that are required by specifications to be expiration dated shall have not less than 30 months of the expiration dating period remaining the time of delivery to the receiving activity(ies) as outlined in the contract.

The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions shall be obtained from the procuring activity or as directed by the contracting officer.)

Specification	(for)	National Stock Number
MIL-C-37183		6545-01-094-6142
MIL-I-3685		6545-01-094-6136
A-A-51387		6505-00-914-3593

AMSC N/A FSC 6545

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

Defense Medical Purchase Description No. 1 Dated 26 December 1972	6508-00-116-1473
MIL-D-37165	6510-00-159-4883
DDD-B-61	6510-00-200-3180
Defense Medical Purchase Description No. 13 Dated 17 August 1987	6510-00-201-1755
DDD-B-0035	6510-00-913-7909
Medical Procurement Item Description No.3 Dated 10 June 1987	6530-00-042-8421
MIL-F-36043	6545-00-853-6309
MIL-G-283	6850-00-985-7166
Medical Procurement Item Description No. 3 Dated 10 June 1987	6530-00-042-8416

Sample Available, see page 4 Graphic Training Aid 21 - 45

The Kit shall be packaged in accordance with the packing list. The proper quantity (QTY) is indicated from the unit of issue (U/I) for each item on the packing list.

The contractor shall print a packing list bearing the following information:

"FOLLOWING ARE THE COMPONENTS FOR:

6545-01-094-8412 FIRST AID KIT, INDIVIDUAL

6545-01-094-6142	CASE, MEDICAL INSTRUMENT AND SUPPLY SET, NYLON, NONRIGID, NO. 8, 4-3/4 INCHES LONG BY 5-1/2 INCHES HIGH	1	EA
6545-01-094-6136	INSERT, FIRST AID KIT CASE, PLASTIC, OLIVE DRAB, 4-3/4 INCHES LONG BY 2-3/8 INCHES WIDE BY 5-1/2 INCHES HIGH	1	EA

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

6505-00-914-3593	POVIDINE-IODINE SOLUTION, USP, 10%, 1/2 FL OZ (14.8 ML), 50S	1/50	BX
6508-00-116-1473	LIPSTICK, ANTICHAP, HOT CLIMATE, 3.7 GRAM, 100S	1/100	BX
6510-00-159-4883	DRESSING, FIRST AID, FIELD, INDIVIDUAL TROOP, CAMOUFLAGED, 4 BY 6-1/4 - 7-1/4 INCHES	2	EA
6510-00-200-3180	BANDAGE, GAUGE, COMPRESSED, CAMOUFLAGED, 2 INCHES BY 6 YARDS	1	EA
6510-00-201-1755	BANDAGE, MUSLIN, COMPRESSED, CAMOUFLAGED, 37 BY 37 BY 52 INCHES	1	EA
6510-00-913-7909	BANDAGE, ADHESIVE, 3/4 BY 3 300S	18/300	BX
*6530-00-042-8421	BOTTLE, SAFETY CAP, LIGHT-RESISTANT, PLASTIC, TABLET AND CAPSULE ROUND, 7 DRAM (26 ML), 200S	2/200	PG
6545-00-853-6309	FIRST AID KIT, EYE DRESSING	1	EA
6850-00-985-7166	WATER PURIFICATION TABLET, IODINE, 8 MG, 50S	1	BT
V**	INSTRUCTION CARD, ARTIFICIAL RESPIRATION, MOUTH-TO-MOUTH RESUSCITATION (GRAPHIC TRAINING AID 21-45)		

**REPLACEMENT CARDS MAY BE OBTAINED THROUGH APPROPRIATE TRAINING-AUDIOVISUAL SUPPORT CENTERS.

*6530-00-042-8416	BOTTLE, SAFETY CAP, PLASTIC, LIGHT-RESISTANT, TABLET AND CAPSULE, ROUND, 4 DRAM (15 ML), 200S MAY BE SUBSTITUTED UNTIL STOCKS ARE DEPLETED.	

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

ASSEMBLY BUILT ACCORDING TO UA CODE DECK 7175 AS FURNISHED BY DIRECTORATE OF MEDICAL MATERIEL, DEFENSE PERSONNEL SUPPORT CENTER, DATED 1 NOVEMBER 1980.

<NAME OF ASSEMBLY COMPANY - CONTRACT NUMBER>

PAGE 1 OF 1 PAGE"

Graphic Training Aid shall be reproduced in color. Samples are available from the procurement activity. The Graphic Training Aid produced by or for the contract must match the sample from the procurement activity.

The packing list shall be printed in a neat and legible fashion. The packing list shall be printed on a 3 inch by 5 inch piece of paper.

A procedure shall be established by the contractor to verify in writing the correct items are being used, the correct quantities are placed in each Kit and the appropriate shelf life remains on items with potency dating as outlined above. This procedure shall be verified before assembly of the kits begins. The contractor shall maintain records and implement the controls necessary to insure the quality of the end item.

The Government reserves the right to change any component(s), the packing list, procurement specifications or packing instruction as deemed necessary.

Packing instruction. Shall be packed to contain the quantity of each individual NSN as specified on the packing list. Where two numbers appear, the quantity given is the first number per kit. Two First Aid Field Dressings and one Muslin Bandage shall be packaged inside the Plastic Insert to one side. On the other side of the Insert the two plastic Bottles shall be placed at the bottom. The Povidone-Iodine shall be placed into one of the empty plastic Bottles (Do not open the Povidine-Iodine, It should fit into the plastic Bottle) and place the Water Purification Tablets (Do Not Open the Water Purification Tablets) into the other empty plastic Bottle. The two plastic Safety Caps shall be placed into the Insert. The Antichap Lipstick and three packages containing six Adhesive Bandages each shall be placed into the Insert. The Gauze Bandage shall be pressed into the lid of the Insert.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

The Insert shall be closed and snapped tight. No materials should be visible outside of the plastic Insert. The Insert should then be placed into the Nylon Case. The convex portion of the Insert shall be placed rightside up into the front of the Case. The front of the Case is considered to be the side where the male portion of the Curtain Snaps are located.

Behind the Insert inside of the Case, (concave side of the Insert) is where the remaining components should be neatly placed. These components are: Eye Dressing First Aid Kit, Graphic Training Aid 21-45, and the Component Packing List for the Individual First Aid Kit.

The completed Kits shall be snapped closed and visually examined for non-conformances. The components should be packaged snugly and not fall out or open accidentally during any phase of shipment or handling.

Storage of the completed Kits shall be in a controlled environmental temperature range of 59 to 86 degrees Fahrenheit. The Kits shall not be exposed to extreme temperature ranges with the exception of shipment to the Government.

Workmanship. The Individual First Aid Kit shall be free from defects which detract from it's appearance or impair it's serviceability.

Unit. One kit (KT), as specified, constitutes one unit.

Packing Quantities. Twenty-four units shall be packed in an exterior container.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspection shall be at the contractors plant. For the purpose of examination only, the special inspection level S-4 shall be used with an Acceptable Quality Level (AQL) of 4.0.

For the purpose of testing, special inspection level S-1 shall be used with an AQL of 1.0. It shall be the discretion of the Government Inspector whether or not to use the same samples from the previous inspection.

Sampling for test. Sampling for tests shall be conducted in accordance with MIL-STDs-105 and as specified herein. The unit of product for tests purposes shall be one Individual First Aid Kit.

Shelf life testing. Completed assemblies shall have a shelf life of not less than 30 months at the time of delivery to the first Government activity(ies) as outlined in the contract.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate package labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6545-01-094-8412 FIRST AID KIT, INDIVIDUAL D 764

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

AGENT:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6545-0090

Location: Disk # 10

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.