A-A-0053209 (DM)
24 September 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

CANNULA, SUCTION (Frazier)

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a suction tube of the Frazier design, with cleaning stylet.

Salient characteristics.

Sizes. Tubes shall be supplied in the following sizes, as specified in the contract or purchase order:

Size 4-1/2 4-1/2 ± 1/4 inches long

Size 7-1/2 7-1/2 ± 1/4 inches long

Material. The cannula tube, tubing connector, stylet, and stylet handle shall be manufactured from austenitic corrosion-resistant steel in accordance with American Society for Testing and Materials (ASTM) Standard F899. As an alternate, the cannula tube may be manufactured from unplated nickel-copper alloy (Monel) of composition in accordance with ASTM B165, and the tubing connector and stylet handle manufactured from nickel plated brass in accordance with Federal Specification GG-I-526.

Dimensions and design. The cannula shall consist of the tube, the tubing connector with thumb control port, the cannula, and the stylet. The cannula tube shall be an "L" shaped tube 0.144 ± 0.015 inches in outside diameter. The 90 degrees shall be formed with a bend of 1 ± 1/8 inch radius of curvature. The distal leg shall be of the length as designated by the nominal size designation, with the measurement made from the distal tip to the plane parallel to the farthest side of the proximal leg. The proximal leg, including tubing connector, shall be 4 ± 1/2 inches in length when measured in a similar fashion.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

The tubing connector shall be permanently attached to the cannula tubing. The connector shall be designed to provide a mechanically secure and leakproof connection with suction tubing ranging in inside diameter from 1/4 to 5/16 inch (as a minimum).

The thumb control port shall be integral to the tubing connector, or permanently attached to the cannula tube within 2 inches from the connector. The port shall be a 0.050 inch diameter hole (minimum) into the lumen of the cannula. The hole shall be centered on a thumb rest plate of 7/16 inch width (minimum).

The proximal end of the tubing connector lumen shall be a female Luer taper in accordance with American National Standards Institute/Health Industry Manufacturer's Association (ANSI/HIMA) MD70.1.

The stylet shall consist of a 1/32 inch (nominal) steel wire permanently attached to a handle or knob. The handle or knob shall be knurled on the circumference and shall have a male Luer taper on the distal end. The Luer taper shall be in accordance with ANSI/HIMA MD70.1. The wire shall be of sufficient length that it protrudes slightly (less than 3/16 inch) from the cannula when the Luer tapers are seated.

Finish. The finish of steel components shall be satin. The finish of all components shall be uniform and free from burrs, pits, grind or tool marks, rough areas, or corrosion. All components shall be free from sharp edges.

Passivation. Cannulas manufactured from austenitic corrosion-resisting steel shall be properly passivated.

Markings. The cannula shall bear the manufacturer's name or registered trademark and the French size. Cannulas manufactured from austenitic corrosion-resistant steel shall bear the word "stainless". Cannulas of foreign manufacture shall bear the name of the country of manufacture.

Workmanship. The cannula and stylet shall be processed in such a manner as to be uniform and free from defects which affect life, utility, or appearance.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Each (EA). One cannula, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. The cannulas shall be examined to determine compliance with specification requirements.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:16Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one cannula. Sampling shall be level S-1 with an AQL of 1.0.

Corrosion test. Cannulas shall be tested in accordance with Resistance to Boil Test of Federal Specification GG-I-526. In addition, Monel and brass components shall be tested in accordance with the Resistance to Salt Spray Test of GG-I-526.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standard, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:16Z
Check the source to verify that this is the currentversion before use.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-139-8257	CANNULA, SUCTION, Frazier, Angular, 11 French, 4-1/2 Inch (11.4 cm) Long
6515-01-139-8258	CANNULA, SUCTION, Frazier, Angular, 11 French, 7-1/2 Inch (191 mm) Long

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4204

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:16Z
Check the source to verify that this is the current version before use.