A-A-51394A
1 June 1988
SUPERSEDING
A-A-0051394(DM)
3 April 1986

COMMERCIAL ITEM DESCRIPTION

BANDAGE, GAUZE (TWO-PLY)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a sterile, roller-type, elastic gauze bandage, suitable for medical use.

Salient Characteristics:

The bandage shall be a sterile, roller-type, two-ply elastic (stretchable), conformable, self adhering gauze bandage, 5 yards minimum length in the stretched condition, suitable for medical use and for securing dressings in place and for bandaging bony prominences.

The finished (treated) gauze shall meet the requirements of Absorbent Gauze, USP, except for thread count and weight and except that water extract shall not exceed 28 mg and except that the gauze shall be all cotton. The gauze shall have undergone treatment causing it to shrink and have an excess crimp in the warp direction or in the warp and fill directions. The finished gauze shall have a minimum of 12 yarns per inch in the fill and a minimum of 14 yarns per inch in the warp direction with the gauze in the relaxed condition. The weight of the finished gauze in the relaxed condition shall range from 20 to 25 grams per square meter. The length of the unstretched bandage shall be a minimum of 120 inches long.

The bandage shall be made by folding the treated gauze to two-ply by turning the lateral edges inwardly toward each other to an abutting position at approximately the center line of the bandage. The folded gauze shall be rolled with the folded edges to the outside of the roll. The edges of the rolled bandage shall be uniformly aligned. The gauze shall not be stretched in the roll, but shall be wound under sufficient tension to keep the roll from gapping.

FSC 6510

AMSC/NA

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

The finished bandage shall contain no rubber, adhesive or sizing materials; shall be of one continuous piece, shall not be cut on the bias, and shall be free from cuts, loose threads, and ravelings.

Workmanship. The Bandage shall be free from defects which detract from its appearance or impair its serviceability.

TABLE I

Width (+/- 1/8 inch)	SIZE
2 inches	1
3 inches	2
4 inches	3
6 inches	4

Unit. Package (PG).

Each bandage shall be individually packaged in a peel-open package which shall maintain sterility of contents unless package is damaged or opened.

For Size 1, one package containing 96 individually packaged bandages, as specified, constitutes one unit.

For Size 2 and Size 3, one package containing 72 individually packaged bandages, as specified, constitutes one unit.

For Size 4, one package containing 48 individually packaged bandages, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Testing and examination shall be conducted to determine compliance with all specification requirements. Test methods shall be suitable, accurate, and reproducible.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for examination shall be one individually packaged bandage. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5 maximum percent defective.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:55Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	SIZE
6510-00-715-3102	BANDAGE, GAUZE, Two-Ply 2 inches by 5 yards, 96s	1
6510-00-717-5094	BANDAGE, GAUZE, Two-Ply 3 inches by 5 yards, 72s	2
6510-00-717-5095	BANDAGE, GAUZE, Two-Ply 4 inches by 5 yards, 72s	3
6510-00-724-1017	BANDAGE, GAUZE, Two-Ply 6 inches by 5 yards, 48s	4

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 05050

Project No. 6510-0741

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51394A	BANDAGE, GAUZE (TWO-PLY)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION (Mark one)
	☐ VENDOR
	☐ USER
b. ADDRESS (Street, City, State, ZIP Code)	☐ MANUFACTURER
	☐ OTHER (Specify)

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER (Last, First, MI) – Optional	b. WORK TELEPHONE NUMBER (Include Area Code) – Optional
c. MAILING ADDRESS (Street, City, State, ZIP Code) – Optional	8. DATE OF SUBMISSION (YYMMDD)

DD FORM 1426

Source: https://assist.dla.mil – Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.