A-A-53180
14 September 1987

COMMERCIAL ITEM DESCRIPTION

SEALER-HOLDER, CAPILLARY TUBE

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a plastic disposable tray containing a trough of sealant material for flameless sealing of blood filled microhematocrit tubes.

Salient characteristics:

Shall be a disposable plastic tray containing sealant material that is suitable for both flameless sealing of blood-filled microhematocrit tubes and retention of the sealed tubes until they are ready to be centrifuged.

Shall be a disposable 24 place sealer-holder for flameless sealing of blood filled microhematocrit tubes having internal diameters ranging from 0.8 to 1.85 millimeters. Shall effectively seal and retain an inserted microhematocrit tube in a vertical position until it is ready for centrifuging.

Shall consist of a rigid, flat bottomed plastic tray with a trough (compartment) containing sealant material that is white in color. Shall be supplied with protective covering that:

1. Covers the sealant material.

2. Adheres to the tray so that it does not slip off, but can be easily removed when the tray is ready for use.

The sealant material shall be white in color to provide ready contrast with a blood column. The sealant material shall be a putty-like material that is:

1. Sufficiently plastic and pliable to allow insertion, plugging, and retention of microhematocrit tubes.

2. Free of lumps, grit, and any impurities which may interfere with laboratory procedures.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

When a blood filled microhematocrit tube is pushed into the sealant with a slight twist, the sealant shall form a small plug in that end of the tube which:

1. Seals that end of the tube against leakage.

2. Does not dislodge or spin out during centrifuging.

The overall dimensions of the tray shall be nominally 4-7/8 long by 1-13/16 wide by 5/16 high (inches). The topside of the tray shall be permanently and distinctly numbered from 1 to at least 24. The numbering shall be spaced so that:

1. Each numbered sector can accommodate two microhematocrit tubes in the trough.

2. Each inserted microhematocrit tube can be identified in terms of where its being positioned with respect to one of the numbered sectors.

Suitable instructions shall be supplied on the proper use of the sealer-holder with microhematocrit tubes. Instructions may be supplied with each sealer-holder or as an insert in the unit package.

Workmanship. The sealer-holder shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing six sealer-holders, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with commercial practice.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

6640-01-010-4122

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.)

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6640-1453

Location: Disk # 9

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be [illegible], folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 8, be as specific as possible about particular problem areas such as wording which required interpretation, was rigid, [illegible], loose, ambiguous, or was not possible, and give proposed wording changes which would alleviate the problems. Enter in block 8 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgment will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4066 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1. DOCUMENT NUMBER
A-A-53189

2. DOCUMENT TITLE
SEALER-HOLDER, CAPILLARY TUBE

3a. NAME OF SUBMITTING ORGANIZATION

4. TYPE OF ORGANIZATION *(Mark one)*
- ☐ VENDOR
- ☐ USER
- ☐ MANUFACTURER
- ☐ OTHER *(Specify)*

3b. ADDRESS *(Street, City, State, ZIP Code)*

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. P / - Wording

c. Reason/Rationale for P - [illegible]

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* - Optional

b. WORK TELEPHONE NUMBER *(Include Area Code)* - Optional

c. MAILING ADDRESS *(Street, City, State, ZIP Code)* - Optional

8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426

PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:16Z
Check the source to verify that this is the current version before use.