A-A-53192A
22 January 1988
Superceding
A-A-0053192(DM)
16 September 1987

COMMERCIAL ITEM DESCRIPTION

SPLINT, TRACTION-EXTRICATION

The General Services Administration has authorized the use of this interim commercial item description.

This commercial item description covers a lower extremity immobilizer and traction splint, suitable for emergency stabilization and/or traction of angulated and non-angulated fractures and dislocations.

Salient characteristics:

Shall be a dual hinged, half-ring, articulating, pre-extrication, long bone immobilizer and traction splint, lower extremity, adult size, with a carrying case. Suitable for emergency stabilization and/or traction of angulated and non-angulated fractures and dislocations.

The frame shall be fabricated of aluminum and stainless steel, and shall extend to a minimum of 100 cm (39.37 inches). All metal parts shall have been properly hardened and passivated to resist corrosion.

The splint shall be hinged near the middle to allow angulation at the knee in the lateral, medial, flexed or extended direction.

The splint shall allow for dorsal or volar placement.

Shall have six (6) webbing straps with snap release buckles. Five (5) straps shall be elastic and one (1) ischial strap shall be non-elastic.

Shall have an ankle strap with hook and pile closure and webbing straps with D-rings to provide connection to the traction mechanism for traction application.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:16Z
Check the source to verify that this is the current version before use.

The traction mechanism shall be a detachable traction ratchet, spring loaded for single hand operation, that shall be attached when traction is desired and removed when traction is not required. Traction shall be released by pushing the ratchet hand control against the spring loading while still maintaining hand control of applied traction. The traction mechanism shall include a folding heel stand.

All metal crossmembers shall be padded to prevent undue pressure on soft tissue.

The splint shall have a highly contoured, pivoting, non-displacing ischial pad.

The splint shall fold and compress upon itself for storage within its carrying case.

The carrying case shall be made of fabric with handles or a draw string for carrying or hanging. The case shall have accommodation for the splint, traction mechanism and ankle strap.

Each splint shall contain "Directions for Use".

Packaging and labeling shall be that normally furnished by the supplier as commercial practice.

Workmanship. The splint, traction mechanism, ankle strap and carrying case, shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One splint complete with carrying case and direction for use, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:16Z
Check the source to verify that this is the currentversion before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one splint with traction assembly and ankle strap in carrying case.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-01-250-8936

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

AGENT:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4211

Location: Disk #10

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:16Z
Check the source to verify that this is the currentversion before use.